

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

box if there is
qualified data on
this date.

Meter ID: Betty Boop

Project: SNP RD 2016-05-02-80

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
(2) Always wait for meter to stabilize before recording any readings.
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/16

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	5/5/16	715	22.0	8.7	8.8	101	33	1.4	P	L
ICV				22.0	8.7	8.8	101	33	1.4	P	L
CCV	Terry Riordan	5/5/16	1435	22.0		8.8	100	33	1.4	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25 75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	5/5/16	715	2601878	12/17	1000	1000	P	L
ICV				2510F25	10/17	100	101	P	L
CCV	Terry Riordan	5/5/16	1435	2601878	12/17	1000	1000	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	5/5/16	715	157448	12/17	7.0	7.0	-25	P	L
Calibr.				157226	11/17	4.0	4.0	150	P	L
Calibr.				156442	12/17	10.0	10.0	-199	P	L
ICV				156442	12/17	10.0	10.0	-198	P	L
CCV	Terry Riordan	5/5/16	1435	156442	12/17	10.0	10.0	-202	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

mV pH 4 Range +180 $\pm$ 50;

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

box if there is
qualified data on
this page.

Meter ID:

11E/11C

Project:

201605-02-77 SMP

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

5/16

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp ($^{\circ}$ C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Mike Drennan	5/5/16	21.6	800	8.8	8.8	100	46	1.0	P	L
ICV			21.6	800	8.8	8.8	100	47	1.0	P	L
CCV			1515	22.9	8.6	8.7	101	47	1.0	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Mike Drennan	5/5/16	800	2601878	12/17	1000	1000	P	L
ICV				2510425	10/17	100	100	P	L
CCV			1515	2510425	10/17	100	100	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Mike Drennan	5/5/16	800	157448	12/17	7.0	7.0	-10	P	L
Calibr.				157226	11/17	4.0	4.0	166	P	L
Calibr.				156442	2/17	10.0	10.0	-190	P	L
ICV				156442	12/17	10.0	10.0	-190	P	L
CCV			1515	156442	12/17	10.0	10.2	-204	P	L
CCV				157226	11/17	4.0	4.1	158	P	L

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

mV pH 4 Range +180 $\pm$ 50;

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater