

Chemical Analysis Report

CEN-DIST-2016-01-29-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Canal 485A Springs Grp+HA**
Request ID: **RQ-2016-01-25-12**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 17-FEB-2016 11:06

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: REEDY CREEK LOWER @ US 192

Collection Date/Time: 01/28/2016 10:07

Field ID: G4CE0049

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1766864	EPA 180.1 Rev. 2.0	Turbidity	1.6		NTU	P297805	
	EPA 300.0 Rev. 2.1	Chloride	39	A	mg Cl/L	P298129	
		Sulfate	8.9	A	mg SO4/L	P298126	
	SM 2120 B	Color (true)	370	A	PCU	P297784	
	SM 2320 B-97	Total Alkalinity	25		mg CaCO3/L	P298140	
	SM 2540 C-97	TDS	163		mg/L	P298298	
	SM 2540 D-97	TSS	3	I	mg/L	P298205	
	SM 4500 F-C-97	Fluoride	0.049	I	mg F/L	P298136	
1766865	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	I	mg N/L	P298069	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P298032	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.094		mg N/L	P297927	
	EPA 365.1 Rev. 2.0	Total-P	0.043		mg P/L	P298030	
	SM 5310 B-00	Organic Carbon	33		mg C/L	P298151	
1766866	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.014		mg P/L	P297798	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

SM 4500 F-C-97: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P297805

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P298126

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P298129

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P298069

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P298032

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P297927

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P298030

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P297798

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P297784

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P298140

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P298298

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P298205

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P298136

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P298151

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P298126

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P298129

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	106		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P298069

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P298032

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P297927

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P298030

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P297798

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P298140

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	95 - 105

Reference Method: SM 4500 F-C-97
Batch ID: P298136

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.5		P	94 - 108

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P298151

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P298126

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1766864	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P298069

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1766273	Ammonia-N	94.5	97.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P298032

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1766273	Kjeldahl Nitrogen	93.6	93.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P297927

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1766272	NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P298030

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1766865	Total-P	100	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P297798

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1766866	O-Phosphate-P	98.6	99.6	P/P	90 - 110

Reference Method: SM 5310 B-00
Batch ID: P298151

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1767035	Organic Carbon	96.0	96.4	P/P	85 - 115

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P297805

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766849	Turbidity	0.0	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P298126

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766864	Sulfate	0.00335	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P298129

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766864	Chloride	0.257	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P298069

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766273	Ammonia-N	2.43	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P298032

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766273	Kjeldahl Nitrogen	0.154	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P297927

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766272	NO2NO3-N	0.320	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P298030

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766865	Total-P	2.04	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P297798

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766866	O-Phosphate-P	0.884	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P297784

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766864	Color (true)	2.12	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P298140

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766855	Total Alkalinity	0.342	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P298298

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1767033	TDS	2.63	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P298136

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1766855	Fluoride	0.513	Spike	P	0 - 3.7

Reference Method: SM 5310 B-00
Batch ID: P298151

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1767035	Organic Carbon	0.371	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A67348
Included Lab Sample IDs: 1766864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	104	102	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A67353
Included Lab Sample IDs: 1766866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.0	99.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A67357

Included Lab Sample IDs: 1766864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A67405

Included Lab Sample IDs: 1766865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A67453

Included Lab Sample IDs: 1766865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.0	98.0	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A67468

Included Lab Sample IDs: 1766864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	105	105	P/P	90 - 110
Sulfate	101	97.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A67469

Included Lab Sample IDs: 1766864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	97.5	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A67470

Included Lab Sample IDs: 1766864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.1	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A67475

Included Lab Sample IDs: 1766865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.4	97.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A67479

Included Lab Sample IDs: 1766865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A67488

Included Lab Sample IDs: 1766865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.0	
EPA 300.0 Rev. 2.1	Chloride	106			0.257	
	Sulfate	102	101		0.0033	
EPA 350.1 Rev. 2.0	Ammonia-N	95.9	94.5 97.6			2.43
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.0	93.6 93.3			0.154
EPA 353.2 Rev. 2.0	NO2NO3-N	102	101 102			0.320
EPA 365.1 Rev. 2.0	Total-P	103	100 103			2.04
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	98.6 99.6			0.884
SM 2120 B	Color (true)				2.12	
SM 2320 B-97	Total Alkalinity	101			0.342	
SM 2540 C-97	TDS				2.63	
SM 4500 F-C-97	Fluoride	96.5				0.513
SM 5310 B-00	Organic Carbon	97.5	96.0 96.4			0.371

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1766864
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1766864
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1766864
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1766865
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1766865
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1766865
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1766865
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1766866
SM 2120 B / E31780	True Color measured at 450 nm	1766864
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1766864
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1766864
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1766864
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1766864
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1766865

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	01/29/2016			01/29/2016 13:56	Sima Feizi	1766864
EPA 300.0 Rev. 2.1	01/29/2016			02/05/2016	Ping Hua	1766864
EPA 350.1 Rev. 2.0	01/29/2016			02/04/2016	Diana M. Turner	1766865
EPA 351.2 Rev. 2.0	01/29/2016	02/04/2016	Christopher Armour	02/05/2016	Bryan A. Werth	1766865
EPA 353.2 Rev. 2.0	01/29/2016			02/02/2016	Peter D. Kaus	1766865
EPA 365.1 Rev. 2.0	01/29/2016	02/04/2016	Christopher Armour	02/05/2016	Lipi Saha	1766865
EPA 365.1 Rev. 2.0 dissolved	01/29/2016			01/29/2016 14:21	Julia Storbeck	1766866
SM 2120 B	01/29/2016			01/29/2016 12:06	Rochelle Y Jackson	1766864
SM 2320 B-97	01/29/2016			02/02/2016	Dale L. Simmons	1766864
SM 2540 C-97	01/29/2016			02/03/2016	Yijie Li	1766864
SM 2540 D-97	01/29/2016			02/03/2016	Yijie Li	1766864
SM 4500 F-C-97	01/29/2016			02/02/2016	Dale L. Simmons	1766864
SM 5310 B-00	01/29/2016			02/04/2016	Sofia Elnemr	1766865