

Chemical Analysis Report

CEN-DIST-2016-04-07-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Blue Cove/Lil Jones**

Request ID: **RQ-2016-04-04-67**

Customer: **CEN-DIST**

Project ID: **SMP**

Send Reports to:

FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact

Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 04-MAY-2016 11:24



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Blue Cove @ 150m NE of Center

Collection Date/Time: 04/06/2016 11:51

Field ID: G4CE0021

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1787529	EPA 180.1 Rev. 2.0	Turbidity	3.6		NTU	P301858	
	EPA 300.0 Rev. 2.1	Chloride	2.7		mg Cl/L	P302028	
		Sulfate	0.22	I	mg SO4/L	P302026	
	SM 2120 B	Color (true)	9.0		PCU	P302067	
	SM 2320 B-97	Total Alkalinity	38		mg CaCO3/L	P302191	
	SM 2540 C-97	TDS	50		mg/L	P302607	
	SM 2540 D-97	TSS	4	I	mg/L	P302327	
	SM 4500 F-C-97	Fluoride	0.28		mg F/L	P302194	
1787531	EPA 350.1 Rev. 2.0	Ammonia-N	0.006		mg N/L	P302272	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.96		mg N/L	P302675	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P302089	
	EPA 365.1 Rev. 2.0	Total-P	0.15		mg P/L	P302352	
	SM 5310 B-00	Organic Carbon	6.9		mg C/L	P302248	
1787533	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.068		mg P/L	P301831	

Ref. Method and Comment:

SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

SM 5310 B-00: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Sample Location: Little Jones Creek @ DS of Loop Bridge

Collection Date/Time: 04/06/2016 13:12

Field ID: G4CE0024

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1787530	EPA 180.1 Rev. 2.0	Turbidity	0.70		NTU	P301858	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P302028	
		Sulfate	14		mg SO4/L	P302026	
	SM 2120 B	Color (true)	51		PCU	P302067	
	SM 2320 B-97	Total Alkalinity	159		mg CaCO3/L	P302191	
	SM 2540 C-97	TDS	225	A	mg/L	P302607	
	SM 2540 D-97	TSS	2	U	mg/L	P302327	
	SM 4500 F-C-97	Fluoride	0.15		mg F/L	P302194	
1787532	EPA 350.1 Rev. 2.0	Ammonia-N	0.020		mg N/L	P302272	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.47		mg N/L	P302675	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.75		mg N/L	P302089	
	EPA 365.1 Rev. 2.0	Total-P	0.057		mg P/L	P302352	
	SM 5310 B-00	Organic Carbon	6.1		mg C/L	P302248	
1787534	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.046		mg P/L	P301831	

Ref. Method and Comment:

SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

SM 5310 B-00: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P301858

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302026

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302028

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P302272

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P302675

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P302089

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P302352

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P301831

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P302067

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P302191

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P302607

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P302327

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P302194

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P302248

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302026

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302028

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P302272

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P302675

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P302675

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.4		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P302089

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P302352

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P301831

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	103		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P302191

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	95 - 105

Reference Method: SM 4500 F-C-97
Batch ID: P302194

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.7		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P302248

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302026

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787480	Sulfate	93.8		P	90 - 110
1787486	Sulfate	96.4		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302028

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787480	Chloride	98.2		P	90 - 110
1787486	Chloride	96.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P302272

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787508	Ammonia-N	99.8	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P302675

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787503	Kjeldahl Nitrogen	88.2	94.6	F/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P302089

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787507	NO2NO3-N	97.6	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P302352

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787504	Total-P	103	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P301831

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787516	O-Phosphate-P	99.9	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P302194

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1787570	Fluoride	96.9	96.9	P/P	91 - 105

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P301858

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787497	Turbidity	0.477	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302026

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787480	Sulfate	0.410	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P302028

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787480	Chloride	0.694	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P302272

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787508	Ammonia-N	0.608	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P302675

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787503	Kjeldahl Nitrogen	5.71	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P302089

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787507	NO2NO3-N	2.06	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P302352

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787504	Total-P	2.61	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P301831

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787516	O-Phosphate-P	1.19	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P302067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787499	Color (true)	0.972	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P302191

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787570	Total Alkalinity	0.214	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P302607

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787530	TDS	0.889	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P302194

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787570	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P302248

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1787457	Organic Carbon	0.342	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A68729
Included Lab Sample IDs: 1787533, 1787534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.5	97.6	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A68738
Included Lab Sample IDs: 1787529, 1787530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	99.0	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A68804
Included Lab Sample IDs: 1787529, 1787530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	99.9	96.4	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A68821

Included Lab Sample IDs: 1787529, 1787530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	106	105	P/P	85 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A68825

Included Lab Sample IDs: 1787531, 1787532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	98.5	100	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A68859

Included Lab Sample IDs: 1787529, 1787530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	99.6	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A68859

Included Lab Sample IDs: 1787529, 1787530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A68878

Included Lab Sample IDs: 1787531, 1787532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.7	96.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A68885

Included Lab Sample IDs: 1787531, 1787532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A68913

Included Lab Sample IDs: 1787531, 1787532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A69032

Included Lab Sample IDs: 1787531, 1787532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	100	99.3	P/P	90 - 110
Kjeldahl Nitrogen	105	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision		MS
					LCS	SMP	
EPA 180.1 Rev. 2.0	Turbidity					0.477	
EPA 300.0 Rev. 2.1	Chloride	98.7	98.2	96.9		0.694	
	Sulfate	97.2	93.8	96.4		0.410	
EPA 350.1 Rev. 2.0	Ammonia-N	102	99.8	100			0.608
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.4	94.6	88.2			5.71
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	97.6	100			2.06
EPA 365.1 Rev. 2.0	Total-P	97.4	103	106			2.61
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	103	99.9	101			1.19
SM 2120 B	Color (true)					0.972	
SM 2320 B-97	Total Alkalinity	101				0.214	
SM 2540 C-97	TDS					0.889	
SM 4500 F-C-97	Fluoride	98.7	96.9	96.9			0.0
SM 5310 B-00	Organic Carbon	96.0					0.342

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1787529, 1787530
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1787529, 1787530
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1787529, 1787530
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1787531, 1787532
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1787531, 1787532
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1787531, 1787532
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1787531, 1787532
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1787533, 1787534
SM 2120 B / E31780	True Color measured at 450 nm	1787529, 1787530
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1787529, 1787530
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1787529, 1787530
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1787529, 1787530
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1787529, 1787530
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1787531, 1787532

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	04/07/2016			04/07/2016 17:11	Sima Feizi	1787529, 1787530
EPA 300.0 Rev. 2.1	04/07/2016			04/12/2016	Ping Hua	1787529, 1787530
EPA 350.1 Rev. 2.0	04/07/2016			04/14/2016	Diana M. Turner	1787531, 1787532
EPA 351.2 Rev. 2.0	04/07/2016	04/21/2016	Sofia Elnemr	04/22/2016	Rochelle Y Jackson	1787531, 1787532
EPA 353.2 Rev. 2.0	04/07/2016			04/12/2016	Peter D. Kaus	1787531, 1787532
EPA 365.1 Rev. 2.0	04/07/2016	04/15/2016	Lipi Saha	04/15/2016	Lipi Saha	1787531, 1787532
EPA 365.1 Rev. 2.0 dissolved	04/07/2016			04/07/2016 14:05	Ping Hua	1787533, 1787534
SM 2120 B	04/07/2016			04/07/2016 15:02	Blanca Fach	1787529
	04/07/2016			04/07/2016 15:03	Blanca Fach	1787530
SM 2320 B-97	04/07/2016			04/13/2016	Dale L. Simmons	1787529, 1787530
SM 2540 C-97	04/07/2016			04/11/2016	Sima Feizi	1787529, 1787530
SM 2540 D-97	04/07/2016			04/11/2016	Sima Feizi	1787529, 1787530
SM 4500 F-C-97	04/07/2016			04/13/2016	Dale L. Simmons	1787529, 1787530
SM 5310 B-00	04/07/2016			04/13/2016	Sofia Elnemr	1787531, 1787532