

Chemical Analysis Report

CEN-DIST-2016-07-26-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Lake Charles**
Request ID: **RQ-2016-07-25-32**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Timothy Fitzpatrick, Program Administrator

Date Certified: 17-AUG-2016 10:06

A handwritten signature in black ink, appearing to read 'Timothy W. Fitzpatrick', is written over a faint, light-colored rectangular stamp.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: LAKE CHARLES @ 400M S. OF CENTER

Collection Date/Time: 07/25/2016 12:35

Field ID: G1CE0030

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1818842	EPA 180.1 Rev. 2.0	Turbidity	2.7		NTU	P308648	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P309263	
		Sulfate	3.3		mg SO4/L	P309266	
	SM 2120 B	Color (true)	610		PCU	P308650	
	SM 2320 B-97	Total Alkalinity	23		mg CaCO3/L	P308881	
	SM 2540 C-97	TDS	84		mg/L	P309040	
	SM 2540 D-97	TSS	6	I	mg/L	P309026	
	SM 4500 F-C-97	Fluoride	0.11		mg F/L	P308884	
1818843	EPA 350.1 Rev. 2.0	Ammonia-N	0.017		mg N/L	P308934	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.7		mg N/L	P309017	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.045		mg N/L	P309208	MS
	EPA 365.1 Rev. 2.0	Total-P	0.063		mg P/L	P309146	
	SM 5310 B-00	Organic Carbon	45		mg C/L	P309107	
1818844	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.030		mg P/L	P308633	

Ref. Method and Comment:

EPA 365.1 Rev. 2.0: Total-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P308648

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P309263

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P309266

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P308934

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P309017

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309208

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309146

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P308633

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P308650

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P308881

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P309040

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P309026

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P308884

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P309107

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P309263

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P309266

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	95.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P308934

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P309017

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	109		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P309208

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P309146

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P308633

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2320 B-97
 Batch ID: P308881

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	95 - 108

Reference Method: SM 4500 F-C-97
 Batch ID: P308884

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.8		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P309107

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	102		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P309263

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818562	Chloride	97.8		P	90 - 110
1818618	Chloride	98.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P309266

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818562	Sulfate	94.3		P	90 - 110
1818618	Sulfate	94.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P308934

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818721	Ammonia-N	99.1	99.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P309017

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1817997	Kjeldahl Nitrogen	101	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P309208

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818823	NO2NO3-N	88.5	88.5	F/F	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P308633

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818865	O-Phosphate-P	98.0	98.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P308884

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818863	Fluoride	98.1	97.6	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P309107

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1818479	Organic Carbon	95.9	96.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P308648

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818822	Turbidity	0.692	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309263

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818562	Chloride	0.0540	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309266

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818562	Sulfate	0.922	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308934

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818721	Ammonia-N	0.0585	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309017

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1817997	Kjeldahl Nitrogen	0.309	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309208

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818823	NO2NO3-N	0.0	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309146

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818772	Total-P	0.498	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P308633

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818865	O-Phosphate-P	0.305	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P308650

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818873	Color (true)	7.52	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P308881

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818863	Total Alkalinity	0.0457	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P309040

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818644	TDS	3.90	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P309026

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818644	TSS	7.41	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P308884

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818863	Fluoride	0.482	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P309107

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1818479	Organic Carbon	0.222	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A70763

Included Lab Sample IDs: 1818844

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.0	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A70769

Included Lab Sample IDs: 1818842

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	99.5	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A70770

Included Lab Sample IDs: 1818842

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.4	99.6	P/P	90 - 115

Reference Method: SM 2320 B-97

Run ID: A70836

Included Lab Sample IDs: 1818842

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A70836

Included Lab Sample IDs: 1818842

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A70861

Included Lab Sample IDs: 1818843

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	99.7	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A70885

Included Lab Sample IDs: 1818842

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	99.9	P/P	90 - 110
Sulfate	98.6	96.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00
Run ID: A70922
Included Lab Sample IDs: 1818843

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.8	98.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A70949
Included Lab Sample IDs: 1818843

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A70977
Included Lab Sample IDs: 1818843

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.4	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A71029
Included Lab Sample IDs: 1818843

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.2	93.4	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.692	
EPA 300.0 Rev. 2.1	Chloride	98.3	97.8 98.0		0.0540	
	Sulfate	95.0	94.3 94.5		0.922	
EPA 350.1 Rev. 2.0	Ammonia-N	102	99.1 99.1			0.0585
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	109	101 100			0.309
EPA 353.2 Rev. 2.0	NO2NO3-N	102	88.5 88.5			0.0
EPA 365.1 Rev. 2.0	Total-P	101				0.498
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	98.0 98.4			0.305
SM 2120 B	Color (true)				7.52	
SM 2320 B-97	Total Alkalinity	104			0.0457	
SM 2540 C-97	TDS				3.90	
SM 2540 D-97	TSS				7.41	
SM 4500 F-C-97	Fluoride	97.8	98.1 97.6			0.482
SM 5310 B-00	Organic Carbon	102	95.9 96.2			0.222

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1818842
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1818842
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1818842
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1818843
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1818843
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1818843
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1818843
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1818844
SM 2120 B / E31780	True Color measured at 450 nm	1818842
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1818842
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1818842
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1818842
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1818842
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1818843

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/26/2016			07/26/2016 13:07	Sima Feizi	1818842
EPA 300.0 Rev. 2.1	07/26/2016			08/04/2016	Ping Hua	1818842
EPA 350.1 Rev. 2.0	07/26/2016			07/29/2016	Julie Michelle Frank	1818843
EPA 351.2 Rev. 2.0	07/26/2016	08/02/2016	Sofia Elnemr	08/04/2016	Christopher Armour	1818843
EPA 353.2 Rev. 2.0	07/26/2016			08/04/2016	Beverly Y. Sanders	1818843
EPA 365.1 Rev. 2.0	07/26/2016	08/04/2016	Vijayalakshmi Reddy	08/04/2016	Lipi Saha	1818843
EPA 365.1 Rev. 2.0 dissolved	07/26/2016			07/26/2016 14:33	Julia Storbeck	1818844
SM 2120 B	07/26/2016			07/26/2016 16:01	Rochelle Y Jackson	1818842
SM 2320 B-97	07/26/2016			07/29/2016	Dale L. Simmons	1818842
SM 2540 C-97	07/26/2016			07/27/2016	Yijie Li	1818842
SM 2540 D-97	07/26/2016			07/27/2016	Yijie Li	1818842
SM 4500 F-C-97	07/26/2016			07/29/2016	Dale L. Simmons	1818842
SM 5310 B-00	07/26/2016			08/02/2016	Sofia Elnemr	1818843