

Chemical Analysis Report

CEN-DIST-2016-09-13-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Holly Hill+HA**
Request ID: **RQ-2016-09-12-83**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 07-OCT-2016 14:18



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Holly Hill Ditch @ Alta Road

Collection Date/Time: 09/12/2016 10:45

Field ID: G5CE0028

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1830634	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P311085	
	EPA 300.0 Rev. 2.1	Chloride	58		mg Cl/L	P311403	
		Sulfate	27		mg SO4/L	P311409	
	SM 2120 B	Color (true)	45	A	PCU	P311036	
	SM 2320 B-97	Total Alkalinity	231		mg CaCO3/L	P311383	
	SM 2540 C-97	TDS	391	A	mg/L	P311522	
	SM 2540 D-97	TSS	5	I	mg/L	P311502	
	SM 4500 F-C-97	Fluoride	0.10		mg F/L	P311387	
1830635	EPA 350.1 Rev. 2.0	Ammonia-N	0.38		mg N/L	P311262	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.97		mg N/L	P311199	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.20		mg N/L	P311592	
	EPA 365.1 Rev. 2.0	Total-P	0.15		mg P/L	P311498	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P311341	
1830636	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.12		mg P/L	P311065	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P311085

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P311403

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P311409

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P311262

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P311199

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P311592

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P311498

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P311065

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P311036

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P311383

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P311522

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P311502

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P311387

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P311341

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P311403

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P311409

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P311262

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.5		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P311199

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P311592

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P311498

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P311065

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P311383

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P311387

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.5		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P311341

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	104		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P311403

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830659	Chloride	98.5		P	90 - 110
1830660	Chloride	98.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P311409

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830659	Sulfate	99.2		P	90 - 110
1830660	Sulfate	98.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P311262

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830472	Ammonia-N	99.3	96.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P311592

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830584	NO2NO3-N	105	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P311498

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830635	Total-P	97.2	98.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P311065

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830504	O-Phosphate-P	99.4	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P311387

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1831043	Fluoride	104	103	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P311341

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1830573	Organic Carbon	102	103	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P311085

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830534	Turbidity	0.514	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P311403

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830659	Chloride	0.193	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P311409

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830659	Sulfate	0.265	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P311262

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830472	Ammonia-N	3.16	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P311199

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830585	Kjeldahl Nitrogen	0.776	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P311592

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830584	NO2NO3-N	2.01	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P311498

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830635	Total-P	0.844	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P311065

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830504	O-Phosphate-P	0.702	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P311036

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830634	Color (true)	4.77	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P311383

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1831043	Total Alkalinity	1.44	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P311522

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830634	TDS	2.56	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P311387

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1831043	Fluoride	0.487	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P311341

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1830573	Organic Carbon	0.480	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A71576

Included Lab Sample IDs: 1830634

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	101	97.4	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A71581

Included Lab Sample IDs: 1830634

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	94.6	95.0	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A71595

Included Lab Sample IDs: 1830636

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	99.3	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A71599

Included Lab Sample IDs: 1830634

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	100	P/P	95 - 105

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A71658

Included Lab Sample IDs: 1830635

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.0	98.7	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A71682

Included Lab Sample IDs: 1830635

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	106	107	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A71690

Included Lab Sample IDs: 1830635

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	99.3	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A71696

Included Lab Sample IDs: 1830634

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97
 Run ID: A71696
 Included Lab Sample IDs: 1830634

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Run ID: A71734
 Included Lab Sample IDs: 1830635

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	103	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Run ID: A71755
 Included Lab Sample IDs: 1830635

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.9	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
 Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
 Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.514	
EPA 300.0 Rev. 2.1	Chloride	100	98.5 98.5		0.193	
	Sulfate	98.6	99.2 98.1		0.265	
EPA 350.1 Rev. 2.0	Ammonia-N	97.5	99.3 96.0			3.16
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.0				0.776
EPA 353.2 Rev. 2.0	NO2NO3-N	106	105 102			2.01
EPA 365.1 Rev. 2.0	Total-P	100	97.2 98.4			0.844
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	99.4 100			0.702
SM 2120 B	Color (true)				4.77	
SM 2320 B-97	Total Alkalinity	103			1.44	
SM 2540 C-97	TDS				2.56	
SM 4500 F-C-97	Fluoride	98.5	104 103			0.487
SM 5310 B-00	Organic Carbon	104	102 103			0.480

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1830634
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1830634
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1830634
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1830635

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1830635
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1830635
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1830635
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1830636
SM 2120 B / E31780	True Color measured at 450 nm	1830634
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1830634
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1830634
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1830634
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1830634
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1830635

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/13/2016			09/13/2016 16:45	Sima Feizi	1830634
EPA 300.0 Rev. 2.1	09/13/2016			09/17/2016	Ping Hua	1830634
EPA 350.1 Rev. 2.0	09/13/2016			09/14/2016	Joseph Suarez	1830635
EPA 351.2 Rev. 2.0	09/13/2016	09/15/2016	Adrian Shelby	09/16/2016	Virginia P. Brown	1830635
EPA 353.2 Rev. 2.0	09/13/2016			09/20/2016	Beverly Y. Sanders	1830635
EPA 365.1 Rev. 2.0	09/13/2016	09/20/2016	Lipi Saha	09/21/2016	Lipi Saha	1830635
EPA 365.1 Rev. 2.0 dissolved	09/13/2016			09/13/2016 18:47	Julia Storbeck	1830636
SM 2120 B	09/13/2016			09/13/2016 15:12	Julie Michelle Frank	1830634
SM 2320 B-97	09/13/2016			09/17/2016	Dale L. Simmons	1830634
SM 2540 C-97	09/13/2016			09/15/2016	Yijie Li	1830634
SM 2540 D-97	09/13/2016			09/15/2016	Yijie Li	1830634
SM 4500 F-C-97	09/13/2016			09/17/2016	Dale L. Simmons	1830634
SM 5310 B-00	09/13/2016			09/16/2016	Sofia Elnemr	1830635