

Chemical Analysis Report

CEN-DIST-2016-11-17-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Sweetwater Creek+SCI**
Request ID: **RQ-2016-11-07-67**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 14-DEC-2016 13:22



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: GROOVER BRANCH @ AIRPORT RD. BRIDGE Collection Date/Time: 11/16/2016 14:06

Field ID: G5CE0040

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1850923	EPA 180.1 Rev. 2.0	Turbidity	3.1	A	NTU	P315416	
	EPA 300.0 Rev. 2.1	Chloride	70		mg Cl/L	P316064	
		Sulfate	22		mg SO4/L	P316067	
	SM 2120 B	Color (true)	75	A	PCU	P315394	
	SM 2320 B-97	Total Alkalinity	93		mg CaCO3/L	P315673	
	SM 2540 C-97	TDS	244		mg/L	P316003	
	SM 2540 D-97	TSS	6	I	mg/L	P315755	
	SM 4500 F-C-97	Fluoride	0.064	I	mg F/L	P315676	
1850924	EPA 350.1 Rev. 2.0	Ammonia-N	0.098		mg N/L	P315786	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.79		mg N/L	P316206	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.072		mg N/L	P315841	
	EPA 365.1 Rev. 2.0	Total-P	0.077		mg P/L	P315899	
	SM 5310 B-00	Organic Carbon	18		mg C/L	P315408	
1850925	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.029		mg P/L	P315387	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P315416

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P316064

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P316067

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P315786

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P316206

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P315841

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315899

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315387

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P315394

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P315673

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P316003

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P315755

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P315676

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P315408

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316064

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316067

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315786

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316206

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P315841

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315899

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315387

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	104		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P315673

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P315676

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.3		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P315408

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	103		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P316064

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851195	Chloride	101		P	90 - 110
1851196	Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P316067

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851195	Sulfate	97.1		P	90 - 110
1851196	Sulfate	96.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P315786

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850629	Ammonia-N	97.6	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P316206

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851183	Kjeldahl Nitrogen	100	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P315841

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850630	NO2NO3-N	103	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P315899

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850861	Total-P	97.7	98.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P315387

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850862	O-Phosphate-P	99.0	99.7	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P315676

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851136	Fluoride	100	101	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P315408

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850840	Organic Carbon	97.2	99.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P315416

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850923	Turbidity	0.969	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316064

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851195	Chloride	0.112	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851195	Sulfate	1.53	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315786

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850629	Ammonia-N	2.17	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316206

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851183	Kjeldahl Nitrogen	0.860	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P315841

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P315841

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850630	NO2NO3-N	1.41	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315899

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850861	Total-P	0.387	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315387

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850862	O-Phosphate-P	0.621	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P315394

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850923	Color (true)	0.704	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P315673

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851136	Total Alkalinity	0.452	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P316003

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850879	TDS	2.14	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P315676

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851136	Fluoride	0.494	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P315408

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850840	Organic Carbon	2.45	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A72868

Included Lab Sample IDs: 1850925

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	101	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A72873

Included Lab Sample IDs: 1850923

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.6	100	P/P	85 - 115

Reference Method: SM 5310 B-00

Run ID: A72882

Included Lab Sample IDs: 1850924

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	102	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A72884

Included Lab Sample IDs: 1850923

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	99.5	P/P	95 - 105

Reference Method: SM 2320 B-97

Run ID: A72982

Included Lab Sample IDs: 1850923

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A72982

Included Lab Sample IDs: 1850923

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A73030

Included Lab Sample IDs: 1850924

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A73050

Included Lab Sample IDs: 1850924

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A73099

Included Lab Sample IDs: 1850924

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	103	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A73131

Included Lab Sample IDs: 1850923

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	100	P/P	90 - 110
Sulfate	98.3	95.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73229

Included Lab Sample IDs: 1850924

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	92.3	94.6	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.969	
EPA 300.0 Rev. 2.1	Chloride	101	101	101	0.112	
	Sulfate	96.4	97.1	96.7	1.53	
EPA 350.1 Rev. 2.0	Ammonia-N	103	97.6	104		2.17
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	100	100	101		0.860
EPA 353.2 Rev. 2.0	NO2NO3-N	103	103	104		1.41
EPA 365.1 Rev. 2.0	Total-P	98.8	97.7	98.1		0.387
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	104	99.0	99.7		0.621
SM 2120 B	Color (true)				0.704	
SM 2320 B-97	Total Alkalinity	102			0.452	
SM 2540 C-97	TDS				2.14	
SM 4500 F-C-97	Fluoride	97.3	100	101		0.494
SM 5310 B-00	Organic Carbon	103	97.2	99.7		2.45

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1850923
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1850923
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1850923
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1850924
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1850924
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1850924
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1850924
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1850925
SM 2120 B / E31780	True Color measured at 450 nm	1850923
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1850923
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1850923
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1850923
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1850923
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1850924

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/17/2016			11/17/2016 16:41	Sima Feizi	1850923
EPA 300.0 Rev. 2.1	11/17/2016			12/01/2016	Ping Hua	1850923
EPA 350.1 Rev. 2.0	11/17/2016			11/28/2016	Sofia Elnemr	1850924
EPA 351.2 Rev. 2.0	11/17/2016	12/06/2016	Martin Acosta	12/07/2016	Joseph Suarez	1850924
EPA 353.2 Rev. 2.0	11/17/2016			11/29/2016	Beverly Y. Sanders	1850924
EPA 365.1 Rev. 2.0	11/17/2016	11/30/2016	Adrian Shelby	11/30/2016	Lipi Saha	1850924
EPA 365.1 Rev. 2.0 dissolved	11/17/2016			11/17/2016 15:29	Julia Storbeck	1850925
SM 2120 B	11/17/2016			11/17/2016 15:57	Jared I. Manley	1850923
SM 2320 B-97	11/17/2016			11/23/2016	Dale L. Simmons	1850923
SM 2540 C-97	11/17/2016			11/21/2016	Sima Feizi	1850923
SM 2540 D-97	11/17/2016			11/21/2016	Sima Feizi	1850923
SM 4500 F-C-97	11/17/2016			11/23/2016	Dale L. Simmons	1850923
SM 5310 B-00	11/17/2016			11/17/2016	Julie Michelle Frank	1850924