

Chemical Analysis Report

CEN-DIST-2016-11-18-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Glencoe Ditches**
Request ID: **RQ-2016-11-14-62**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Environmental Administrator

Date Certified: 15-DEC-2016 14:35

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Glencoe Ditches @ 225m South of Turnbull Bay Rd

Collection Date/Time: 11/17/2016 10:50

Field ID: G5CE0077

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851457	EPA 180.1 Rev. 2.0	Turbidity	11		NTU	P315470	
	EPA 300.0 Rev. 2.1	Chloride	100		mg Cl/L	P316065	
		Sulfate	7.4		mg SO4/L	P316134	
	SM 2120 B	Color (true)	17	A	PCU	P315464	
	SM 2320 B-97	Total Alkalinity	249		mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	500		mg/L	P315924	
	SM 2540 D-97	TSS	5	I	mg/L	P315927	
	SM 4500 F-C-97	Fluoride	0.059	I	mg F/L	P316309	
1851458	EPA 350.1 Rev. 2.0	Ammonia-N	0.17		mg N/L	P315790	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.53		mg N/L	P316207	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.025		mg N/L	P316013	
	EPA 365.1 Rev. 2.0	Total-P	0.11		mg P/L	P315973	
	SM 5310 B-00	Organic Carbon	7.7		mg C/L	P315485	
1851459	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.026		mg P/L	P315473	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P315470

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P316065

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P316134

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P315790

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P316207

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316013

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315973

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315473

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P315464

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P315809

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P315924

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P315927

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P315485

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316065

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316134

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315790

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316207

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.1		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316013

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315973

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315473

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P315809

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.8		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P315485

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.7		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316065

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1849598	Chloride	100		P	90 - 110
1851137	Chloride	95.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316134

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854429	Sulfate	107		P	90 - 110
1854432	Sulfate	104		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315790

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851139	Ammonia-N	99.2	97.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316207

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851511	Kjeldahl Nitrogen	101	97.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316013

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851458	NO2NO3-N	98.8	98.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315973

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851394	Total-P	105	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315473

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851516	O-Phosphate-P	103	103	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1852055	Fluoride	96.1	97.8	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P315485

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851394	Organic Carbon	92.4	93.9	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P315470

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851427	Turbidity	0.441	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316065

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849598	Chloride	0.0783	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316134

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854432	Sulfate	0.466	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315790

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851139	Ammonia-N	1.94	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316207

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851511	Kjeldahl Nitrogen	3.82	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316013

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316013

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851458	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315973

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851394	Total-P	0.258	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315473

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851516	O-Phosphate-P	0.583	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P315464

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851457	Color (true)	2.56	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P315809

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852055	Total Alkalinity	0.568	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P315924

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851427	TDS	6.38	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852055	Fluoride	1.55	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P315485

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851394	Organic Carbon	1.57	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A72897
Included Lab Sample IDs: 1851457

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	100	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A72900
Included Lab Sample IDs: 1851457

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.0	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A72902
Included Lab Sample IDs: 1851459

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.0	96.7	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A72907
Included Lab Sample IDs: 1851458

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.6	96.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A73030
Included Lab Sample IDs: 1851458

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A73035
Included Lab Sample IDs: 1851457

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A73111
Included Lab Sample IDs: 1851458

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	98.0	96.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A73131

Included Lab Sample IDs: 1851457

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	98.9	P/P	90 - 110
Sulfate	105	108	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A73161

Included Lab Sample IDs: 1851458

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	104	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A73217

Included Lab Sample IDs: 1851457

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73229

Included Lab Sample IDs: 1851458

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	95.5	92.4	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					0.441	
EPA 300.0 Rev. 2.1	Chloride	101	100	95.4		0.0783	
	Sulfate	101	104	107		0.466	
EPA 350.1 Rev. 2.0	Ammonia-N	101	99.2	97.2			1.94
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.1	101	97.3			3.82
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.0	98.8	98.8			0.0
EPA 365.1 Rev. 2.0	Total-P	105	105	105			0.258
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	103	103			0.583
SM 2120 B	Color (true)					2.56	
SM 2320 B-97	Total Alkalinity	103				0.568	
SM 2540 C-97	TDS					6.38	
SM 4500 F-C-97	Fluoride	98.8	96.1	97.8			1.55
SM 5310 B-00	Organic Carbon	97.7	92.4	93.9			1.57

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1851457
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1851457
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1851457
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1851458
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1851458
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1851458
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1851458
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1851459
SM 2120 B / E31780	True Color measured at 450 nm	1851457
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1851457
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1851457
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1851457
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1851457
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1851458

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/18/2016			11/18/2016 14:00	Sima Feizi	1851457
EPA 300.0 Rev. 2.1	11/18/2016			12/02/2016	Ping Hua	1851457
	11/18/2016			12/03/2016	Ping Hua	1851457
EPA 350.1 Rev. 2.0	11/18/2016			11/28/2016	Sofia Elnemr	1851458
EPA 351.2 Rev. 2.0	11/18/2016	12/06/2016	Martin Acosta	12/07/2016	Joseph Suarez	1851458
EPA 353.2 Rev. 2.0	11/18/2016			12/01/2016	Beverly Y. Sanders	1851458
EPA 365.1 Rev. 2.0	11/18/2016	12/01/2016	Christopher Armour	12/05/2016	Lipi Saha	1851458
EPA 365.1 Rev. 2.0 dissolved	11/18/2016			11/18/2016 16:23	Julia Storbeck	1851459
SM 2120 B	11/18/2016			11/18/2016 14:21	Jared I. Manley	1851457
SM 2320 B-97	11/18/2016			11/28/2016	Dale L. Simmons	1851457
SM 2540 C-97	11/18/2016			11/22/2016	Yijie Li	1851457
SM 2540 D-97	11/18/2016			11/22/2016	Yijie Li	1851457
SM 4500 F-C-97	11/18/2016			12/07/2016	Dale L. Simmons	1851457
SM 5310 B-00	11/18/2016			11/18/2016	Julie Michelle Frank	1851458