

Chemical Analysis Report

CEN-DIST-2016-12-28-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Weohykapka Creek+SCI**
Request ID: **RQ-2016-12-12-25**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 11-JAN-2017 11:02

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: WEOHYAKAPKA CREEK @ 800M N OF 60

Collection Date/Time: 12/27/2016 13:49

Field ID: G4CE0108

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1860246	EPA 180.1 Rev. 2.0	Turbidity	5.0		NTU	P317504	
	EPA 300.0 Rev. 2.1	Chloride	13		mg Cl/L	P317849	
		Sulfate	12		mg SO4/L	P317846	
	SM 2120 B	Color (true)	51		PCU	P317508	
	SM 2320 B-97	Total Alkalinity	15		mg CaCO3/L	P317538	
	SM 2540 C-97	TDS	84		mg/L	P317970	
	SM 2540 D-97	TSS	8	I	mg/L	P317940	
	SM 4500 F-C-97	Fluoride	0.064	I	mg F/L	P317539	
1860247	EPA 350.1 Rev. 2.0	Ammonia-N	0.020		mg N/L	P317751	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.63		mg N/L	P317515	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.044		mg N/L	P317609	
	EPA 365.1 Rev. 2.0	Total-P	0.033		mg P/L	P317522	
	SM 5310 B-00	Organic Carbon	9.8		mg C/L	P317612	
1860248	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	I	mg P/L	P317507	

Ref. Method and Comment:

EPA 350.1 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P317504

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P317846

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P317849

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P317751

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P317515

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P317609

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P317522

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P317507

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P317508

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P317538

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P317970

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P317940

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P317539

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P317612

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P317846

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P317849

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P317751

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P317515

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	107		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P317609

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P317522

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	92.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P317507

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	103		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P317538

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	99.4		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P317539

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.0		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P317612

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P317846

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860246	Sulfate	95.4		P	90 - 110
1860746	Sulfate	96.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P317849

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860246	Chloride	99.3		P	90 - 110
1860746	Chloride	100		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P317515

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860217	Kjeldahl Nitrogen	105	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P317609

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1858677	NO2NO3-N	106	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P317522

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860247	Total-P	94.2	91.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P317507

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860243	O-Phosphate-P	94.5	93.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P317539

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860214	Fluoride	98.9	99.5	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P317612

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1860198	Organic Carbon	95.7	94.8	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P317504

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860194	Turbidity	10.7	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P317846

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860746	Sulfate	0.248	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P317849

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860746	Chloride	0.0155	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P317751

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860201	Ammonia-N	0.343	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P317515

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860217	Kjeldahl Nitrogen	1.27	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P317609

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1858677	NO2NO3-N	0.862	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P317522

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860247	Total-P	3.10	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P317507

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860243	O-Phosphate-P	0.673	Spike	P	0 - 20

Reference Method: SM 2120 B

Batch ID: P317508

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860194	Color (true)	0.190	Sample	P	0 - 20

Reference Method: SM 2320 B-97

Batch ID: P317538

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860214	Total Alkalinity	0.140	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97

Batch ID: P317970

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860197	TDS	2.38	Sample	P	0 - 10

Reference Method: SM 2540 D-97

Batch ID: P317940

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860197	TSS	6.67	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97

Batch ID: P317539

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860214	Fluoride	0.509	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00

Batch ID: P317612

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1860198	Organic Carbon	0.473	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A73621

Included Lab Sample IDs: 1860246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A73622

Included Lab Sample IDs: 1860248

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.7	98.9	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A73623

Included Lab Sample IDs: 1860246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.7	99.2	P/P	85 - 115

Reference Method: SM 2320 B-97

Run ID: A73634

Included Lab Sample IDs: 1860246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.9	99.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A73634

Included Lab Sample IDs: 1860246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.6	97.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73662

Included Lab Sample IDs: 1860247

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A73672

Included Lab Sample IDs: 1860247

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00
 Run ID: A73673
 Included Lab Sample IDs: 1860247

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.0	96.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Run ID: A73703
 Included Lab Sample IDs: 1860247

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	96.5	96.8	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Run ID: A73745
 Included Lab Sample IDs: 1860247

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	99.7	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Run ID: A73754
 Included Lab Sample IDs: 1860246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	95.9	99.1	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
 Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
 Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				10.7	
EPA 300.0 Rev. 2.1	Chloride	102	100	99.3	0.0155	
	Sulfate	97.2	96.8	95.4	0.248	
EPA 350.1 Rev. 2.0	Ammonia-N	102				0.343
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	107	105	104		1.27
EPA 353.2 Rev. 2.0	NO2NO3-N	106	106	105		0.862
EPA 365.1 Rev. 2.0	Total-P	92.4	94.2	91.0		3.10
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	103	94.5	93.8		0.673
SM 2120 B	Color (true)				0.190	
SM 2320 B-97	Total Alkalinity	99.4			0.140	
SM 2540 C-97	TDS				2.38	
SM 2540 D-97	TSS				6.67	
SM 4500 F-C-97	Fluoride	96.0	98.9	99.5		0.509
SM 5310 B-00	Organic Carbon	101	95.7	94.8		0.473

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1860246
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1860246
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1860246
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1860247
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1860247
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1860247
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1860247
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1860248
SM 2120 B / E31780	True Color measured at 450 nm	1860246
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1860246
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1860246
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1860246
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1860246
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1860247

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	12/28/2016			12/28/2016 14:58	Sima Feizi	1860246
EPA 300.0 Rev. 2.1	12/28/2016			01/05/2017	Ping Hua	1860246
EPA 350.1 Rev. 2.0	12/28/2016			01/04/2017	Sofia Elnemr	1860247
EPA 351.2 Rev. 2.0	12/28/2016	12/29/2016	Adrian Shelby	12/30/2016	Joseph Suarez	1860247
EPA 353.2 Rev. 2.0	12/28/2016			12/30/2016	Beverly Y. Sanders	1860247
EPA 365.1 Rev. 2.0	12/28/2016	12/29/2016	Adrian Shelby	01/03/2017	Lipi Saha	1860247
EPA 365.1 Rev. 2.0 dissolved	12/28/2016			12/28/2016 15:32	Lipi Saha	1860248
SM 2120 B	12/28/2016			12/28/2016 15:24	Julie Michelle Frank	1860246
SM 2320 B-97	12/28/2016			12/28/2016	Dale L. Simmons	1860246
SM 2540 C-97	12/28/2016			12/30/2016	Yijie Li	1860246
SM 2540 D-97	12/28/2016			12/30/2016	Yijie Li	1860246
SM 4500 F-C-97	12/28/2016			12/28/2016	Dale L. Simmons	1860246
SM 5310 B-00	12/28/2016			12/29/2016	Julie Michelle Frank	1860247