

Field Meter Calibration Log

Meter ID: YSI multi probe

	Date	2-24-16	2-25-16	2-26-16	3/3/16	3/15/16
AM	pH 4.00	4	4	4	4.00	4.00
	pH 7.00	7	7	7	7.00	7.00
	pH 10.00	10	10	10	10.00	10.00
	Cond 147					
	Cond 717.5					
	Cond 1413					
	@ Temp					
	DO %	99.6	100.3	100.3	100.3	100.2
	DO @ Temp	23.62	21.51	21.43	21.59	24.77
PM	pH	4/7/10	4.1/7.1/10.1	4.1/7.1/10.1	3.9/6.9/9.9	4/7/10
	Cond	—	—	—	—	—
	DO% @ Temp	99.2/23.17	100.4/14.41	100.4/22.57	100.6/26.21	100.3/31.60
	Batteries	50%	49%	49%	—	48%
	Comments				pH +/- Sample / temp in lab / water & rebo.	
	User	J.D.F.	J.D.F.	J.D.F.	S.F.	J.D.F.