

Field Meter Calibration Log

Meter ID:

YSI multiprobe

		Date	3-29-16	3-30-16	3-31-16	4-14-16	4-26-16
AM	pH 4.00		4	4	4	4	4
	pH 7.00		7	7	7	7	7
	pH 10.00		10	10	10	10	10
	Cond 147						
	Cond 717.5						
	Cond 1413						
	@ Temp						
	DO %		100.2	100.4	100.3	100.1	100.6
	DO @ Temp		25.41	23.78	24.69	24.61	24.77
PM	pH all		4/7/10	4/7/10	4/7/10	4/7/9.99	4/7/10
	Cond		—	—	—	—	—
	DO% @ Temp		100.3 / 26.48	100.4 / 25.54	100.3 / 25.81	100.1 / 26.00	99.9 / 25.96
	Batteries		48%	48%	48%	48%	46%
	Comments						
	User		J.D.F.	J.D.F.	J.D.F.	J.D.F.	J.D.F.