

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 7

Project: SMP: Lake Adelaide

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____
 Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 3/12/16

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kaitlin DeBeth	5/16/16	0710	23.6	8.48	8.48	100.0	75.5		☒ F ☐ F	☐ F ☒ F
ICV	"	"	0713	23.5	8.49	8.51	100.0	75.7		☒ F ☐ F	☐ F ☒ F
CCV	Corey Anderson	5/16/16	1612	30.2	7.5	7.6	101	96	0.9	☒ F ☐ F	☐ F ☒ F
CCV										☐ F ☒ F	☐ F ☒ F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.
 DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Optical: DO gain range 0.85 to 1.15; DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Kaitlin DeBeth	5/16/16	0715	160418A	4/17	1,000	1,000	☒ F ☐ F	☐ F ☒ F
ICV	"	"	0716	150915E	9/16	500	513	☒ F ☐ F	☐ F ☒ F
CCV	Corey Anderson	5/16/16	1616	160318	9/16	100	102	☒ F ☐ F	☐ F ☒ F
CCV								☐ F ☒ F	☐ F ☒ F

Report specific conductance as a whole number with no decimal figure. Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Kaitlin DeBeth	5/16/16	0718	161523	3/18	7.0	7.0	-11.5	☒ F ☐ F	☐ F ☒ F
Calibr.	"	"	719	160222A	9/17	4.0	4.0	162.6	☒ F ☐ F	☐ F ☒ F
Calibr.	"	"	720	150915B	3/17	10.0	10.0	-181.2	☒ F ☐ F	☐ F ☒ F
ICV	"	"	721	150915B	3/17	10.0	9.9	-182.5	☒ F ☐ F	☐ F ☒ F
CCV	Corey Anderson	5/16/16	1618	160222A	9/17	4.0	4.2	147	☒ F ☐ F	☐ F ☒ F
CCV									☐ F ☒ F	☐ F ☒ F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$;
 mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly) Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: RA-2016-05-16-13