

Field Meter Calibration Log

Meter ID:

YSI multi probe

		Date	4-22-16	5-24-16	5-26-16	6-13-16	6-21-16
AM	pH 4.00		4	4	4	4	4
	pH 7.00		7	7	7	7	7
	pH 10.00		10	10	10	10	10
	Cond 147						
	Cond 717.5						
	Cond 1413						
	@ Temp						
PM	DO %		100.0	100.2	100.3	100.2	100.6
	DO @ Temp		24.81	24.76	25.22	26.41	24.15
	pH		4/7/10	4/7/10	4/7/10	4.01/7/10	4/7.04/10.0
	Cond		—	—	—	—	—
	DO% @ Temp		$\frac{100.0}{26.75}$	$\frac{100.4}{27.52}$	$\frac{100.4}{27.50}$	$\frac{100.1}{26.40}$	$\frac{100.5}{25.41}$
	Batteries		46%	46%	46%	46%	45%
	Comments						
	User		J.D.F.	J.D.F.	S.F./J.D.F.	J.D.F.	J.D.F.