

# CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page

Meter ID:

5

Project:

Smp: Howard, Indian, Phlippi

Notes: (1) Numbers  $\leq 4$ , are rounded down; numbers  $\geq 5$  are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7-7-16

| DO DEP SOP FT 1500 | Name           | Date   | Time (ET) | Temp. (°C) | D.O. Chart mg/L | Meter D.O. mg/L | % DO | Probe Charge | Probe Gain | Pass / Fail | Lab / Field |
|--------------------|----------------|--------|-----------|------------|-----------------|-----------------|------|--------------|------------|-------------|-------------|
| Calibr.            | Corey Anderson | 8/9/16 | 735       | 23.2       | 8.5             | 8.6             | 100  | 45           |            | (P) F       | (L) F       |
| ICV                | "              | "      | 745       | 23.2       | 8.5             | 8.5             | 100  |              |            | (P) F       | (L) F       |
| CCV                | Corey Anderson | 8/9/16 | 1436      | 21.7       | 8.8             | 8.8             | 100  | 44           |            | (P) F       | (L) F       |
| CCV                |                |        |           |            |                 |                 |      |              |            | P F         | L F         |

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table  $\pm 0.3$  mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.  
Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

| Spec. Cond. FT 1200 | Name           | Date   | Time (ET) | Lot #   | Expir. Date | Standard $\mu$ mhos/cm | Meter Reading $\mu$ mhos/cm | Pass / Fail | Lab / Field |
|---------------------|----------------|--------|-----------|---------|-------------|------------------------|-----------------------------|-------------|-------------|
| Calibr.             | Corey Anderson | 8-9-16 | 751       | 160418A | 4/17        | 1000                   | 1000                        | (P) F       | (L) F       |
| ICV                 | "              | "      | 752       | 160318  | 9/16        | 100                    | 104                         | (P) F       | (L) F       |
| CCV                 | Corey Anderson | 8/9/16 | 1437      | 160418A | 4/17        | 1000                   | 1005                        | (P) F       | (L) F       |
| CCV                 |                |        |           |         |             |                        |                             | P F         | L F         |

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria  $\pm 5\%$

| pH DEP SOP FT 1100 | Name           | Date   | Time (ET) | Lot #   | Expir. Date | pH Buffer SU | Meter reading SU | mV   | Pass / Fail | Lab / Field |
|--------------------|----------------|--------|-----------|---------|-------------|--------------|------------------|------|-------------|-------------|
| Calibr.            | Corey Anderson | 8-9-16 | 755       | 160222A | 9/17        | 7.0          | 7.0              | -14  | (P) F       | (L) F       |
| Calibr.            | "              | "      | 756       | 161523  | 3/18        | 4.0          | 4.0              | 164  | (P) F       | (L) F       |
| Calibr.            | "              | "      | 757       | 160126  | 2/18        | 10.0         | 10.0             | -191 | (P) F       | (L) F       |
| ICV                | "              | "      | 800       | 160126  | 2/18        | 10.0         | 10.0             | -191 | (P) F       | (L) F       |
| CCV                | Corey Anderson | 8/9/16 | 1438      | 161523  | 3/18        | 4.0          | 4.0              | 162  | (P) F       | (L) F       |
| CCV                |                |        |           |         |             |              |                  |      | P F         | L F         |

Report pH with one decimal place; pH Acceptance criteria  $\pm 0.2$  SU;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

mV pH 4 Range +180 $\pm$ 50;

Depth (Quarterly)

Date of Last Depth Verification

| Depth Sensor (Daily) | Name | Date | Time CT-ET | Zero the Sensor | ICV Value | Pass / Fail | Lab / Field |
|----------------------|------|------|------------|-----------------|-----------|-------------|-------------|
| Pressure mode in air |      |      |            | 0.000           |           |             |             |

Notes:

RR-2016-07-25-07

ICV acceptance criteria  $\pm 5\%$  or  $\pm 0.05$ m, whichever is greater