

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: SMP B4

Date: 4/21/16

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	MM	4/21/16	0900			8.51	23.2	74.6	8.54	P
CAL	ICV	CCV	RA	4/21/16	1445			8.45	23.4		8.514	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Acceptance Criteria: $\pm 0.3 \text{ mg/L}$

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	--	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

Acceptance Criteria: $\pm 5\%$

CAL	ICV	CCV	NA	4/21/16	0900	1413	7/8/16	KCL 16-2			1413	P
CAL	ICV	CCV	RA	4/21/16	1445	1413	7/8/16	KCL 16-2			1420	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	--	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

Acceptance Criteria: $\pm 0.2 \text{ SU}$

CAL	ICV	CCV	MM	4/21/16	0900	7	6/16	4405608			7.0	P
CAL	ICV	CCV	RA	4/21/16	1445	7	6/16	4405608			7.0	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: