

Be sure this box is filled in if there is qualified data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: RQ-2016-06-06-65 Date: 6/7/16 Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	RA	6/7/16	0810			26.7	100	8.0	P
CAL	ICV	CCV	RA	6/8/16	0835			25.0	100	8.2	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	RA	6/7/16	0811	1413	2/13/17	16-4			1436-1413	P
CAL	ICV	CCV	RA	6/7/16	0812	100	11/17	251794			101	P
CAL	ICV	CCV	RA	6/8/16	0835	1413	2/13/17	16-4			1402	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	RA	6/7/16	0815	7	5/17	2506518			-25.7	P
CAL	ICV	CCV	RA	6/7/16	0815	4	5/17	2507749			152.8	P
CAL	ICV	CCV	RA	6/7/16	0819	10	0/16	2503279			-197.8	P
CAL	ICV	CCV	RA	6/7/16	0820	4	7/17	2507749			4.0	P
CAL	ICV	CCV	RA	6/8/16	0940	7	5/17	2506518			7.1	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Kawick Lake > 101
Bear Lake
Harrison Lake

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: RQ-2016-06-06-22 Date: 6/2/16 Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0720</u>			<u>8.1</u>	<u>25.7</u>	<u>99</u>	<u>8.158</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>1430</u>			<u>8.3</u>	<u>24.3</u>	<u>99</u>	<u>8.371</u>	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0722</u>	<u>12900</u>	<u>1/21/17</u>	<u>KCL 14-1</u>			<u>12900</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0723</u>	<u>1413</u>	<u>2/13/17</u>	<u>KCL 14-4</u>			<u>1414</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>1435</u>	<u>1200</u>	<u>11/17</u>	<u>2511997</u>			<u>104</u>	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0725</u>	<u>7</u>	<u>05/17</u>	<u>2506518</u>		<u>-49.9</u>	<u>7</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0726</u>	<u>4</u>	<u>07/17</u>	<u>2507749</u>		<u>137.5</u>	<u>4</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0727</u>	<u>10</u>	<u>07/16</u>	<u>2503774</u>		<u>-208.8</u>	<u>10</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>0730</u>	<u>4</u>	<u>07/16</u>	<u>2507749</u>			<u>4.0</u>	P
CAL	ICV	CCV	<u>me</u>	<u>6/2/16</u>	<u>1440</u>	<u>10</u>	<u>09/16</u>	<u>2503774</u>			<u>10.0</u>	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: 167.9 Specific Conductance Probe Cleaned? Yes (No) Dissolved Oxygen Membrane Changed: Yes (No)

Notes: changed batteries