

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: RQ-2016-06-06-22 Date: 6/23/16 Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV <u>CCV</u>		RA	6/23/16	0855			8.5	23.6	101	8.82	P
CAL ICV <u>CCV</u>		RA	6/23/16	0920			8.4	23.9		8.43	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL ICV CCV		RA	6/23/16	0900	1413	2/13/17	16-9			1409-1413	P
CAL ICV CCV		RA	6/23/16	0910	100	11/17	2511994			101	P
CAL ICV CCV		RA	6/23/16	0925	1417	2/13/17				1479	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV		RA	6/23/16	0915	7	5/17	2506518			7.06-7.00	P
CAL ICV CCV		RA	6/23/16	0916	4	3/18	2603491			3.86-4.00	P
CAL ICV CCV		RA	6/23/16	0920	10	9/16	2503774			4.78-4.98	P
CAL ICV CCV		RA	6/23/16	0921	4	3/18	2603491			4.0	P
CAL ICV CCV		RA	6/23/16	0925	7	5/17	2506518			7.0	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No