

Below "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: STATUS / Z1-LC-100005 Date: 8-1-2016 Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen

DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV	FB	8/1/16	8:10			8.23	25.4	100.5	8.23	P
CAL ICV CCV	FB	8/1/16	18:45			8.2	25.7	100.7	8.2	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Specific Conductance

DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL ICV CCV	FB	8/1/16	8:10	100	11/17	2511994	5		103	P
CAL ICV CCV	FB	8/1/16	8:10	1413	11/17	100165			1994	P
CAL ICV CCV	FB	8/1/16	18:45	100	11/17	2511994	5		103	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

pH

DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV	FB	8/1/16	8:10	7.0	5/20/17	2506518		-20.2	7.0	P
CAL ICV CCV	FB	8/1/16	8:10	4.0	5/20/17	2506518		153.0	4.0	P
CAL ICV CCV	FB	8/1/16	8:10	10.0	5/20/17	2506518		108.3	10.0	P
CAL ICV CCV	FB	8/1/16	8:15	7.0	5/20/17	2506518			7.07	P
CAL ICV CCV	FB	8/1/16	18:50	4.0	3/18	2603491			4.1	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: