

Botany X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # **DEP 132237**

Project/Site: Stahys Lakes / Homew 26-6-10003 Date: 8/3/2016

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	FB	8/3/16	8:30			25.4	100	8.253	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	FB	8/3/16	8:15	14.3	7/14/17	KCL 16-5			14.3	P
CAL	ICV	CCV	FB	8/3/16	8:30	14.0	8/17	2511924	5		9.7	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	FB	8/3/16	8:15	7.0	5/20/17	2546518			7.0	P
CAL	ICV	CCV	FB	8/3/16	8:30	4.0	3/20/18	2603491			4.0	P
CAL	ICV	CCV	FB	8/3/16	8:30	10.0	4/20/16	2543774			10.0	P
CAL	ICV	CCV	FB	8/3/16	8:30	7.0	5/20/17	2546518			7.0	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: 210 not sample