

Bolus, X" this box  
if there is qualified  
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: Eglw Run Date: 8/8/16

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

**Dissolved Oxygen**

| DEP SOP FT 1500    | Initials  | Date       | Time        | Probe Charge | Probe Gain | mg/L       | Temp °C     | % DO      | Saturation mg/L (from chart) | Pass or Fail |
|--------------------|-----------|------------|-------------|--------------|------------|------------|-------------|-----------|------------------------------|--------------|
| CAL ICV <u>CCV</u> | <u>mm</u> | <u>8/8</u> | <u>0745</u> |              |            | <u>7.9</u> | <u>27.3</u> | <u>99</u> | <u>7.926</u>                 | P            |
| CAL ICV <u>CCV</u> | <u>mm</u> | <u>8/8</u> | <u>1630</u> |              |            | <u>7.8</u> | <u>27.2</u> | <u>99</u> | <u>7.940</u>                 | P            |
| CAL ICV CCV        |           |            |             |              |            |            |             |           |                              | P            |
| CAL ICV CCV        |           |            |             |              |            |            |             |           |                              | P            |
| CAL ICV CCV        |           |            |             |              |            |            |             |           |                              | P            |
| CAL ICV CCV        |           |            |             |              |            |            |             |           |                              | P            |
| CAL ICV CCV        |           |            |             |              |            |            |             |           |                              | P            |
| CAL ICV CCV        |           |            |             |              |            |            |             |           |                              | P            |

Acceptance Criteria:  $\pm 0.3 \text{ mg/L}$

**Specific Conductance**

| DEP SOP FT 1200    | Initials  | Date       | Time        | Standard $\mu\text{mhos/cm}$ | Exp. Date       | Lot #           | Bottle # | Cell Constant | Reading $\mu\text{mhos/cm}$ | Pass or Fail |
|--------------------|-----------|------------|-------------|------------------------------|-----------------|-----------------|----------|---------------|-----------------------------|--------------|
| <u>CAL</u> ICV CCV | <u>mm</u> | <u>8/8</u> | <u>0750</u> | <u>12900</u>                 | <u>01/21/17</u> | <u>KCL 16-1</u> |          |               | <u>12100</u>                | P            |
| CAL <u>ICV</u> CCV | <u>mm</u> | <u>8/8</u> | <u>0751</u> | <u>1413</u>                  | <u>07/14/17</u> | <u>KCL 16-5</u> |          |               | <u>1405</u>                 | P            |
| CAL ICV <u>CCV</u> | <u>mm</u> | <u>8/8</u> | <u>1635</u> | <u>100</u>                   | <u>11/20/17</u> | <u>2511994</u>  |          |               | <u>103</u>                  | P            |
| CAL ICV CCV        |           |            |             |                              |                 |                 |          |               |                             | P            |
| CAL ICV CCV        |           |            |             |                              |                 |                 |          |               |                             | P            |
| CAL ICV CCV        |           |            |             |                              |                 |                 |          |               |                             | P            |
| CAL ICV CCV        |           |            |             |                              |                 |                 |          |               |                             | P            |

Acceptance Criteria:  $\pm 5\%$

**pH**

| DEP SOP FT 1100    | Initials  | Date       | Time        | Standard SU | Exp. Date       | Lot #          | Bottle # | mV / Slope    | Reading SU  | Pass or Fail |
|--------------------|-----------|------------|-------------|-------------|-----------------|----------------|----------|---------------|-------------|--------------|
| <u>CAL</u> ICV CCV | <u>mm</u> | <u>8/8</u> | <u>0755</u> | <u>7</u>    | <u>05/20/17</u> | <u>2505576</u> |          | <u>-27.2</u>  | <u>7.0</u>  | P            |
| CAL ICV CCV        | <u>mm</u> | <u>8/8</u> | <u>0756</u> | <u>4</u>    | <u>03/20/18</u> | <u>2603491</u> |          | <u>152.7</u>  | <u>4.0</u>  | P            |
| CAL ICV CCV        | <u>mm</u> | <u>8/8</u> | <u>0757</u> | <u>10</u>   | <u>09/12/16</u> | <u>2503774</u> |          | <u>-194.5</u> | <u>10.0</u> | P            |
| CAL <u>ICV</u> CCV | <u>mm</u> | <u>8/8</u> | <u>0758</u> | <u>4</u>    | <u>03/20/18</u> | <u>2603491</u> |          |               | <u>4.0</u>  | P            |
| CAL ICV <u>CCV</u> | <u>mm</u> | <u>8/8</u> | <u>1640</u> | <u>10</u>   | <u>05/20/17</u> | <u>2503774</u> |          |               | <u>10.0</u> | P            |
| CAL ICV CCV        |           |            |             |             |                 |                |          |               |             | P            |
| CAL ICV CCV        |           |            |             |             |                 |                |          |               |             | P            |
| CAL ICV CCV        |           |            |             |             |                 |                |          |               |             | P            |
| CAL ICV CCV        |           |            |             |             |                 |                |          |               |             | P            |

Acceptance Criteria:  $\pm 0.2 \text{ SU}$

Maintenance: Weekly pH Slope: 167.3 Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode: **CAL** - Calibrate  
Perform only in Run Mode: **ICV** - Initial Calibration Verification  
Perform only in Run Mode: **CCV** - Continuing Calibration Verification