

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: Snp Local

Date: 8/18/16

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	nm	8/18/16	0900			8.3	24.0	99	8.48	P
CAL	ICV	CCV										P
CAL	ICV	CCV	nm	8/18/16	1500			8.3	24.9	100	8.27	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Acceptance Criteria: $\pm 0.3 \text{ mg/L}$

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	nm	8/18/16	0900	1413	7/14/17	MC11605			154	P
CAL	ICV	CCV				100	7/14/17	2511494			104	P
CAL	ICV	CCV			1500	100	7/14/17	2511494				P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Acceptance Criteria: $\pm 5\%$

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	nm	8/18/16	0900	7	5/17	2501608			7.6	P
CAL	ICV	CCV				7	3/18	2602491			9.0	P
CAL	ICV	CCV				10	9/16	2503714			10.9	P
CAL	ICV	CCV				4	3/18	2602491			7.1	P
CAL	ICV	CCV			1500	7	5/17	2501608				P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Acceptance Criteria: $\pm 0.2 \text{ SU}$

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Below X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

21-00-10013 / HANCO

Project/Site: Reference Lake Status Date: 8/18/16

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
Acceptance Criteria: ± 0.3 mg/L												
CAL	ICV	CCV	mm	8/18	0720		1,00738	8.4	24.1	100.1	8.403	P
CAL	ICV	CCV										P
CAL	ICV	CCV	FS	8/19	0730			8.73	22.2	99.6	8.71	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
Acceptance Criteria: $\pm 5\%$												
CAL	ICV	CCV	mm	8/18	0721	12900	1/21/17	KCL 16-1			12900	P
CAL	ICV	CCV	mm	8/18	0722	1413	7/14/17	KCL 16-5			1416	P
CAL	ICV	CCV	FS	8/19	0730	100	11/20/17	2511994	5		143	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
Acceptance Criteria: ± 0.2 SU												
CAL	ICV	CCV	mm	8/18	0723	7	08/20/17	2506518			7.0	P
CAL	ICV	CCV	mm	8/18	0724	4	03/20/18	2603491			4.0	P
CAL	ICV	CCV	mm	8/18	0725	10	09/10/16	2563774			10.0	P
CAL	ICV	CCV	mm	8/18	0730	4	03/20/18	2603491			4.0	P
CAL	ICV	CCV	FS	8/19	0730	7	05/20/17	2506518			7.0	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: _____

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Initial Calibration Verification: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification