

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: SMP Salo

Date: 9/7/16

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen

DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV	NM	9/7/16	0900			8.31	24.4	100	8.336	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV	NM	9/8/16	0745			8.99	23.2	99	8.946	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Specific Conductance

DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
CAL ICV CCV	NM	9/7/16	0900	1413	7/14/17	N4165			1413	P
CAL ICV CCV				100	9/1/16	280308			104	P
CAL ICV CCV										P
CAL ICV CCV	NM	9/8/16	0745	100	9/1/16	280308			103	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

pH

DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV	NM	9/7/16	0900	7	5/17	280658			7.0	P
CAL ICV CCV				4	3/18	280308			4.0	P
CAL ICV CCV				4	3/18	280308			4.0	P
CAL ICV CCV				4	3/18	280308			4.05	P
CAL ICV CCV	NM	9/8/16	0745	7	5/17	280658			7.12	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Maintenance: Weekly pH Slope: _____

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: