

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # **DEP 113529**

Project/Site: Stokes Small Lake/21-56-10013 Date: 8-10-11-2016

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	FB	10-11-16	8:40			8.5	23.3	99.2	8.53	P
CAL	ICV	CCV										P
CAL	ICV	CCV	FB	10-11-16	16:35			8.25	24.7	99.2	8.3	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	FB	10-11-16	8:40	1403	7-14-17	KCL16-5	6		1413	P
CAL	ICV	CCV				100	11-20-17	2511994	6		1002	P
CAL	ICV	CCV				100	11-20-17	2511994	6		101	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	FB	10-11-16	8:40	7.0	5-20-17	2506518			7.0	P
CAL	ICV	CCV				7.0	3-20-18	2603491			7.0	P
CAL	ICV	CCV				10.0	10-20-17	1605846			10.0	P
CAL	ICV	CCV				4.0	3-20-18	2603491			4.0	P
CAL	ICV	CCV				7.0	5-20-17	2506518			7.05	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: BATTERIES changed on 8/20/2016.

Below, in this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: BQ. 2016 - 10-10-51 Date: 10/11/16 Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
											Acceptance Criteria: $\pm 0.3 \text{ mg/L}$	
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0705</u>			<u>8.3</u>	<u>24.2</u>	<u>99</u>	<u>8.387</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>1645</u>			<u>8.4</u>	<u>24.0</u>	<u>100</u>	<u>8.418</u>	<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
											Acceptance Criteria: $\pm 5\%$	
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0710</u>	<u>12900</u>	<u>KCL 16-6</u>	<u>10/4/17</u>			<u>12900</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0711</u>	<u>1413</u>	<u>KCL 16-6</u>	<u>7/14/17</u>			<u>1416</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>1650</u>	<u>100</u>	<u>11/20/17</u>	<u>2511994</u>			<u>102</u>	<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
											Acceptance Criteria: $\pm 0.2 \text{ SU}$	
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0715</u>	<u>7</u>	<u>05/20/17</u>	<u>2506518</u>		<u>-30.8</u>	<u>7.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0716</u>	<u>4</u>	<u>03/20/18</u>	<u>2603491</u>		<u>146.4</u>	<u>4.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0717</u>	<u>10</u>	<u>10/20/17</u>	<u>1605846</u>		<u>-19.17</u>	<u>10.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>0718</u>	<u>4</u>	<u>03/20/18</u>	<u>2603491</u>			<u>4.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>10/11</u>	<u>1655</u>	<u>10</u>	<u>10/20/17</u>	<u>1605846</u>			<u>10.0</u>	<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>
CAL	ICV	CCV										<u>P</u>

Maintenance: Weekly pH Slope: 17.2 Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification