

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box if there is qualified data on this page.

Project/Site: Johns Small / 21-SC-10032 Date: 10/23/2016

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	FS	10/23/16	8:40		8.67	22.4	100	8.67	(P) F
CAL	ICV	CCV									P F
CAL	ICV	CCV	FS	10/23/16	16:45		8.27	21.6	99.9	8.3	(P) F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard umhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading umhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	FS	10/23/16	8:40	1413	KCL 16-8	10/4/2017		1413	(P) F
CAL	ICV	CCV				100	11/2017	2511999	6	101	(P) F
CAL	ICV	CCV	FS	10/23/16	16:45	100	11/2017	2511999	6	103	(P) F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	FS	10/23/16	11:40	7.0	4/2008	2605804		7.0	(P) F
CAL	ICV	CCV				9.0	3/2018	2603491		9.0	(P) F
CAL	ICV	CCV				10.0	10/2017	1605846		10.0	(P) F
CAL	ICV	CCV				9.0	3/2018	2603491		9.03	(P) F
CAL	ICV	CCV	FS	10/23/16	16:45	7.0	4/2017	2605804		7.0	(P) F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

Maintenance: Weekly pH Slope: Buffalo Buffalo Lake Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No