

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # DEP 132237

Date: 11-15-2016

Project/Site: SMP/2016-10-21-08

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	FB	11/15/16	9:15			21.7	90.3	8.79	P
CAL	ICV	CCV									P
CAL	ICV	CCV	MM	11/16	0800			21.9	99	8.76	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
CAL	ICV	CCV	FB	11/15/16	9:15	12900	10/4/17	10CL 16-6			12900	P
CAL	ICV	CCV	FB	11/16/16	9:15	1413	10/4/17	10CC 16-8			1385	P
CAL	ICV	CCV	MM	11/16/16	0800	58640	10/4/17	10CL 16-7			57400	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	FB	11/15/16	9:15	7.0	04/20/18	2605805			6.94	P
CAL	ICV	CCV				7.0	05/20/18	26053491				P
CAL	ICV	CCV				7.0	10/20/17	1605540			6.0	P
CAL	ICV	CCV	FB	11/15	9:15	7.0	04/20/18	2605805			7.09	P
CAL	ICV	CCV	MM	11/16	0800	7.0	10/17	16053491			9.98	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: Tree Lake, Big Lagoon, Paradise Bay/Levee

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification