

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site:

RA-2016-11-07-29

Date:

11/17/16

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV	CCV	MM	11/17	0825			8.70	22.2	100	8.11	P
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV	MM	11/17	1155			8.62	23.1	101	8.56	P
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV										F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL ICV CCV	CCV	MM	11/17	0825	1413	10/4/17	KCL 16-7			1413	P
CAL ICV CCV	CCV	X	11/17	1155	100	11/1/17	2511999			104	P
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV	MM	11/17	1155	5840	10/4/17	KCL 16-7			5988	P
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV										F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV	CCV	MM	11/17	0825	7	4/18	260804			7.0	P
CAL ICV CCV	CCV	X	11/17	1155	4	8/18	2608103			4.0	P
CAL ICV CCV	CCV	X	11/17	1155	10	10/17	1608349			16.0	P
CAL ICV CCV	CCV					4/18	2608104			7.02	P
CAL ICV CCV	CCV	MM	11/17	1155	4	8/18	2608103			4.1	P
CAL ICV CCV	CCV										F
CAL ICV CCV	CCV										F

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: Audit