

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # **DEP 132237**

Date: 11/21/16

Date: 11/21/16

Project/Site: MP Elin

For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	---------	------	------------------------------	--------------

CAL ICV CCV	MM	11/21	0730				21.1	99	8.84	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV	MM	11/21	1400				18.8	99	9.31	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL ICV CCV	MM	11/21	0730		1413	10/4/17	10116-8			1413	P
CAL ICV CCV					100	11/17	2511994			100	P
CAL ICV CCV											P
CAL ICV CCV	MM	11/21	1400		100	11/17	2511994			100	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV CCV	MM	11/21	0730		7	4/18	210804		-39.5	7.0	P
CAL ICV CCV					7	8/18	210803			4.0	P
CAL ICV CCV					7	10/17	210804			10.0	P
CAL ICV CCV					7	4/18	210804		-39.5	7.0	P
CAL ICV CCV	MM	11/21	1400		4	8/18	210803		190.9	4.00	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: _____

CAL - Calibrate
ICV - Initial Calibration Verification
CCV - Continuing Calibration Verification