

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: SMP Egin

Date: 12/7/16

Meter # DEP 113529

For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	NM	12/7	0730			21.3	8.95	8.86	P
CAL	ICV	CCV									P
CAL	ICV	CCV	NM	12/8	1300			20.0	9.27	9.09	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	NM	12/7	0735	1413	10/4/17	KC16-8			1913	P
CAL	ICV	CCV			0736	100	11/17	2511994			101	P
CAL	ICV	CCV	NM	12/8	1302	100	11/17	2511994			104	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	NM	12/7	0740	7	4/18	2608804			147.6	P
CAL	ICV	CCV			0741	4	8/18	2608803			147.6	P
CAL	ICV	CCV			0742	19	8/18	2608804			147.6	P
CAL	ICV	CCV			0743	7	8/18	2608804			147.6	P
CAL	ICV	CCV	NM	12/8	1306	7	4/18	2608804			147.6	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification