

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "V" this box
if there is qualified
data on this page.

Project/Site:

SMP FWS/ROCKY B/CHARTER

Meter #

DEP 113529

Temperature (Quarterly)

For Date of Last Temperature Verification see

In log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Change	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	--------------------	----------	------	------	-----------------	---------------	------	---------	------	------------------------------------	-----------------

CAL	ICV	CCV	MM	12/12/16	0736		8.98	20.9	101	8.932	P
CAL	ICV	CCV	W	12/13	1645		8.7	21.0	99	8.812	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard µmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading µmhos/cm	Pass or Fail
----------------------	--------------------	----------	------	------	----------------------	-----------	-------	----------	------------------	---------------------	-----------------

CAL	ICV	CCV	MM	12/12/16	0735	1413	10/4/17		ACU 16-8	1415	P
CAL	ICV	CCV	W	12/13	0736	1000	11/17		BS1994	100	P
CAL	ICV	CCV	W	12/13	1650	5850	10/4/17		KC10-7	58793	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	--------------------	----------	------	------	----------------	-----------	-------	----------	------------	---------------	-----------------

CAL	ICV	CCV	MM	12/12/16	0740	7	4/18		260804	70	P
CAL	ICV	CCV	W	12/13	0741	4	8/18		2608503	9.0	P
CAL	ICV	CCV	W	12/13	0742	10	10/17		2608314	10.0	P
CAL	ICV	CCV	W	12/13	0743	4	0/18		2608403	148.7	P
CAL	ICV	CCV	W	12/13	1653	10	10/20/17		1605744	10.0	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: