

Chemical Analysis Report

NW-DIST-2016-06-21-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **TMDL Bioassessment**
Request ID: **RQ-2016-06-06-22**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 W Government St
Suite 208A
Pensacola, FL 32502-5794
Attn: Michael King

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Environmental Administrator

Date Certified: 08-JUL-2016 16:15

A handwritten signature in black ink, appearing to read "Colin Wright", is written on a light-colored rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Bray Mill Creek North of Oil Plant Road

Collection Date/Time: 06/20/2016 11:10

Field ID: G4NW0066

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1809529	EPA 180.1 Rev. 2.0	Turbidity	19	A	NTU	P306570	
	EPA 300.0 Rev. 2.1	Chloride	9.3		mg Cl/L	P307114	
		Sulfate	3.8		mg SO4/L	P307112	
	SM 2120 B	Color (true)	2.5	U	PCU	P306577	
	SM 2320 B-97	Total Alkalinity	5.8		mg CaCO3/L	P306823	
	SM 2540 C-97	TDS	64		mg/L	P307012	
	SM 2540 D-97	TSS	34		mg/L	P306990	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P307268	
1809531	EPA 350.1 Rev. 2.0	Ammonia-N	0.12		mg N/L	P306895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.17	I	mg N/L	P306604	
	EPA 353.2 Rev. 2.0	NO2NO3-N	3.7		mg N/L	P306831	
	EPA 365.1 Rev. 2.0	Total-P	0.019		mg P/L	P306661	
	SM 5310 B-00	Organic Carbon	0.89	I	mg C/L	P307065	
1809533	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P306560	

Ref. Method and Comment:

SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Sample Location: Gum Creek abv Bartlett Rd.

Collection Date/Time: 06/20/2016 14:00

Field ID: G4NW0323

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1809530	EPA 180.1 Rev. 2.0	Turbidity	2.6		NTU	P306570	
	EPA 300.0 Rev. 2.1	Chloride	4.8		mg Cl/L	P307114	
		Sulfate	0.23	I	mg SO4/L	P307112	
	SM 2120 B	Color (true)	160	A	PCU	P306574	
	SM 2320 B-97	Total Alkalinity	1.7	I	mg CaCO3/L	P306823	
	SM 2540 C-97	TDS	44		mg/L	P307012	
	SM 2540 D-97	TSS	9	I	mg/L	P306990	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P307268	
1809532	EPA 350.1 Rev. 2.0	Ammonia-N	0.043		mg N/L	P306895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.71		mg N/L	P306604	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.27		mg N/L	P306831	
	EPA 365.1 Rev. 2.0	Total-P	0.038		mg P/L	P306661	
	SM 5310 B-00	Organic Carbon	16		mg C/L	P307065	
1809534	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.016		mg P/L	P306560	

Ref. Method and Comment:

SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Matrix spike recovery data is unavailable due to high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P306570

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P307112

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P307114

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P306895

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P306604

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P306831

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P306661

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P306560

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P306574

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report Method Blank Results

Reference Method: SM 2120 B
Batch ID: P306577

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P306823

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P307012

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P306990

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P307268

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P307065

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P307112

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P307114

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P306895

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P306895

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	94.3		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P306604

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P306831

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P306661

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P306560

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2320 B-97
 Batch ID: P306823

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
 Batch ID: P307268

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.3		P	92 - 103

Reference Method: SM 5310 B-00
 Batch ID: P307065

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P307112

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P307112

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1809380	Sulfate	101		P	90 - 110
1809401	Sulfate	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P307114

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1809380	Chloride	102		P	90 - 110
1809401	Chloride	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P306831

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1809500	NO2NO3-N	95.8	97.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P306661

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1809580	Total-P	95.4	95.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P306560

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1809533	O-Phosphate-P	102	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P307268

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1810894	Fluoride	97.8	99.6	P/P	91 - 105

Reference Method: SM 5310 B-00
 Batch ID: P307065

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1809003	Organic Carbon	96.6	103	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P306570

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809529	Turbidity	3.60	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P307112

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809380	Sulfate	1.33	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P307114

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809380	Chloride	0.0617	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P306895

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809531	Ammonia-N	0.0139	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P306604

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809500	Kjeldahl Nitrogen	3.26	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P306831

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809500	NO2NO3-N	1.03	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P306661

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809580	Total-P	0.0748	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P306560

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809533	O-Phosphate-P	0.0	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P306574

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809530	Color (true)	0.982	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
 Batch ID: P306577

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809716	Color (true)	3.51	Sample	P	0 - 20

Reference Method: SM 2320 B-97
 Batch ID: P306823

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809379	Total Alkalinity	0.217	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
 Batch ID: P307012

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809494	TDS	2.76	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
 Batch ID: P307268

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1810894	Fluoride	1.45	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
 Batch ID: P307065

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1809003	Organic Carbon	3.51	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
 Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Run ID: A70146
 Included Lab Sample IDs: 1809533, 1809534

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.1	99.1	P/P	90 - 110
O-Phosphate-P	99.7	99.1	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
 Run ID: A70148
 Included Lab Sample IDs: 1809529, 1809530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	101	P/P	95 - 105

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A70149

Included Lab Sample IDs: 1809529, 1809530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.6	96.4	P/P	90 - 115
Color (true)	98.8	97.6	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A70200

Included Lab Sample IDs: 1809531, 1809532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	100	P/P	90 - 110
Total-P	99.2	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A70202

Included Lab Sample IDs: 1809531, 1809532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	105	108	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A70213

Included Lab Sample IDs: 1809531, 1809532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A70215

Included Lab Sample IDs: 1809529, 1809530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	106	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A70237

Included Lab Sample IDs: 1809531, 1809532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	95.8	95.4	P/P	90 - 110
Ammonia-N	96.0	95.8	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A70300

Included Lab Sample IDs: 1809531, 1809532

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	104	105	P/P	90 - 110
Organic Carbon	105	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1
 Run ID: A70313
 Included Lab Sample IDs: 1809529, 1809530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	101	P/P	90 - 110
Sulfate	96.3	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Run ID: A70370
 Included Lab Sample IDs: 1809529, 1809530

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.5	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
 Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
 Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				3.60	
EPA 300.0 Rev. 2.1	Chloride	103	102 102		0.0617	
	Sulfate	102	101 101		1.33	
EPA 350.1 Rev. 2.0	Ammonia-N	94.3				0.0139
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103				3.26
EPA 353.2 Rev. 2.0	NO2NO3-N	98.5	95.8 97.2			1.03
EPA 365.1 Rev. 2.0	Total-P	97.0	95.4 95.4			0.0748
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	102 102			0.0
SM 2120 B	Color (true)				0.982	
	Color (true)				3.51	
SM 2320 B-97	Total Alkalinity	103			0.217	
SM 2540 C-97	TDS				2.76	
SM 4500 F-C-97	Fluoride	96.3	97.8 99.6			1.45
SM 5310 B-00	Organic Carbon	101	96.6 103			3.51

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1809529, 1809530
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1809529, 1809530
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1809529, 1809530
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1809531, 1809532
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1809531, 1809532
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1809531, 1809532
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1809531, 1809532

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1809533, 1809534
SM 2120 B / E31780	True Color measured at 450 nm	1809529, 1809530
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1809529, 1809530
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1809529, 1809530
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1809529, 1809530
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1809529, 1809530
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1809531, 1809532

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/21/2016			06/21/2016 15:30	Sima Feizi	1809529, 1809530
EPA 300.0 Rev. 2.1	06/21/2016			06/29/2016	Ping Hua	1809529, 1809530
EPA 350.1 Rev. 2.0	06/21/2016			06/24/2016	Diana M. Turner	1809531, 1809532
EPA 351.2 Rev. 2.0	06/21/2016	06/22/2016	Sofia Elnemr	06/23/2016	Christopher Armour	1809531, 1809532
EPA 353.2 Rev. 2.0	06/21/2016			06/24/2016	Virginia P. Brown	1809531, 1809532
EPA 365.1 Rev. 2.0	06/21/2016	06/22/2016	Vijayalakshmi Reddy	06/23/2016	Lipi Saha	1809531, 1809532
EPA 365.1 Rev. 2.0 dissolved	06/21/2016			06/21/2016 14:24	Julia Storbeck	1809533
	06/21/2016			06/21/2016 14:43	Julia Storbeck	1809534
SM 2120 B	06/21/2016			06/21/2016 16:12	Chrisoulla Rakowski	1809529
	06/21/2016			06/21/2016 16:20	Chrisoulla Rakowski	1809530
SM 2320 B-97	06/21/2016			06/23/2016	Rochelle Y Jackson	1809529, 1809530
SM 2540 C-97	06/21/2016			06/22/2016	Yijie Li	1809529, 1809530
SM 2540 D-97	06/21/2016			06/22/2016	Yijie Li	1809529, 1809530
SM 4500 F-C-97	06/21/2016			06/30/2016	Dale L. Simmons	1809529
	06/21/2016			07/01/2016	Dale L. Simmons	1809530
SM 5310 B-00	06/21/2016			06/28/2016	Diana M. Turner	1809531, 1809532