



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 148287

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 5/24/16 TIME 8:18

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 5/24/16 TIME 1600 hrs

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.43	23.612	99.7%	☒
CALIBRATION	8.50	23.575	100.0%	☒
PRE-VERIFICATION	8.50	23.530	100.1%	☒
POST-VERIFICATION	8.18	26.243	100.3%	☒

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	500	2512A07	12/17	496.5	☒
CALIBRATION	500	2512A07	12/17	500.0	☒
PRE-VERIFICATION	500	2512A07	12/17	500.0	☒
POST-VERIFICATION	500	2512A07	12/17	520.0	☒

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.0	152111	04/17	7.01	23.188	☒
CALIBRATION	7.0	152111	04/17	7.00	23.191	☒
CALIBRATION	4.0					
CALIBRATION	10.0	158638	01/18	10.00	23.182	☒
PRE-VERIFICATION	10.0	158638	01/18	9.99	23.185	☒
POST-VERIFICATION	7.0	152111	4/17	7.05	23.543	☒

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO ☐

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☐ NO ☒ N/A

RQ & SITE IDS

RQ-2015-09-21-14

SPRG1507



BIG BLUE SPRING (JEFFERSON)

SPRG1507 BIG BLUE SPRING (JEFFERSON)



QAQC BLANK 3

SPRG1504



WACISSA HEAD SPRING NUMBER

SPRG1507 WACISSA HEAD SPRING NUMBER 2

SAMPLER NAMES & SIGNATURES (All team members present)

Mike Dalis *[Signature]*