CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

box if there is qualified data on this page.

Meter ID: 4 Don Project: 3/16/16

Notes: (1) Numbers ≤4, are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2). RQ-2016-03-14-44
(2) Always wait for meter to stabilize before recording any readings.
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 1/4

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	A Salmeron	3/16	0840	23.2	8.5	8.5	100	52	1.0	P	L
ICV				23.3	8.5	8.5	100	51			
CCV			1342	23.2	8.5	8.5	100	51			
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.
DO Acceptance criteria from Table ±0.3 mg/L. **Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15; DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	A. Salmeron	3/16	0816	151028E	11/16	1600	1000	P	L
ICV				151028D	5/16	100	100		
CCV			1344	151028F	11/16	500	510		
CCV									

Report specific conductance as a whole number with no decimal figure. Conductivity Acceptance criteria ± 5%

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	A. Salmeron	3/16	0818	157448	12/17	7.0	7.0	-15	P	L
Calibr.				157226	11/17	4.0	4.0	162		
Calibr.				151028A	5/17	10.0	10.0	-181		
ICV				151028A	5/17	10.0	10.0	-181		
CCV			1343	157226	11/17	4.0	4.0	160		
CCV										

Report pH with one decimal place; pH Acceptance criteria ±0.2 SU; mV pH7 Range 0±50; mV pH 4 Range +180±50;
mV pH 10 Range -180±50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly) Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria ±5 % or ±0.05m, whichever is greater