

Meter ID: W1 Jon

Project: SNAP

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g. 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 3/30

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	A. Salmeron	6/27	0742	23.7	8.5	8.5	100	5	1.4	P	L
ICV				23.7	8.5	8.5	100	51			
CCV	S. Clingerman	1	1415	24.7	8.3	8.4	101	51		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.
 DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Optical: DO gain range 0.85 to 1.15; DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	A. Salmeron	6/27	0744	2401A06	05/18	1000	1000	P	L
ICV				26031142	02/18	100	104		
CCV	S. Clingerman	1	1418	151028F	11/16	500	485	P	L
CCV									

Report specific conductance as a whole number with no decimal figure. Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	A. Salmeron	6/27	0746	158437	01/18	7.0	7.0	-30	P	L
Calibr.				158982	01/18	4.0	4.0	150		
Calibr.				156442	12/17	10.0	10.0	-220		
ICV				156442	12/17	10.0	10.0	-220		
CCV	S. Clingerman		1421	158982	01/18	4.0	4.0	144	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$;
 mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly) Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater