

Chemical Analysis Report

SW-DIST-2016-10-27-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Bellows, Lk Thomas Outlet**
Request ID: **RQ-2016-10-24-68**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Judy Ashton

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Environmental Administrator

Date Certified: 21-NOV-2016 15:41

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: BELLOWS LAKE/EAST LAKE

Collection Date/Time: 10/26/2016 09:25

Field ID: G1SW0044

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1845016	EPA 180.1 Rev. 2.0	Turbidity	20	A	NTU	P314064	
	EPA 300.0 Rev. 2.1	Chloride	5.7		mg Cl/L	P314989	
		Sulfate	5.7		mg SO4/L	P314992	
	SM 2120 B	Color (true)	52		PCU	P314096	
	SM 2320 B-97	Total Alkalinity	46		mg CaCO3/L	P314176	
	SM 2540 C-97	TDS	71		mg/L	P314712	
	SM 2540 D-97	TSS	30		mg/L	P314673	
	SM 4500 F-C-97	Fluoride	0.12		mg F/L	P314181	
1845018	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P314906	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	3.4		mg N/L	P315116	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P314242	
	EPA 365.1 Rev. 2.0	Total-P	0.051		mg P/L	P314771	
	SM 5310 B-00	Organic Carbon	16		mg C/L	P314166	
1845020	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P314069	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: LAKE THOMAS OUTLET

Collection Date/Time: 10/26/2016 11:45

Field ID: G1SW0050

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1845017	EPA 180.1 Rev. 2.0	Turbidity	3.9		NTU	P314064	
	EPA 300.0 Rev. 2.1	Chloride	29		mg Cl/L	P314989	
		Sulfate	3.0		mg SO4/L	P314992	
	SM 2120 B	Color (true)	90		PCU	P314096	
	SM 2320 B-97	Total Alkalinity	30		mg CaCO3/L	P314176	
	SM 2540 C-97	TDS	108		mg/L	P314712	
	SM 2540 D-97	TSS	18		mg/L	P314673	
	SM 4500 F-C-97	Fluoride	0.072	I	mg F/L	P314181	
1845019	EPA 350.1 Rev. 2.0	Ammonia-N	0.063		mg N/L	P314906	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P315116	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.012		mg N/L	P314242	
	EPA 365.1 Rev. 2.0	Total-P	0.033		mg P/L	P314771	
	SM 5310 B-00	Organic Carbon	14		mg C/L	P314166	
1845021	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.007	I	mg P/L	P314069	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P314064

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314989

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314992

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P314906

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P315116

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P314242

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P314771

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P314069

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P314096

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P314176

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2540 C-97
Batch ID: P314712

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P314673

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P314181

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P314166

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314989

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314992

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	94.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P314906

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P315116

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P314242

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P314242

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P314771

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P314069

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.2		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P314176

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P314181

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P314166

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314989

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1844798	Chloride	97.6		P	90 - 110
1845290	Chloride	96.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314992

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1844798	Sulfate	94.6		P	90 - 110
1845290	Sulfate	93.8		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P314906

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1845004	Ammonia-N	93.9	91.9	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P315116

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1844800	Kjeldahl Nitrogen	103	98.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P314242

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1845018	NO2NO3-N	94.5	94.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P314771

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1844917	Total-P	96.1	97.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P314069

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1845081	O-Phosphate-P	97.4	98.6	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P314181

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1845002	Fluoride	97.8	99.2	P/P	91 - 105

Reference Method: SM 5310 B-00
 Batch ID: P314166

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1845018	Organic Carbon	91.7	92.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P314064

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845016	Turbidity	4.06	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314989

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845290	Chloride	0.0504	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P314992

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845290	Sulfate	1.85	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P314906

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845004	Ammonia-N	2.07	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P315116

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1844800	Kjeldahl Nitrogen	3.04	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P314242

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845018	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P314771

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1844917	Total-P	0.777	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P314069

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845081	O-Phosphate-P	0.920	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P314096

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1844916	Color (true)	5.21	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P314176

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845002	Total Alkalinity	0.107	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P314712

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845002	TDS	3.72	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P314181

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845002	Fluoride	0.946	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P314166

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1845018	Organic Carbon	0.440	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A72478
Included Lab Sample IDs: 1845016, 1845017

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	102	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A72479
Included Lab Sample IDs: 1845020, 1845021

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	98.8	P/P	90 - 110

Reference Method: SM 2120 B
Run ID: A72485
Included Lab Sample IDs: 1845016, 1845017

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	99.7	P/P	85 - 115

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00

Run ID: A72513

Included Lab Sample IDs: 1845018, 1845019

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.1	97.3	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A72515

Included Lab Sample IDs: 1845016, 1845017

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	97.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A72515

Included Lab Sample IDs: 1845016, 1845017

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A72533

Included Lab Sample IDs: 1845018, 1845019

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	96.0	95.5	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A72715

Included Lab Sample IDs: 1845016, 1845017

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	99.1	97.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A72730

Included Lab Sample IDs: 1845018

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A72736

Included Lab Sample IDs: 1845018, 1845019

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.5	99.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A72760

Included Lab Sample IDs: 1845019

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A72832
Included Lab Sample IDs: 1845018, 1845019

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	105	107	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				4.06	
EPA 300.0 Rev. 2.1	Chloride	98.4	96.6 97.6		0.0504	
	Sulfate	94.3	93.8 94.6		1.85	
EPA 350.1 Rev. 2.0	Ammonia-N	101	93.9 91.9			2.07
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	103 98.9			3.04
EPA 353.2 Rev. 2.0	NO2NO3-N	99.5	94.5 94.5			0.0
EPA 365.1 Rev. 2.0	Total-P	98.2	96.1 97.1			0.777
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.2	97.4 98.6			0.920
SM 2120 B	Color (true)				5.21	
SM 2320 B-97	Total Alkalinity	104			0.107	
SM 2540 C-97	TDS				3.72	
SM 4500 F-C-97	Fluoride	101	97.8 99.2			0.946
SM 5310 B-00	Organic Carbon	97.0	91.7 92.4			0.440

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1845016, 1845017
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1845016, 1845017
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1845016, 1845017
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1845018, 1845019
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1845018, 1845019
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1845018, 1845019
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1845018, 1845019
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1845020, 1845021
SM 2120 B / E31780	True Color measured at 450 nm	1845016, 1845017
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1845016, 1845017
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1845016, 1845017
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1845016, 1845017
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1845016, 1845017
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1845018, 1845019

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	10/27/2016			10/27/2016 15:34	Sima Feizi	1845016, 1845017
EPA 300.0 Rev. 2.1	10/27/2016			11/10/2016	Ping Hua	1845016, 1845017
EPA 350.1 Rev. 2.0	10/27/2016			11/09/2016	Sofia Elnemr	1845018, 1845019
EPA 351.2 Rev. 2.0	10/27/2016	11/15/2016	Adrian Shelby	11/16/2016	Joseph Suarez	1845018, 1845019
EPA 353.2 Rev. 2.0	10/27/2016			10/31/2016	Beverly Y. Sanders	1845018, 1845019
EPA 365.1 Rev. 2.0	10/27/2016	11/08/2016	Adrian Shelby	11/09/2016	Lipi Saha	1845018
	10/27/2016	11/08/2016	Adrian Shelby	11/10/2016	Lipi Saha	1845019
EPA 365.1 Rev. 2.0 dissolved	10/27/2016			10/27/2016 15:52	Julia Storbeck	1845020
	10/27/2016			10/27/2016 15:53	Julia Storbeck	1845021
SM 2120 B	10/27/2016			10/27/2016 15:00	Julie Michelle Frank	1845016
	10/27/2016			10/27/2016 15:01	Julie Michelle Frank	1845017
SM 2320 B-97	10/27/2016			10/29/2016	Dale L. Simmons	1845016, 1845017
SM 2540 C-97	10/27/2016			11/01/2016	Yijie Li	1845016, 1845017
SM 2540 D-97	10/27/2016			11/01/2016	Yijie Li	1845016, 1845017
SM 4500 F-C-97	10/27/2016			10/29/2016	Dale L. Simmons	1845016, 1845017
SM 5310 B-00	10/27/2016			10/28/2016	Julie Michelle Frank	1845018, 1845019