

Chemical Analysis Report

SW-DIST-2016-11-15-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **NorthCreek**
Request ID: **RQ-2016-11-14-36**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Judy Ashton

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 07-DEC-2016 11:43

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: North Creek

Collection Date/Time: 11/14/2016 11:10

Field ID: G3SW0034

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1850029	EPA 180.1 Rev. 2.0	Turbidity	31		NTU	P315248	
	EPA 300.0 Rev. 2.1	Chloride	110		mg Cl/L	P315883	
		Sulfate	410		mg SO4/L	P315885	
	SM 2120 B	Color (true)	55		PCU	P315226	
	SM 2320 B-97	Total Alkalinity	357		mg CaCO3/L	P315336	
	SM 2540 C-97	TDS	1.11E+03		mg/L	P315617	
	SM 2540 D-97	TSS	58		mg/L	P315628	
	SM 4500 F-C-97	Fluoride	0.46		mg F/L	P315338	
1850030	EPA 350.1 Rev. 2.0	Ammonia-N	0.090		mg N/L	P315661	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.3		mg N/L	P315873	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.50		mg N/L	P315532	
	EPA 365.1 Rev. 2.0	Total-P	0.31		mg P/L	P315764	
	SM 5310 B-00	Organic Carbon	17		mg C/L	P315311	
1850031	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.059		mg P/L	P315189	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P315248

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P315883

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P315885

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P315661

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P315873

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P315532

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P315764

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315189

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P315226

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P315336

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P315617

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P315628

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P315338

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P315311

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P315883

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	96.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P315885

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	93.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P315661

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P315873

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P315532

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P315764

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P315189

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2320 B-97

Batch ID: P315336

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	95 - 108

Reference Method: SM 4500 F-C-97

Batch ID: P315338

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.4		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P315311

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P315883

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1848212	Chloride	99.7		P	90 - 110
1850222	Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P315885

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1848212	Sulfate	95.7		P	90 - 110
1850222	Sulfate	96.7		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P315873

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1849988	Kjeldahl Nitrogen	97.0	91.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P315532

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1849828	NO2NO3-N	99.0	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P315764

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850173	Total-P	92.0	92.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P315189

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850031	O-Phosphate-P	95.8	96.9	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P315338

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1850105	Fluoride	95.7	96.8	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P315311

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1849688	Organic Carbon	98.4	99.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P315248

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849998	Turbidity	0.304	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P315883

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1848212	Chloride	0.299	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P315885

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1848212	Sulfate	0.894	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315661

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849829	Ammonia-N	0.222	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P315873

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849988	Kjeldahl Nitrogen	5.43	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P315532

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849828	NO2NO3-N	0.541	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P315764

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850173	Total-P	0.0276	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P315189

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850031	O-Phosphate-P	0.710	Spike	P	0 - 20

Reference Method: SM 2120 B

Batch ID: P315226

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850001	Color (true)	1.81	Sample	P	0 - 20

Reference Method: SM 2320 B-97

Batch ID: P315336

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850105	Total Alkalinity	2.52	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97

Batch ID: P315617

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849687	TDS	2.89	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97

Batch ID: P315338

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1850105	Fluoride	1.04	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00

Batch ID: P315311

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849688	Organic Carbon	0.498	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A72815

Included Lab Sample IDs: 1850031

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	100	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A72826

Included Lab Sample IDs: 1850029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	102	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A72831

Included Lab Sample IDs: 1850029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	100	P/P	95 - 105

Reference Method: SM 5310 B-00

Run ID: A72849

Included Lab Sample IDs: 1850030

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	100	101	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A72857

Included Lab Sample IDs: 1850029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A72857

Included Lab Sample IDs: 1850029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	99.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A72925

Included Lab Sample IDs: 1850030

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.0	98.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A72977

Included Lab Sample IDs: 1850030

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A73043

Included Lab Sample IDs: 1850030

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A73064

Included Lab Sample IDs: 1850029

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	98.2	96.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73124

Included Lab Sample IDs: 1850030

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.6	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.304	
EPA 300.0 Rev. 2.1	Chloride	96.0	99.7 101		0.299	
	Sulfate	93.4	95.7 96.7		0.894	
EPA 350.1 Rev. 2.0	Ammonia-N	99.9				0.222
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	100	97.0 91.0			5.43
EPA 353.2 Rev. 2.0	NO2NO3-N	103	99.0 100			0.541
EPA 365.1 Rev. 2.0	Total-P	101	92.0 92.0			0.0276
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	95.8 96.9			0.710
SM 2120 B	Color (true)				1.81	
SM 2320 B-97	Total Alkalinity	101			2.52	
SM 2540 C-97	TDS				2.89	
SM 4500 F-C-97	Fluoride	98.4	95.7 96.8			1.04
SM 5310 B-00	Organic Carbon	99.4	98.4 99.2			0.498

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1850029
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1850029
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1850029

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1850030
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1850030
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1850030
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1850030
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1850031
SM 2120 B / E31780	True Color measured at 450 nm	1850029
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1850029
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1850029
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1850029
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1850029
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1850030

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/15/2016			11/15/2016 16:29	Sima Feizi	1850029
EPA 300.0 Rev. 2.1	11/15/2016			11/30/2016	Ping Hua	1850029
EPA 350.1 Rev. 2.0	11/15/2016			11/21/2016	Sofia Elnemr	1850030
EPA 351.2 Rev. 2.0	11/15/2016	11/30/2016	Adrian Shelby	12/01/2016	Peter D. Kaus	1850030
EPA 353.2 Rev. 2.0	11/15/2016			11/21/2016	Beverly Y. Sanders	1850030
EPA 365.1 Rev. 2.0	11/15/2016	11/28/2016	Adrian Shelby	11/29/2016	Lipi Saha	1850030
EPA 365.1 Rev. 2.0 dissolved	11/15/2016			11/15/2016 16:16	Julia Storbeck	1850031
SM 2120 B	11/15/2016			11/15/2016 17:09	Jared I. Manley	1850029
SM 2320 B-97	11/15/2016			11/17/2016	Dale L. Simmons	1850029
SM 2540 C-97	11/15/2016			11/18/2016	Yijie Li	1850029
SM 2540 D-97	11/15/2016			11/18/2016	Yijie Li	1850029
SM 4500 F-C-97	11/15/2016			11/17/2016	Dale L. Simmons	1850029
SM 5310 B-00	11/15/2016			11/16/2016	Julie Michelle Frank	1850030