

RQ - 2016-05-16-37

Boldly "X" this box if there is qualified data on this page.

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID:

A23

Project:

SMP Orange

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration

or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

4/7

DO DEP SOP FT 1500	Name	Date	Time GT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Janes P.H.S	5/16	8:00	24.0	8.4	8.4	100	43.1	1.084	P	L
ICV	" "			24.1	8.4	8.4	100	43.1		P	L
CCV		5/16	17:50	25.5	8.2	8.2	99	43.1		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time GT-ET	Lot #	Expir. Date	Standard $\mu\text{mhos/cm}$	Meter Reading $\mu\text{mhos/cm}$	Pass / Fail	Lab / Field
Calibr.	Janes P.H.S	5/16	8:10	2602002	02/18	1000	1000	P	L
ICV	" "			2603082	3/18	100	101	P	L
CCV		5/16	17:51	2602002	2/18	1000	973	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1400	Name	Date	Time GT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Janes P.H.S	5/16	8:15	156651	10/17	7.0	7.0	-25.2	P	L
Calibr.	" "			158982	01/18	4.0	4.0	151.1	P	L
Calibr.	" "			156700	10/17	10.0	9.99	-200	P	L
ICV	" "			" "	" "	10.0	9.93	-198.8	P	L
CCV		5/16	17:05	158982	1/18	4.0	4.1	147.4	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

4/7

Depth Sensor (Daily)	Name	Date	Time GT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	Janes P.H.S	5/16	8:30	0.000	0.000	P	L

ICV acceptance criteria $\pm 5\%$ or $\pm 0.05\text{m}$, whichever is greater