

RQ-2016-06-27-19

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

Lil Gus

Project:

WCA SMP

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/13/16

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chan mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Brewton	6/29/16	9:00	23.1	8.6	8.6	100	48	1.2	P	L
ICV	"	"	9:02	23.1	8.6	8.6	100	48		P	L
CCV	J. Pitts	6/29/16	16:30	22.1	8.7	8.7	100	46		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond FT 1200	Name	Date	Time	Lot #	Expir. Date	Standard umhos/cm	Meter Reading umhos/cm	Pass / Fail	Lab / Field
Calibr.	Brewton	6/29/16	9:05	2606003	03/18	1000	1000	P	L
ICV	"	"	9:10	2603002	03/18	100	100	P	L
CCV	J. Pitts	6/29/16	16:32	2606003	02/18	1000	998	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1400	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter Reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Brewton	6/29/16	9:07	1506051	10/17	7.0	7.0	-7	P	L
Calibr.	"	"	9:08	1615023	02/18	4.0	4.0	170	P	L
Calibr.	"	"	9:09	1510700	10/17	10.0	10.0	-86	P	L
ICV	"	"	9:12	150700	10/17	10.0	10.0		P	L
CCV	J. Pitts	6/29/16	16:34	161523	3/18	4.0	4.2	155.5	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;mV pH7 Range 0 ± 50 ;mV pH 4 Range $+180 \pm 50$;mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification 6/13/16

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	Brewton	6/29/16	9:15	0.000	0.000	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater