

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

AE3

Project:

SMP-POWER Run

Everstaes

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration

or

Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

7/11/10

DO FDEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	DO mg/L	DO mg/L	%DO	Probe date	Probe gain	Pass / Fail	Lab / Field
Calibr.	Brewton	8/17/10	8:00	22.8	8.10	8.10	100	58	1.2	P	L
ICV	"	"	8:02	22.8	8.10	8.10	100			P	L
CCV	Brewton	8/17/10	15:30	24.8	8.3	8.3	100			P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond FDEP 1500	Name	Date	Time CT-ET	Temp °C	DO mg/L	%DO	Probe date	Probe gain	Pass / Fail	Lab / Field
Calibr.	Brewton	8/17/10	8:03	22.8	2000	2000	1000	1000	P	L
ICV	"	"	8:05	22.8	2000	2000	1000	1000	P	L
CCV	Brewton	8/17/10	15:32	24.8	2000	2000	1000	1000	P	L
CCV										

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1000	Name	Date	Time CT-ET	Temp °C	pH	pH	mV	Probe date	Probe gain	Pass / Fail	Lab / Field
Calibr.	Brewton	8/17/10	8:10	10.1	7.0	7.0	-20	P	L		
Calibr.	Brewton	"	8:11	10.1	4.0	4.0	151	P	L		
Calibr.	"	"	8:12	10.1	10.0	10.0	-203	P	L		
ICV	"	"	8:15	"	"	10.0	10.0	P	L		
CCV	Brewton	8/17/10	15:35	10.1	4.0	4.1		P	L		
CCV											

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

7/11/10

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass/ Fail	Lab/ Field
Pressure mode in air	Brewton	8/17/10	8:17	0.000	0.000	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

L11. GVS.

Project:

IRL/Brewer SMP

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

6/13/16

DO FDEP SOP FD 1500	Name	Date	Time ET-ET	Temp °C	DO Gain	DO Charge	% DO	DO Gain	DO Charge	Pass Fail	Lab Field
Calibr.	J. P. NTS	8/16	8:37	23.6	8.5	8.5	100	52.3	1.16	P	L
ICV	" "	" "	8:39	23.7	8.5	8.5	100	52.3	-	P	L
CCV	Brewer	8/16/16	1500	20.6	8.0	8.0	100			P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Specific Conductance FDEP 1400	Name	Date	Time ET-ET	Temp °C	Specific Conductance	Specific Conductance	Specific Conductance	Specific Conductance	Specific Conductance	Pass Fail	Lab Field
Calibr.	J. P. NTS	8/16	8:40	26.0	203	2/18	1000	1000		P	L
ICV	" "	" "	8:42	26.0	203	3/18	100	103		P	L
CCV	Brewer	8/16/16	1503	26.0	203	2/18	1000	1000		P	L
CCV											

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FD 1400	Name	Date	Time ET-ET	Temp °C	pH	pH	pH	pH	pH	Pass Fail	Lab Field
Calibr.	J. P. NTS	8/16	8:43	16.1	856	4/18	7.0	7.0	-14.6	P	L
Calibr.	" "	" "	8:45	16.1	523	3/18	4.0	4.0	163.7	P	L
Calibr.	" "	" "	8:46	15.6	700	10/17	10.0	10.0	-191.3	P	L
ICV	" "	" "	8:47	15.6	700	10/17	10.0	10.0	-191.1	P	L
CCV	Brewer	8/16/16	1505	16.1	523	3/18	4.0	4.2		P	L
CCV											

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

6/13/16

Depth Sensor (Daily)	Name	Date	Time ET-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	J. P. NTS	8/16	8:49	0.000	0.00	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater