

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

AES

Project:

North Brevard

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or

Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

7/11/16

DO FDEP SOP FT 1400	Name	Date	Time ET/ET	Temp C/F	pH Chan mV/L	Meter DO mg/L	WDO	Probe Chan	Probe Gain	Pass/ Fail	Lab Field
Calibr.	J. PINS	8/30	7:38	24.7	8.3	8.3	100	55.3	1.21	P	L
ICV	" "	" "	7:40	24.6	8.3	8.3	100	55.3	—	P	L
CCV	Brewster	8/30/16	10:00	23.1	8.6	8.7	101			P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Specific Conductance FDEP SOP FD 1400	Name	Date	Time ET/ET	Temp C/F	pH Chan mV/L	Meter DO mg/L	WDO	Probe Chan	Probe Gain	Pass/ Fail	Lab Field
Calibr.	J. PINS	8/30	7:42	26.0	8.3	2/18	1000	1000		P	L
ICV	" "	" "	7:44	26.0	8.3	3/18	100	100		P	L
CCV	Brewster	8/30/16	10:05	25.0	8.6	11/17	5900	49,915		P	L
CCV				25.1	8.3						

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FD 1400	Name	Date	Time ET/ET	Temp C/F	pH Chan mV/L	Meter DO mg/L	WDO	Probe Chan	Probe Gain	Pass/ Fail	Lab Field
Calibr.	J. PINS	8/30	7:46	16.1	8.56	4/18	7.0	7.0	-26.4	P	L
Calibr.			7:48	16.1	5.23	3/18	4.0	4.0	151.6	P	L
Calibr.			7:50	15.6	7.00	10/17	10.0	10.0	-20.34	P	L
ICV			7:52	15.6	7.00	10/17	10.0	10.0	-20.27	P	L
CCV	Brewster	8/30/16	10:10	14.5	5.23	3/18	4.0	4.1		P	L
CCV											

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

7/11/16

Depth Sensor (Daily)	Name	Date	Time ET/ET	Zero the Sensor	ICV Value	Pass/ Fail	Lab/ Field
Pressure mode in air	J. PINS	8/30	7:54	0.000	0.00	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater