

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

AE3

Project:

Taylor Creek SMP

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or

Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

7/11/16

DO FDEP SOP FT 1500	Name	Date	Time GT-ET	Temp. T	DO Conc. mg/L	Meter DO mg/L	KDO	Dose Charge	P. DO Gain	Pass / Fail	Lab / Field
Calibr.	J. Piers	9/14	8:21	21.8	8.8	8.8	100	50.2	1.14	P	L
ICV	"		8:23	21.8	8.8	8.8	100	50.2	-	P	L
CCV	in Serial	9/14	1530	25.9	8.1	8.1	100	50.2		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Conduct. FT 1200	Name	Date	Time GT-ET	Temp. T	DO Conc. mg/L	Meter DO mg/L	Standard DO mg/L	Meter DO mg/L	Pass / Fail	Lab / Field
Calibr.	J. Piers	9/14	8:24	26.06	0.3	0.288	1000	1000	P	L
ICV	"		8:26	26.03	0.32	3/18	100	103	P	L
CCV	in serial	9/14	1530	151.18	0.17		500	505	P	L
CCV										

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1000	Name	Date	Time GT-ET	Temp T	pH Buffer	pH Buffer SL	Meter Reading SL	mV	Pass / Fail	Lab / Field
Calibr.	J. Piers	9/14	8:30	161.856	4/18	7.0	7.0	-31.0	P	L
Calibr.			8:31	161.523	3/18	4.0	4.0	145.4	P	L
Calibr.			8:33	160.910	3/18	10.0	10.0	-208.0	P	L
ICV			8:34	160.910	3/18	10.0	10.0	-207.1	P	L
CCV	in serial	9/14	1530	161.523	3/18	4.0	4.1	134	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

7/11/16

Depth Sensor (Daily)	Name	Date	Time GT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	J. Piers	9/14	8:35	0.000	0.00	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

VISIT ID: 77439