

RQ-2016-10-03-13

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID:

A123

Project:

SMP

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

7/11

DO FDEP SOP FT 1500	Name	Date	Time CT-ET	Temp	DO Gain mg/L	DO mg/L	% DO	Probe check	Probe gain	Pass/ Fail	Lab/ Field
Calibr.	Matthew Sewell	10/4	9:45	22.7	8.6	8.6	100	75	1.16	P	L
ICV	Matthew Sewell	10/4	9:45	22.7	8.6	8.6	100	75		P	L
CCV	T WIDPIC	10/4	1600	24.3	8.4	8.5	102	74		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond FT 1200	Name	Date	Time CT-ET	Temp	DO Gain mg/L	DO mg/L	% DO	Probe check	Probe gain	Pass/ Fail	Lab/ Field
Calibr.	Math Sewell	10/4	9:45	22.7	2.18	2.18	1000	1000		P	L
ICV	Math Sewell	10/4	9:45	22.7	2.18	2.18	100	101		P	L
CCV	T WIDPIC	10/4	1600	26.0	2.18	2.18	1000	997		P	L
CCV											

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1000-4	Name	Date	Time CT-ET	Temp	pH Gain	pH mg/L	mV	Probe check	Probe gain	Pass/ Fail	Lab/ Field
Calibr.	Math Sewell	10/4	9:45	16.8	4.18	7.0	7.0	-35		P	L
Calibr.	Matthew Sewell	10/4	9:45	16.5	3.18	4.0	4.0	140		P	L
Calibr.	Math S	10/4	9:50	16.9	3.18	10.0	10.0	-210		P	L
ICV	Math Sewell	10/4	9:50	16.9	3.18	6.0	10.0	210		P	L
CCV	T WIDPIC	10/4	1601	16.5	3.18	4.0	4.0	141		P	L
CCV											

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;mV pH7 Range 0 $\pm$ 50;mV pH 4 Range +180 $\pm$ 50;mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

7/11

Depth Sensor (Paly)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass/ Fail	Lab/ Field
Pressure mode in air	Math Sewell	10/4	9:50	0.000	0.0	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater