

RQ-2016-03-16

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

Lil. Gue

Project:

P. Beach / Everglades

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or

Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

9/20/16

DO FDEP SOP FT 1000	Name	Date	Time CT-F	Temp C/F	DO Gain mV/L	DO Charge mV/L	DO Gain mV/L	DO Charge mV/L	DO Gain mV/L	DO Charge mV/L	Pass / Fail	Lab / Field
Calibr.	A. Hurley	10/4/16	0800	24.1	8.4	8.4	100	90	1.2		P	L
ICV				24.1	8.4	8.4	100				P	L
CCV				1600	22.9	8.6	8.7	101			P	L
CCV												

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Specific Conductance FDEP SOP FT 1000	Name	Date	Time CT-F	Temp C/F	Specific Conductance mS/cm	Specific Conductance mS/cm	Specific Conductance mS/cm	Specific Conductance mS/cm	Specific Conductance mS/cm	Specific Conductance mS/cm	Pass / Fail	Lab / Field
Calibr.	A. Hurley	10/4/16	0800	24.06003	2/18	1000	1000				P	L
ICV				24.03042	3/18	100	103				P	L
CCV				1600	26.01525	12/17	5000	5010			P	L
CCV												

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1000	Name	Date	Time CT-F	Temp C/F	pH Gain mV	pH Charge mV	pH Gain mV	pH Charge mV	pH Gain mV	pH Charge mV	Pass / Fail	Lab / Field
Calibr.	A. Hurley	10/4/16	0800	16.856	4/18	7.0	7.0	-16			P	L
Calibr.				16.1523	3/18	4.0	4.0	163			P	L
Calibr.				16.0910	3/18	10.0	10.0	-195			P	L
ICV				16.0910	3/18	10.0	10.0				P	L
CCV				1600	16.1523	3/18	4.0	4.1			P	L
CCV												

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;mV pH 4 Range $+180 \pm 50$;mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

9/20/16

Depth Sensor (Daily)	Name	Date	Time CT-F	Zero the Sensor	ICV Value	Pass/ Fail	Lab/ Field
Pressure mode in air	A. Hurley	10/4/16	0830	0.000	0.000	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater