

Chemical Analysis Report

SE-DIST-2016-07-28-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SMP-Vero/Sebastian**

Request ID: **RQ-2016-06-20-19**

Customer: **SE-DIST**

Project ID: **SMP**

Send Reports to:


FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
2199 South Rock Road
Fort Pierce, FL 34945
Attn: Matthew Sewell

For additional information please contact

Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Timothy Fitzpatrick, Program Administrator

Date Certified: 19-AUG-2016 16:16

A handwritten signature in black ink, appearing to read 'Timothy W. Fitzpatrick', with a stylized flourish at the end.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: SOUTH PRONG SEBASTIAN RIVER @ FELLSMANE RD.

Collection Date/Time: 07/27/2016 10:00

Field ID: G4SE0042

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1819849	EPA 180.1 Rev. 2.0	Turbidity	4.9		NTU	P308804	
	EPA 300.0 Rev. 2.1	Chloride	470		mg Cl/L	P309378	
		Sulfate	110		mg SO4/L	P309382	
	SM 2120 B	Color (true)	41		PCU	P308810	
	SM 2320 B-97	Total Alkalinity	183		mg CaCO3/L	P308919	
	SM 2540 C-97	TDS	1.19E+03		mg/L	P309555	
	SM 2540 D-97	TSS	9	I	mg/L	P309445	
	SM 4500 F-C-97	Fluoride	0.47		mg F/L	P308922	
1819852	EPA 350.1 Rev. 2.0	Ammonia-N	0.030		mg N/L	P308956	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.48		mg N/L	P309245	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.10		mg N/L	P309034	
	EPA 365.1 Rev. 2.0	Total-P	0.10		mg P/L	P309283	
	SM 5310 B-00	Organic Carbon	8.5		mg C/L	P309198	
1819855	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.059		mg P/L	P308824	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: NORTH CANAL @ US1

Collection Date/Time: 07/27/2016 10:35

Field ID: G5SE0043

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1819858	EPA 180.1 Rev. 2.0	Turbidity	4.7	A	NTU	P308805	
	EPA 300.0 Rev. 2.1	Chloride	190		mg Cl/L	P309378	
		Sulfate	57		mg SO4/L	P309382	
	SM 2120 B	Color (true)	71		PCU	P308810	
	SM 2320 B-97	Total Alkalinity	163		mg CaCO3/L	P308919	
	SM 2540 C-97	TDS	573		mg/L	P309555	
	SM 2540 D-97	TSS	5	I	mg/L	P309445	
	SM 4500 F-C-97	Fluoride	0.30		mg F/L	P308922	
1819859	EPA 350.1 Rev. 2.0	Ammonia-N	0.045		mg N/L	P308956	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.67		mg N/L	P309245	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.14		mg N/L	P309034	
	EPA 365.1 Rev. 2.0	Total-P	0.16		mg P/L	P309283	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P309198	
1819860	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.088		mg P/L	P308824	
1819878	EPA 200.7 Rev. 4.4	Calcium	83.2		mg/L	P308969	
		Iron	670		ug/L	P308969	
		Magnesium	20.7		mg/L	P308969	
		Potassium	6.2		mg/L	P308969	
		Sodium	102		mg/L	P308969	
		Zinc	5.0	U	ug/L	P308969	
	EPA 200.8 Rev. 5.4	Arsenic	0.89		ug/L	P308969	

Field ID: G5SE0043

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1819878	EPA 200.8 Rev. 5.4	Cadmium	0.020	U	ug/L	P308969	
		Chromium	0.78		ug/L	P308969	
		Copper	1.05		ug/L	P308969	
		Lead	0.11	I	ug/L	P308969	
		Nickel	0.35		ug/L	P308969	
		Silver	0.0050	U	ug/L	P308969	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: MAIN CANAL @ US1

Collection Date/Time: 07/27/2016 11:25

Field ID: G5SE0045

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1819850	EPA 180.1 Rev. 2.0	Turbidity	3.7		NTU	P308804	
	EPA 300.0 Rev. 2.1	Chloride	310		mg Cl/L	P309378	
		Sulfate	75		mg SO4/L	P309382	
	SM 2120 B	Color (true)	51		PCU	P308810	
	SM 2320 B-97	Total Alkalinity	182		mg CaCO3/L	P308919	
	SM 2540 C-97	TDS	798		mg/L	P309555	
	SM 2540 D-97	TSS	4	I	mg/L	P309445	
	SM 4500 F-C-97	Fluoride	0.43		mg F/L	P308922	
1819853	EPA 350.1 Rev. 2.0	Ammonia-N	0.053		mg N/L	P308956	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.65		mg N/L	P309245	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.054		mg N/L	P309034	
	EPA 365.1 Rev. 2.0	Total-P	0.096		mg P/L	P309283	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P309198	
1819856	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.052		mg P/L	P308824	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: SOUTH CANAL @ US1

Collection Date/Time: 07/27/2016 13:00

Field ID: G5SE0044

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1819851	EPA 180.1 Rev. 2.0	Turbidity	9.1		NTU	P308804	
	EPA 300.0 Rev. 2.1	Chloride	200		mg Cl/L	P309378	
		Sulfate	58		mg SO4/L	P309382	
	SM 2120 B	Color (true)	77		PCU	P308810	
	SM 2320 B-97	Total Alkalinity	191		mg CaCO3/L	P308919	
	SM 2540 C-97	TDS	609		mg/L	P309555	
	SM 2540 D-97	TSS	12		mg/L	P309445	
	SM 4500 F-C-97	Fluoride	0.36		mg F/L	P308922	
1819854	EPA 350.1 Rev. 2.0	Ammonia-N	0.11		mg N/L	P308956	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.90		mg N/L	P309245	

Field ID: G5SE0044

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1819854	EPA 353.2 Rev. 2.0	NO2NO3-N	0.085		mg N/L	P309034	
	EPA 365.1 Rev. 2.0	Total-P	0.18		mg P/L	P309283	
	SM 5310 B-00	Organic Carbon	14		mg C/L	P309198	
1819857	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.060		mg P/L	P308824	

Ref. Method and Comment:
 EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.
 SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P308804

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P308805

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
 Batch ID: P308969

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
 Batch ID: P308969

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.050	U	ug/L
Copper	0.10	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.050	U	ug/L
Silver	0.0050	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P309378

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309382

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308956

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309245

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309034

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309283

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P308824

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P308810

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P308919

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P309555

Component	Result	Code	Units
TDS	15	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2540 D-97
Batch ID: P309445

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P308922

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P309198

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P308969

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	106		P	85 - 115
Iron	105		P	85 - 115
Magnesium	107		P	85 - 115
Potassium	103		P	85 - 115
Sodium	104		P	85 - 115
Zinc	100		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P308969

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	95.8		P	85 - 115
Cadmium	98.3		P	85 - 115
Chromium	101		P	85 - 115
Copper	92.6		P	85 - 115
Lead	94.9		P	85 - 115
Nickel	93.0		P	85 - 115
Silver	94.9		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309378

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309382

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	95.2		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308956

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309245

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309034

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309283

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P308824

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P308919

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P308922

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P309198

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	102		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P308969

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P308969

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819297	Calcium	104		P	80 - 120
1819297	Iron	105		P	80 - 120
1819297	Magnesium	104		P	80 - 120
1819297	Potassium	102		P	80 - 120
1819297	Sodium	102		P	80 - 120
1819297	Zinc	97.2		P	80 - 120
1820259	Calcium	107	110	P/P	80 - 120
1820259	Iron	107	108	P/P	80 - 120
1820259	Magnesium	109	111	P/P	80 - 120
1820259	Potassium	109	110	P/P	80 - 120
1820259	Sodium	105	107	P/P	80 - 120
1820259	Zinc	100	106	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P308969

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819297	Arsenic	96.2		P	80 - 120
1819297	Cadmium	98.2		P	80 - 120
1819297	Chromium	100		P	80 - 120
1819297	Copper	88.7		P	80 - 120
1819297	Lead	96.3		P	80 - 120
1819297	Nickel	90.4		P	80 - 120
1819297	Silver	95.9		P	80 - 120
1820259	Arsenic	95.6	97.7	P/P	80 - 120
1820259	Cadmium	102	101	P/P	80 - 120
1820259	Chromium	105	104	P/P	80 - 120
1820259	Copper	92.9	94.2	P/P	80 - 120
1820259	Lead	98.2	98.0	P/P	80 - 120
1820259	Nickel	92.6	94.7	P/P	80 - 120
1820259	Silver	97.6	97.2	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309378

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1820682	Chloride	98.4		P	90 - 110
1820719	Chloride	99.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309382

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1820682	Sulfate	96.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308956

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819781	Ammonia-N	98.6	99.2	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P309245

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819816	Kjeldahl Nitrogen	91.9	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P309034

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819838	NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P309283

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819838	Total-P	106	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P308824

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819839	O-Phosphate-P	97.6	98.1	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P308922

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819631	Fluoride	98.0	98.5	P/P	91 - 105

Reference Method: SM 5310 B-00
 Batch ID: P309198

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1819637	Organic Carbon	106	106	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P308804

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819775	Turbidity	5.17	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P308805

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819858	Turbidity	0.213	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P308969

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1820259	Calcium	1.98	Spike	P	0 - 20
1820259	Iron	1.38	Spike	P	0 - 20
1820259	Magnesium	1.74	Spike	P	0 - 20
1820259	Potassium	1.58	Spike	P	0 - 20
1820259	Sodium	1.32	Spike	P	0 - 20
1820259	Zinc	5.22	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P308969

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1820259	Arsenic	2.23	Spike	P	0 - 20
1820259	Cadmium	1.73	Spike	P	0 - 20
1820259	Chromium	0.582	Spike	P	0 - 20
1820259	Copper	1.32	Spike	P	0 - 20
1820259	Lead	0.287	Spike	P	0 - 20
1820259	Nickel	2.21	Spike	P	0 - 20
1820259	Silver	0.438	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309378

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1820682	Chloride	0.0581	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309382

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1820682	Sulfate	0.177	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308956

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819781	Ammonia-N	0.426	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309245

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819816	Kjeldahl Nitrogen	5.79	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309034

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819838	NO2NO3-N	0.469	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309283

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819838	Total-P	0.0417	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P308824

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819839	O-Phosphate-P	0.340	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P308810

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819815	Color (true)	0.229	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P308919

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819631	Total Alkalinity	0.260	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P309555

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819780	TDS	1.71	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P308922

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819631	Fluoride	0.482	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P309198

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819637	Organic Carbon	0.497	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A70811

Included Lab Sample IDs: 1819849, 1819850, 1819851, 1819858

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105
Turbidity	99.5	100	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A70813

Included Lab Sample IDs: 1819849, 1819850, 1819851, 1819858

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	99.6	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A70815

Included Lab Sample IDs: 1819855, 1819856, 1819857, 1819860

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.9	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A70856

Included Lab Sample IDs: 1819849, 1819850, 1819851, 1819858

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A70856

Included Lab Sample IDs: 1819849, 1819850, 1819851, 1819858

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A70871

Included Lab Sample IDs: 1819852, 1819853, 1819854, 1819859

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.3	99.9	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A70885

Included Lab Sample IDs: 1819849, 1819850, 1819851, 1819858

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.7	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A70885

Included Lab Sample IDs: 1819849, 1819850, 1819851, 1819858

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	96.4	97.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A70894

Included Lab Sample IDs: 1819852, 1819853, 1819854, 1819859

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A70934

Included Lab Sample IDs: 1819878

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	103	102	P/P	90 - 110
Iron	102	102	P/P	90 - 110
Magnesium	103	103	P/P	90 - 110
Potassium	102	102	P/P	90 - 110
Sodium	101	102	P/P	90 - 110
Zinc	98.9	99.7	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A70940

Included Lab Sample IDs: 1819878

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Cadmium	97.5	97.7	P/P	90 - 110
Chromium	99.9	102	P/P	90 - 110
Lead	93.2	92.7	P/P	90 - 110
Silver	93.2	92.7	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A70945

Included Lab Sample IDs: 1819852, 1819853, 1819854, 1819859

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	102	100	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A70963

Included Lab Sample IDs: 1819878

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	94.7	93.7	P/P	90 - 110
Copper	91.5	91.1	P/P	90 - 110
Nickel	92.3	91.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A70992

Included Lab Sample IDs: 1819852, 1819853, 1819854, 1819859

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	101	P/P	90 - 110
Total-P	102	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A71014

Included Lab Sample IDs: 1819852, 1819853, 1819854, 1819859

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	93.0	94.9	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery				LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity							5.17	
	Turbidity							0.213	
EPA 200.7 Rev. 4.4	Calcium	106	104	107	110				1.98
	Iron	105	105	107	108				1.38
	Magnesium	107	104	109	111				1.74
	Potassium	103	102	109	110				1.58
	Sodium	104	102	105	107				1.32
	Zinc	100	97.2	100	106				5.22
EPA 200.8 Rev. 5.4	Arsenic	95.8	96.2	95.6	97.7				2.23
	Cadmium	98.3	98.2	102	101				1.73
	Chromium	101	100	105	104				0.582
	Copper	92.6	88.7	92.9	94.2				1.32
	Lead	94.9	96.3	98.2	98.0				0.287
	Nickel	93.0	90.4	92.6	94.7				2.21
	Silver	94.9	95.9	97.6	97.2				0.438
EPA 300.0 Rev. 2.1	Chloride	98.9	98.4	99.8				0.0581	
	Sulfate	95.2	96.9					0.177	
EPA 350.1 Rev. 2.0	Ammonia-N	95.9	98.6	99.2					0.426
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	91.9	102					5.79
EPA 353.2 Rev. 2.0	NO2NO3-N	105	104						0.469
EPA 365.1 Rev. 2.0	Total-P	102	106	106					0.0417
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	97.6	98.1					0.340
SM 2120 B	Color (true)							0.229	
SM 2320 B-97	Total Alkalinity	103						0.260	
SM 2540 C-97	TDS							1.71	
SM 4500 F-C-97	Fluoride	101	98.0	98.5					0.482
SM 5310 B-00	Organic Carbon	102	106	106					0.497

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1819849, 1819850, 1819851, 1819858
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	1819878
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	1819878
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1819849, 1819850, 1819851, 1819858
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1819849, 1819850, 1819851, 1819858
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1819852, 1819853, 1819854, 1819859
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1819852, 1819853, 1819854, 1819859
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1819852, 1819853, 1819854, 1819859
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1819852, 1819853, 1819854, 1819859
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1819855, 1819856, 1819857, 1819860
SM 2120 B / E31780	True Color measured at 450 nm	1819849, 1819850, 1819851, 1819858
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1819849, 1819850, 1819851, 1819858
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1819849, 1819850, 1819851, 1819858
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1819849, 1819850, 1819851, 1819858
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1819849, 1819850, 1819851, 1819858
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1819852, 1819853, 1819854, 1819859

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/28/2016			07/28/2016 14:56	Sima Feizi	1819849, 1819850, 1819851, 1819858
EPA 200.7 Rev. 4.4	07/28/2016	08/01/2016	Elliott D. Healy	08/03/2016	Joshua Ayres	1819878
EPA 200.8 Rev. 5.4	07/28/2016	08/01/2016	Elliott D. Healy	08/04/2016	Charles J. Peterson	1819878
EPA 300.0 Rev. 2.1	07/28/2016			08/06/2016	Ping Hua	1819849, 1819850, 1819851, 1819858
EPA 350.1 Rev. 2.0	07/28/2016			08/01/2016	Julie Michelle Frank	1819852, 1819853, 1819854, 1819859
EPA 351.2 Rev. 2.0	07/28/2016	08/05/2016	Sofia Elnemr	08/09/2016	Virginia P. Brown	1819852, 1819853, 1819854, 1819859
EPA 353.2 Rev. 2.0	07/28/2016			08/02/2016	Peter D. Kaus	1819852, 1819853, 1819854, 1819859
EPA 365.1 Rev. 2.0	07/28/2016	08/05/2016	Vijayalakshmi Reddy	08/08/2016	Lipi Saha	1819852, 1819853, 1819854, 1819859
EPA 365.1 Rev. 2.0 dissolved	07/28/2016			07/28/2016 16:29	Julia Storbeck	1819855
	07/28/2016			07/28/2016 16:30	Julia Storbeck	1819856
	07/28/2016			07/28/2016 16:31	Julia Storbeck	1819857
	07/28/2016			07/28/2016 16:32	Julia Storbeck	1819860
SM 2120 B	07/28/2016			07/28/2016 15:11	Blanca Fach	1819849
	07/28/2016			07/28/2016 15:12	Blanca Fach	1819850
	07/28/2016			07/28/2016 15:13	Blanca Fach	1819851, 1819858
SM 2320 B-97	07/28/2016			07/30/2016	Dale L. Simmons	1819849, 1819850, 1819851, 1819858
SM 2540 C-97	07/28/2016			08/01/2016	Sima Feizi	1819849, 1819850, 1819851, 1819858
SM 2540 D-97	07/28/2016			08/01/2016	Sima Feizi	1819849, 1819850, 1819851, 1819858
SM 4500 F-C-97	07/28/2016			07/30/2016	Dale L. Simmons	1819849, 1819850, 1819851, 1819858
SM 5310 B-00	07/28/2016			08/04/2016	Sofia Elnemr	1819852, 1819853, 1819854, 1819859