

RQ-2016-05-09-75

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: Biodex #2Project: SMF

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Curtis Musson	5/11/16	0719	22.65	8.61	8.64	100.0	46.1	-	P	L
ICV	Curtis Musson	5/11/16	0720	22.66	8.61	8.64	100.0	46.1	-	P	L
CCV	Curtis Musson	5/11/16		23.32	8.53	14.76	173.00	51.2	-	F	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.**Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.**Optical:** DO gain range 0.85 to 1.15; DO charge N/A.**Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Curtis Musson	5/11/16	0723	CC14162	10/15/16	996	996	P	L
ICV	Curtis Musson	5/11/16	0724	CC14162	10/15/16	996	996	P	L
CCV	Curtis Musson	5/11/16	15:01	CC13882	7/15/16	1061	97	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Curtis Musson	5/11/16	0726	153035	07/20/17	7.0	7.00	70.4	P	L
Calibr.	Curtis Musson	5/11/16	0730	154634	07/20/17	4.0	4.00	165.3	P	L
Calibr.	Curtis Musson	5/11/16	0732	160330C	10/20/17	10.0	10.00	79.2	P	L
ICV	Curtis Musson	5/11/16	0735	154634	07/20/17	4.00	4.01	164.7	P	L
CCV	Curtis Musson	5/11/16	1504	153035	07/20/17	7.00	7.01	70.5	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50;mV pH 4 Range +180 $\pm$ 50;mV pH 10 Range -180 $\pm$ 50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mVDepth (Quarterly)

Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater