

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter #4 RQ- 2017-01-30-10 Project Swansee N 2

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site N

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 03/04/2016 11/16

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	JPm	11/31/17	710	19.2	9.24	9.22	100	55.3	0.934	P	
ICV	J	2	712	19.3	9.22	9.21	100	55.3	0.934	P	
CCV	AKalmbacher	1-31-17	1525	20.9	8.93	9.01	100.9	55.3	0.934	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	JPm	11/31/17	713	161101A	11/17	1000	1000	P	
ICV	J	2	715	160616B	12/16	100	103	P	
CCV	AKalmbacher	1-31-17	1535	161101A	11/17	999	999	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	JPm	11/31/17	716	2608904	7/18	7.0	7.0	-	P	
Calibr.	J		717	2605918	4/18	4.0	4.0	168.3	P	
Calibr.	J		718	2607973	1/18	10.0	10.0	182.7	P	
ICV	J		719	2605918	4/18	4.0	4.04	165.9	P	
CCV	AKalmbacher	1-31-17	1538	2608904	7/18	7.00	6.99	-7.7	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification 11/16

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	JP	11/31/17	724	0.000	0.00	P	

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater