

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter YSI #4

RQ- 2017-01-30-14

Project SW

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

2/2017

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J Parash	2/13/17	707	22.0	8.74	8.74	100	45.1		P	L
ICV	J	J	708	22.0	8.72	8.72	99.3	44.1		P	L
CCV	J Parash	2/13/17	1435	22.8	8.61	7.65	89	41.0		F	Had bubbles
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J Par	2/13/17	710	161101A	11/17	1000	1000	P	
ICV	J	J	712	12/17	21421017	1000	144	P	
CCV	J Par	2/13/17	1424	161101B	11/17	100	100	P	
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J Par	2/13/17	714	2608904	7/18	7.0	7.00	-9.3	P	L
Calibr.	J	J	716	2605918	4/18	4.0	4.0	164.8	P	L
Calibr.	J	J	718	2607173	11/18	10.0	10.0	-183.9	P	L
ICV	J	J	724	2607173	11/18	10.0	10.0	-183.6	P	L
CCV	J Par	2/13/17	1425	2608904	7/18	7.0	7.08	-14.0	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

2/2017

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	J	2/13/17		0.000	0.00	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater