

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter YSI #4 RQ- 2017-07-25-41 Project NSR

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site L-6

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 6/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	T. Barnish	9/26/17	725	23.2	8.54	8.52	100	57.4	1.0	P	L
ICV	J	J	727	23.3	8.53	8.51	100	57.4	1.0	P	L
CCV	T. Barnish	9/26/17	1325	24.6	8.43	8.32	101.3	57.4	1.0	L	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	T. Barnish	9/26/17	729	170720A	8/18	1000	1000	P	L
ICV	T. Barnish	9/26/17	731	81472017	3/18	147	151	P	L
CCV	J	J	1327	170720A	8/18	1000	990	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	T. Barnish	9/26/17	737	2707256	6/19	7.0	7.0	-23.4	P	L
Calibr.	J	J	742	2707222	6/19	4.0	4.0	152.5	P	L
Calibr.	J	J	746	2707298	9/18	10.0	9.99	-196.1	P	L
ICV	J	J	754	2707256	6/19	7.0	7.0	-21.9	P	L
CCV	J	J	1325	2707222	6/19	4.0	4.0	-152.2	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly) Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

T. Barnish 10/2/17