

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: Bathy 167

Project: SWP

RQ:

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 02/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>Leif Boman</u>	<u>5/8/17</u>	<u>8:30</u>	<u>23.9</u>	<u>8.4</u>	<u>8.4</u>	<u>100</u>	<u>43</u>	<u>1.0</u>	<u>P</u>	<u>L</u>
ICV	<u>1</u>	<u>1</u>	<u>1</u>	<u>23.9</u>	<u>8.4</u>	<u>8.4</u>	<u>100</u>	<u>44</u>		<u>P</u>	<u>L</u>
CCV	<u>1</u>	<u>1</u>	<u>16:00</u>	<u>26.6</u>	<u>8.0</u>	<u>8.0</u>	<u>99</u>	<u>44</u>	<u>1.0</u>	<u>P</u>	<u>L</u>
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>Leif Boman</u>	<u>5/8/17</u>	<u>8:30</u>	<u>170220</u>	<u>03/18</u>	<u>1000</u>	<u>1000</u>	<u>P</u>	<u>L</u>
ICV	<u>1</u>	<u>1</u>	<u>1</u>	<u>2611962</u>	<u>10/18</u>	<u>100</u>	<u>103</u>	<u>P</u>	<u>L</u>
CCV	<u>1</u>	<u>1</u>	<u>16:00</u>	<u>2611962</u>	<u>10/18</u>	<u>100</u>	<u>103</u>	<u>P</u>	<u>L</u>
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>Leif Boman</u>	<u>5/8/17</u>	<u>8:30</u>	<u>165674</u>	<u>08/18</u>	<u>7.0</u>	<u>7.0</u>	<u>-35</u>	<u>P</u>	<u>L</u>
Calibr.	<u>1</u>	<u>1</u>	<u>1</u>	<u>166544</u>	<u>10/18</u>	<u>4.0</u>	<u>4.0</u>	<u>139</u>	<u>P</u>	<u>L</u>
Calibr.	<u>1</u>	<u>1</u>	<u>1</u>	<u>165833</u>	<u>08/18</u>	<u>10.0</u>	<u>10.0</u>	<u>-209</u>	<u>P</u>	<u>L</u>
ICV	<u>1</u>	<u>1</u>	<u>1</u>	<u>166544</u>	<u>10/18</u>	<u>4.0</u>	<u>4.0</u>	<u>139</u>	<u>P</u>	<u>L</u>
CCV	<u>1</u>	<u>1</u>	<u>16:00</u>	<u>165833</u>	<u>08/18</u>	<u>10.0</u>	<u>10.0</u>	<u>-208</u>	<u>P</u>	<u>L</u>
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: B00P NON

Project: SMP/BMAP

RQ: 2017-05-08-37

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site 2/17

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Mike Vreman	5/8	820	23.2	8.5	8.5	100	46	1.1	P	L
ICV				23.3	8.5	8.5	100	46	1.1	P	L
CCV			1300	23.4	8.5	8.5	100	46	1.1	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.			820	170220	3/18	1000	1000	P	L
ICV				2611962	10/18	100	100	P	L
CCV				2611962	10/18	100	101	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.			820	165674	8/18	7.0	7.0	-20	P	L
Calibr.				166544	10/18	4.0	4.0	158	P	L
Calibr.				165833	8/18	10.0	10.0	1990	P	L
ICV				165833	8/18	10.0	10.0	-199	P	L
CCV			1300	165833	8/18	10.0	10.0	-199	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater