

Quarterly Depth Verification Regional Operation Centers

SOP - S&T Sampling Manual and ROC Training Manual
Record depth reading with only one rounded decimal figure
Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up

Meter ID#: <u>Betty/67</u>	Date: <u>3/3/17</u>	Time: <u>1125</u>
Location: <u>Lab</u>		
Date of Last Verification: <u>11/12/16</u>		
Depth using verified weighted line marked in increments of 0.1 m: <u>7</u> m		
Depth read with the meter: <u>7</u> m		
Pass/Fail: <u>Pass</u> (acceptance Criteria 10%)		
Person Performing Verification: <u>Mike Drennan</u>		

- meter cable marked to 7m, in 10ths, double marked every meter

Secchi/Weighted Line ID#: <u>Lead-Foot</u>	Date: <u>3/2/17</u>	Time: <u>1420</u>
Location: <u>Lab</u>		
Date of Last Verification: <u>11/12/16</u>		
Incremental markings of 0.1 m checked <u>Yes</u> yes/no		
Pass/Fail: <u>Pass</u> (acceptable criteria 10%) New increments done: <u>No</u> yes/no/NA		
Total Length of Line marked in 0.1 m increments: <u>10</u> m		
Reference meter stick or metal measuring tape: <u>5</u> m		
Pass/Fail: <u>Pass</u> (acceptable criteria of 5%) New increments done: <u>No</u> yes/no/NA		
Person Performing Verification: <u>Mike Drennan</u>		

Secchi/Weighted Line ID#: <u>Knot-head</u>	Date: <u>3/3/17</u>	Time: <u>1045</u>
Location: <u>Lab</u>		
Date of Last Verification: <u>11/12/16</u>		
Incremental markings of 0.1 m checked <u>Yes</u> yes/no		
Pass/Fail: <u>Pass</u> (acceptable criteria 10%) New increments done: <u>Yes</u> yes/no/NA		
Total Length of Line marked in 0.1 m increments: <u>10</u> m		
Reference meter stick or metal measuring tape: <u>5</u> m		
Pass/Fail: <u>Pass</u> (acceptable criteria of 5%) New increments done: <u>No</u> yes/no/NA		
Person Performing Verification: <u>Mike Drennan</u>		

Comments: New rope, New markings (Knot-head only)
Blue-red-Blue is 10m

Figure 25. Quarterly Depth Measuring Device Verification Log.

This figure is provided as an example only. Please visit the [Watershed Monitoring Information Center](#) to download the most recent version.

Quarterly Depth Verification Regional Operation Centers			
SOP - S&T Sampling Manual and ROC Training Manual Record depth reading with only one rounded decimal figure Numbers ≤ 4 are rounded down; numbers ≥ 5 are rounded up			
Meter ID#: <u>Boop/Non</u>	Date: <u>3/3/17</u>	Time: <u>1135</u>	
Location: <u>Lab</u>	Date of Last Verification: <u>11/12/16</u>		
Depth using verified weighted line marked in increments of 0.1 m: <u>7</u> m		- meter cable marked to 7m, in 10ths, double marked every meter	
Depth read with the meter: <u>7</u> m			
Pass/Fail: <u>Pass</u> (acceptance Criteria 10%)			
Person Performing Verification: <u>Mike Drennon</u>			
Secchi/Weighted Line ID#: <u>Rusty</u> Date: <u>3/3/17</u> Time: <u>1115</u>			
Location: <u>Lab</u>		Date of Last Verification: _____	
Incremental markings of 0.1 m checked <u>Yes</u> yes/no			
Pass/Fail: <u>Pass</u> (acceptable criteria 10%)		New increments done: <u>No</u> yes/no/NA	
Total Length of Line marked in 0.1 m increments: <u>5.8</u> m			
Reference meter stick or metal measuring tape: <u>5</u> m			
Pass/Fail: <u>Pass</u> (acceptable criteria of 5%)		New increments done: <u>No</u> yes/no/NA	
Person Performing Verification: <u>Mike Drennon</u>			
Secchi/Weighted Line ID#: _____ Date: _____ Time: _____			
Location: <u>Lab</u>		Date of Last Verification: _____	
Incremental markings of 0.1 m checked _____ yes/no			
Pass/Fail: _____ (acceptable criteria 10%)		New increments done: _____ yes/no/NA	
Total Length of Line marked in 0.1 m increments: _____ m			
Reference meter stick or metal measuring tape: _____ m			
Pass/Fail: _____ (acceptable criteria of 5%)		New increments done: _____ yes/no/NA	
Person Performing Verification: _____			
Comments: _____			

Page _____
Form effective April 1, 2015