

Chemical Analysis Report

CEN-DIST-2017-03-31-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Minnie / Charm**
Request ID: **RQ-2017-03-27-48**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Michael Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 21-APR-2017 12:42

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light-colored rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: LAKE CHAM @ CENTER

Collection Date/Time: 03/30/2017 10:30

Field ID: G1CE0051

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1884979	EPA 180.1 Rev. 2.0	Turbidity	3.2		NTU	P322687	
	EPA 300.0 Rev. 2.1	Chloride	60		mg Cl/L	P323118	
		Sulfate	29		mg SO4/L	P323121	
		Color (true)	24	A	PCU	P322589	
	SM 2320 B-97	Total Alkalinity	51		mg CaCO3/L	P323004	
	SM 2540 C-97	TDS	207		mg/L	P322986	
	SM 2540 D-97	TSS	3	I	mg/L	P323023	
	SM 4500 F-C-97	Fluoride	0.25		mg F/L	P323007	
1884980	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P322607	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.90		mg N/L	P322633	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P322721	
	EPA 365.1 Rev. 2.0	Total-P	0.033		mg P/L	P322706	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P322613	
1884981	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P322609	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: LAKE MINNIE @ CENTER

Collection Date/Time: 03/30/2017 11:30

Field ID: G2CE0108

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1884976	EPA 180.1 Rev. 2.0	Turbidity	8.8	A	NTU	P322687	
	EPA 300.0 Rev. 2.1	Chloride	53		mg Cl/L	P323118	
		Sulfate	3.4		mg SO4/L	P323121	
		Color (true)	160		PCU	P322589	
	SM 2320 B-97	Total Alkalinity	47		mg CaCO3/L	P323004	
	SM 2540 C-97	TDS	184	A	mg/L	P322986	
	SM 2540 D-97	TSS	8	IA	mg/L	P323023	
	SM 4500 F-C-97	Fluoride	0.073	I	mg F/L	P323007	
1884977	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P322607	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.3		mg N/L	P322633	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P322721	
	EPA 365.1 Rev. 2.0	Total-P	0.042		mg P/L	P322706	
	SM 5310 B-00	Organic Carbon	22		mg C/L	P322613	
1884978	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P322609	
1884989	EPA 200.7 Rev. 4.4	Calcium	22.2		mg/L	P322663	
		Iron	200		ug/L	P322663	
		Magnesium	3.69		mg/L	P322663	
		Potassium	1.8		mg/L	P322663	
		Sodium	28.3		mg/L	P322663	
		Zinc	5.0	U	ug/L	P322663	
	EPA 200.8 Rev. 5.4	Arsenic	0.81		ug/L	P322663	

Field ID: G2CE0108

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1884989	EPA 200.8 Rev. 5.4	Cadmium	0.020	U	ug/L	P322663	
		Chromium	0.40	U	ug/L	P322663	
		Copper	0.75	I	ug/L	P322663	
		Lead	0.24		ug/L	P322663	
		Nickel	0.21	I	ug/L	P322663	
		Silver	0.010	U	ug/L	P322663	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: LAKE MINNIE @ CENTER

Collection Date/Time: 03/30/2017 11:30

Field ID: G2CE0108

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1884985	EPA 8321B	Glyphosate**	10	U	ug/L	P322576	
		Sucralose**	0.094		ug/L	P322530	
	USGS O-2060-01	2,4-D	0.0040	U	ug/L	P322600	
		Diuron	0.0040	U	ug/L	P322600	
		Fenuron	0.0080	U	ug/L	P322600	
		Imidacloprid	0.0040	U	ug/L	P322600	
		Bentazon	0.031		ug/L	P322600	MS
		Linuron	0.0040	U	ug/L	P322600	
		Acetaminophen**	0.010	U	ug/L	P322600	
		MCP	0.0040	U	ug/L	P322600	
		Carbamazepine**	4.0E-04	U	ug/L	P322600	
		Triclopyr**	0.010	U	ug/L	P322600	
		Primidone**	0.0080	U	ug/L	P322600	RPD
		Fluridone**	0.0016	I	ug/L	P322600	
1884986	EPA 8321B	Pyraclorobin**	0.0020	U	ug/L	P322599	
1884987	EPA 8270D	Aldrin	4.2	U	ng/L	P322674	
		Alpha-BHC	2.1	U	ng/L	P322674	
		Beta-BHC	4.2	U	ng/L	P322674	
		Delta-BHC	4.2	U	ng/L	P322674	
		Gamma-BHC	4.2	U	ng/L	P322674	
		Carbophenothion	4.2	U	ng/L	P322674	
		Chlorothalonil	2.1	U	ng/L	P322674	
		Cypermethrin	21	U	ng/L	P322674	
		DDD-p,p'	2.1	U	ng/L	P322674	
		DDE-p,p'	3.2	U	ng/L	P322674	
		DDT-p,p'	3.2	U	ng/L	P322674	
		Dieldrin	4.2	U	ng/L	P322674	
		Endosulfan I	4.2	U	ng/L	P322674	
		Endosulfan II	4.2	U	ng/L	P322674	
		Endosulfan Sulfate	2.1	U	ng/L	P322674	
		Endrin	2.1	U	ng/L	P322674	

Field ID: G2CE0108

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1884987	EPA 8270D	Endrin Aldehyde	2.1	U	ng/L	P322674	MS
		Heptachlor	1.6	U	ng/L	P322674	
		Heptachlor Epoxide	2.1	U	ng/L	P322674	
		Methoxychlor	4.2	U	ng/L	P322674	
		Mirex	2.1	U	ng/L	P322674	
		Permethrin	11	U	ng/L	P322674	
		Trifluralin	1.1	U	ng/L	P322674	
		Alpha-Chlordane	1.1	U	ng/L	P322674	
		Gamma-Chlordane	1.1	U	ng/L	P322674	
		Endrin Ketone	2.1	U	ng/L	P322674	
		Dicofol	32	U	ng/L	P322674	
		Chlordane**	2.6	U	ng/L	P322674	
		Toxaphene**	13	U	ng/L	P322674	
1884988	EPA 8270D	Ametryn	0.32	U	ng/L	P322762	RPD
		Atrazine	29		ng/L	P322762	
		Chlorpyrifos Ethyl	0.25	U	ng/L	P322762	
		Chlorpyrifos Methyl	0.10	U	ng/L	P322762	
		Disulfoton	0.32	U	ng/L	P322762	
		Ethion	0.15	UJ	ng/L	P322762	
		Ethoprop	0.10	U	ng/L	P322762	
		Hexazinone	0.50	U	ng/L	P322762	
		Malathion	0.35	U	ng/L	P322762	
		Metribuzin	0.20	U	ng/L	P322762	
		Mevinphos	0.50	U	ng/L	P322762	
		Norflurazon	0.50	U	ng/L	P322762	
		Phorate	0.25	U	ng/L	P322762	
		Prometryn	0.20	U	ng/L	P322762	
		Simazine	0.50	U	ng/L	P322762	
		Atrazine Desethyl	1.7	I	ng/L	P322762	
		Prometon	0.90	U	ng/L	P322762	
		Fipronil	0.25	U	ng/L	P322762	
		Molinate	0.30	U	ng/L	P322762	
		Pendimethalin	1.5	U	ng/L	P322762	
		Terbufos	0.10	U	ng/L	P322762	
		EPTC	1.2	U	ng/L	P322762	
		Terbuthylazine	0.42	U	ng/L	P322762	
		Bromacil	0.50	U	ng/L	P322762	
		Fipronil Sulfide	0.33	I	ng/L	P322762	
		Fipronil Sulfone	0.44	I	ng/L	P322762	
		Alachlor	0.25	U	ng/L	P322762	
		Azinphos Methyl	2.2	U	ng/L	P322762	
		Butylate	0.40	U	ng/L	P322762	
		Demeton	1.0	U	ng/L	P322762	
		Diazinon	0.12	U	ng/L	P322762	
		Fenamiphos	0.25	U	ng/L	P322762	

Field ID: G2CE0108

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1884988	EPA 8270D	Fonofos	0.20	U	ng/L	P322762	
		Metalaxyl	0.20	U	ng/L	P322762	
		Metolachlor	0.32	I	ng/L	P322762	
		Acetochlor	0.25	U	ng/L	P322762	
		Cyanazine	0.50	U	ng/L	P322762	
		Parathion Ethyl	0.25	U	ng/L	P322762	
		Parathion Methyl	0.25	U	ng/L	P322762	

Ref. Method and Comment:

EPA 8270D: MS accuracy for Atrazine not assessed due to high content of this parameter in the sample spiked. Refer to the narrative for an explanation of QC Codes.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P322687

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P322663

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P322663

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.20	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.20	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P323118

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P323121

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P322607

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P322633

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322721

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P322706

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322609

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270D
Batch ID: P322674

Component	Result	Code	Units
Aldrin	4.0	U	ng/L
Alpha-BHC	2.0	U	ng/L
Alpha-Chlordane	1.0	U	ng/L
Beta-BHC	4.0	U	ng/L
Carbophenothion	4.0	U	ng/L
Chlorothalonil	2.0	U	ng/L
Cypermethrin	20	U	ng/L
DDD-p,p'	2.0	U	ng/L
DDE-p,p'	3.0	U	ng/L
DDT-p,p'	3.0	U	ng/L
Delta-BHC	4.0	U	ng/L
Dicofol	30	U	ng/L
Dieldrin	4.0	U	ng/L
Endosulfan I	4.0	U	ng/L
Endosulfan II	4.0	U	ng/L
Endosulfan Sulfate	2.0	U	ng/L
Endrin	2.0	U	ng/L
Endrin Aldehyde	2.0	U	ng/L
Endrin Ketone	2.0	U	ng/L
Gamma-BHC	4.0	U	ng/L
Gamma-Chlordane	1.0	U	ng/L
Heptachlor	1.5	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270D

Batch ID: P322674

Component	Result	Code	Units
Heptachlor Epoxide	2.0	U	ng/L
Methoxychlor	4.0	U	ng/L
Mirex	2.0	U	ng/L
Permethrin	10	U	ng/L
Trifluralin	1.0	U	ng/L

Reference Method: EPA 8270D

Batch ID: P322762

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.32	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	2.2	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.10	U	ng/L
Cyanazine	0.50	U	ng/L
Demeton	1.0	U	ng/L
Diazinon	0.12	U	ng/L
Disulfoton	0.32	U	ng/L
EPTC	1.2	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Fenamiphos	0.25	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Malathion	0.35	U	ng/L
Metaxyl	0.20	U	ng/L
Metolachlor	0.25	U	ng/L
Metribuzin	0.20	U	ng/L
Mevinphos	0.50	U	ng/L
Molinate	0.30	U	ng/L
Norflurazon	0.50	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.25	U	ng/L
Pendimethalin	1.5	U	ng/L
Phorate	0.25	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Simazine	0.50	U	ng/L
Terbufos	0.10	U	ng/L
Terbuthylazine	0.42	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8276
Batch ID: P322674

Component	Result	Code	Units
Chlordane	2.5	U	ng/L
Toxaphene	12	U	ng/L

Reference Method: EPA 8321B
Batch ID: P322530

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Reference Method: EPA 8321B
Batch ID: P322576

Component	Result	Code	Units
Glyphosate	10	U	ug/L

Reference Method: EPA 8321B
Batch ID: P322599

Component	Result	Code	Units
Pyraclostrobin	0.0020	U	ug/L

Reference Method: SM 2120 B
Batch ID: P322589

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P323004

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P322986

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P323023

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P323007

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report Method Blank Results

Reference Method: SM 5310 B-00
Batch ID: P322613

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: USGS O-2060-01
Batch ID: P322600

Component	Result	Code	Units
2,4-D	0.0040	U	ug/L
Acetaminophen	0.010	U	ug/L
Bentazon	0.0040	U	ug/L
Carbamazepine	4.0E-04	U	ug/L
Diuron	0.0040	U	ug/L
Fenuron	0.0080	U	ug/L
Fluridone	4.0E-04	U	ug/L
Imidacloprid	0.0040	U	ug/L
Linuron	0.0040	U	ug/L
MCP	0.0040	U	ug/L
Primidone	0.0080	U	ug/L
Triclopyr	0.010	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P322663

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	102		P	85 - 115
Iron	104		P	85 - 115
Magnesium	103		P	85 - 115
Potassium	102		P	85 - 115
Sodium	102		P	85 - 115
Zinc	99.7		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P322663

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	96.9		P	85 - 115
Cadmium	93.7		P	85 - 115
Chromium	96.6		P	85 - 115
Copper	95.4		P	85 - 115
Lead	99.2		P	85 - 115
Nickel	97.6		P	85 - 115
Silver	98.1		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P323118

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P323121

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P322607

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P322633

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322721

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P322706

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	107		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322609

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.4		P	90 - 110

Reference Method: EPA 8270D
Batch ID: P322674

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	46.9	62.0	P/P	30 - 83.0
Alpha-BHC	86.3	85.9	P/P	62 - 119
Alpha-Chlordane	92.8	94.9	P/P	56 - 119
Beta-BHC	88.8	83.0	P/P	54 - 141
Carbophenothion	96.6	90.2	P/P	33 - 147
Chlorothalonil	118	102	P/P	25 - 175
Cypermethrin	89.8	89.2	P/P	50 - 155
DDD-p,p'	92.7	91.7	P/P	56 - 135
DDE-p,p'	98.8	99.2	P/P	48 - 120
DDT-p,p'	85.2	87.9	P/P	51 - 121
Delta-BHC	105	97.3	P/P	65 - 163
Dicofol	82.4	89.3	P/P	15 - 112
Dieldrin	93.9	93.9	P/P	59 - 134
Endosulfan I	97.1	97.7	P/P	65 - 142
Endosulfan II	93.2	91.0	P/P	65 - 154
Endosulfan Sulfate	85.8	81.3	P/P	65 - 136

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270D

Batch ID: P322674

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Endrin	103	104	P/P	53 - 156
Endrin Aldehyde	93.5	92.2	P/P	47 - 156
Endrin Ketone	78.8	76.9	P/P	70 - 124
Gamma-BHC	91.1	87.1	P/P	64 - 142
Gamma-Chlordane	92.3	92.5	P/P	55 - 114
Heptachlor	84.5	84.4	P/P	42 - 105
Heptachlor Epoxide	92.7	93.0	P/P	67 - 120
Methoxychlor	90.9	81.1	P/P	63 - 134
Mirex	81.4	80.9	P/P	40 - 90.0
Permethrin	83.1	83.2	P/P	36 - 144
Trifluralin	116	108	P/P	40 - 170

Reference Method: EPA 8270D

Batch ID: P322762

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	92.7	90.5	P/P	78 - 118
Alachlor	93.2	90.9	P/P	77 - 119
Ametryn	113	106	P/P	61 - 136
Atrazine	104	96.6	P/P	73 - 120
Atrazine Desethyl	84.2	75.4	P/P	16 - 122
Azinphos Methyl	144	149	P/P	69 - 186
Bromacil	99.7	76.5	P/P	40 - 125
Butylate	72.2	54.6	P/P	32 - 87.0
Chlorpyrifos Ethyl	103	94.1	P/P	53 - 110
Chlorpyrifos Methyl	97.2	93.1	P/P	31 - 119
Cyanazine	121	116	P/P	36 - 224
Demeton	104	98.2	P/P	39 - 122
Diazinon	90.6	89.9	P/P	75 - 119
Disulfoton	117	115	P/P	42 - 139
EPTC	70.6	56.2	P/P	33 - 96.0
Ethion	119	96.5	P/P	63 - 127
Ethoprop	93.0	91.6	P/P	65 - 107
Fenamiphos	122	113	P/P	64 - 137
Fipronil	103	97.4	P/P	60 - 126
Fipronil Sulfide	102	95.4	P/P	58 - 124
Fipronil Sulfone	96.2	88.0	P/P	57 - 126
Fonofos	91.0	89.2	P/P	65 - 111
Hexazinone	96.9	89.0	P/P	46 - 151
Malathion	117	109	P/P	68 - 132
Metalaxyl	93.9	89.5	P/P	51 - 132
Metolachlor	102	96.6	P/P	78 - 119
Metribuzin	125	128	P/P	64 - 138
Mevinphos	114	110	P/P	34 - 128
Molinate	75.6	68.0	P/P	44 - 101
Norflurazon	115	110	P/P	63 - 129
Parathion Ethyl	99.9	91.1	P/P	80 - 109
Parathion Methyl	85.1	87.5	P/P	74 - 125
Pendimethalin	103	94.4	P/P	48 - 134
Phorate	102	97.0	P/P	52 - 119
Prometon	121	111	P/P	69 - 124
Prometryn	113	108	P/P	66 - 124

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270D
Batch ID: P322762

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Simazine	99.8	95.0	P/P	61 - 134
Terbufos	85.7	83.8	P/P	58 - 115
Terbutylazine	93.8	90.5	P/P	77 - 113

Reference Method: EPA 8276
Batch ID: P322674

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	91.0	91.8	P/P	42 - 113
Toxaphene	92.3	98.8	P/P	40 - 120

Reference Method: EPA 8321B
Batch ID: P322530

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	99.6		P	40 - 140

Reference Method: EPA 8321B
Batch ID: P322576

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Glyphosate	131		P	40 - 140

Reference Method: EPA 8321B
Batch ID: P322599

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Pyraclostrobin	112		P	40 - 140

Reference Method: SM 2320 B-97
Batch ID: P323004

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	105		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P323007

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.4		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P322613

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.8		P	85 - 115

Reference Method: USGS O-2060-01
Batch ID: P322600

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	104		P	40 - 140

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: USGS O-2060-01
Batch ID: P322600

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetaminophen	53.2		P	40 - 140
Bentazon	83.6		P	40 - 140
Carbamazepine	121		P	40 - 140
Diuron	87.6		P	40 - 140
Fenuron	98.0		P	40 - 140
Fluridone	75.6		P	40 - 140
Imidacloprid	85.6		P	40 - 160
Linuron	92.4		P	40 - 140
MCPD	105		P	40 - 140
Primidone	86.8		P	40 - 140
Triclopyr	94.8		P	40 - 140

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P322663

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884303	Calcium	100		P	80 - 120
1884303	Iron	104		P	80 - 120
1884303	Magnesium	101		P	80 - 120
1884303	Potassium	101		P	80 - 120
1884303	Sodium	102		P	80 - 120
1884303	Zinc	98.8		P	80 - 120
1884721	Calcium	99.5		P	80 - 120
1884721	Iron	103	104	P/P	80 - 120
1884721	Magnesium	98.4		P	80 - 120
1884721	Potassium	103	103	P/P	80 - 120
1884721	Sodium	101		P	80 - 120
1884721	Zinc	96.8	98.0	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P322663

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884303	Arsenic	96.6		P	80 - 120
1884303	Cadmium	92.5		P	80 - 120
1884303	Chromium	95.1		P	80 - 120
1884303	Copper	92.2		P	80 - 120
1884303	Lead	97.9		P	80 - 120
1884303	Nickel	93.6		P	80 - 120
1884303	Silver	98.1		P	80 - 120
1884721	Arsenic	98.8	97.2	P/P	80 - 120
1884721	Cadmium	96.3	95.2	P/P	80 - 120
1884721	Chromium	98.1	96.7	P/P	80 - 120
1884721	Copper	92.9	92.5	P/P	80 - 120
1884721	Lead	98.3	98.3	P/P	80 - 120
1884721	Nickel	95.4	94.0	P/P	80 - 120
1884721	Silver	99.2	98.8	P/P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P323118

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885051	Chloride	98.7		P	90 - 110
1885069	Chloride	99.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P323121

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885051	Sulfate	95.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P322607

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884977	Ammonia-N	106	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P322633

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884960	Kjeldahl Nitrogen	103	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322721

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884980	NO2NO3-N	99.5	98.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P322706

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885054	Total-P	108	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322609

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884981	O-Phosphate-P	97.0	98.4	P/P	90 - 110

Reference Method: EPA 8270D
Batch ID: P322674

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885130	Aldrin	71.7	72.5	P/P	30 - 83.0
1885130	Alpha-BHC	88.2	84.9	P/P	62 - 119
1885130	Alpha-Chlordane	98.7	97.0	P/P	56 - 119
1885130	Beta-BHC	91.9	105	P/P	54 - 141
1885130	Carbophenothion	122	116	P/P	33 - 147
1885130	Chlorothalonil	60.9	80.9	P/P	25 - 175
1885130	Cypermethrin	101	89.9	P/P	50 - 155
1885130	DDD-p,p'	93.9	86.8	P/P	56 - 135
1885130	DDE-p,p'	103	99.8	P/P	48 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270D

Batch ID: P322674

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885130	DDT-p,p'	92.9	84.0	P/P	51 - 121
1885130	Delta-BHC	133	109	P/P	65 - 163
1885130	Dicofol	89.1	97.1	P/P	15 - 112
1885130	Dieldrin	98.8	97.5	P/P	59 - 134
1885130	Endosulfan I	111	122	P/P	65 - 142
1885130	Endosulfan II	108	105	P/P	65 - 154
1885130	Endosulfan Sulfate	95.6	87.7	P/P	65 - 136
1885130	Endrin	103	101	P/P	53 - 156
1885130	Endrin Aldehyde	56.3	58.9	P/P	47 - 156
1885130	Endrin Ketone	89.1	84.8	P/P	70 - 124
1885130	Gamma-BHC	112	116	P/P	64 - 142
1885130	Gamma-Chlordane	98.7	95.5	P/P	55 - 114
1885130	Heptachlor	81.8	80.5	P/P	42 - 105
1885130	Heptachlor Epoxide	98.6	96.2	P/P	67 - 120
1885130	Methoxychlor	96.4	85.2	P/P	63 - 134
1885130	Mirex	91.2	83.7	F/P	40 - 90.0
1885130	Permethrin	99.1	91.8	P/P	36 - 144
1885130	Trifluralin	140	136	P/P	40 - 170

Reference Method: EPA 8270D

Batch ID: P322762

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884988	Acetochlor	94.5	94.2	P/P	74 - 123
1884988	Alachlor	98.7	98.5	P/P	73 - 124
1884988	Ametryn	128	130	P/P	56 - 149
1884988	Atrazine Desethyl	67.8	56.4	P/P	12 - 103
1884988	Azinphos Methyl	185	171	P/P	60 - 202
1884988	Bromacil	83.0	83.3	P/P	42 - 133
1884988	Butylate	68.8	57.0	P/P	10 - 85.0
1884988	Chlorpyrifos Ethyl	107	98.9	P/P	53 - 113
1884988	Chlorpyrifos Methyl	95.5	88.2	P/P	40 - 115
1884988	Cyanazine	120	118	P/P	22 - 210
1884988	Demeton	115	107	P/P	17 - 149
1884988	Diazinon	97.8	97.8	P/P	72 - 124
1884988	Disulfoton	137	130	P/P	43 - 139
1884988	EPTC	70.3	60.4	P/P	10 - 86.0
1884988	Ethion	108	75.3	P/P	65 - 131
1884988	Ethoprop	105	99.6	P/P	59 - 110
1884988	Fenamiphos	89.9	89.3	P/P	57 - 138
1884988	Fipronil	103	101	P/P	40 - 130
1884988	Fipronil Sulfide	96.1	86.2	P/P	36 - 128
1884988	Fipronil Sulfone	86.3	79.1	P/P	16 - 145
1884988	Fonofos	103	94.2	P/P	60 - 112
1884988	Hexazinone	93.8	95.6	P/P	69 - 142
1884988	Malathion	113	106	P/P	68 - 132
1884988	Metalaxyl	94.7	96.7	P/P	66 - 118
1884988	Metolachlor	97.0	96.7	P/P	71 - 124
1884988	Metribuzin	130	129	P/P	61 - 140
1884988	Mevinphos	123	110	P/P	38 - 130
1884988	Molinate	79.6	73.1	P/P	23 - 99.0
1884988	Norflurazon	108	107	P/P	39 - 139

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270D
Batch ID: P322762

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884988	Parathion Ethyl	95.3	93.9	P/P	75 - 111
1884988	Parathion Methyl	95.2	95.5	P/P	64 - 130
1884988	Pendimethalin	104	104	P/P	26 - 173
1884988	Phorate	123	118	P/P	43 - 127
1884988	Prometon	114	114	P/P	50 - 144
1884988	Prometryn	114	117	P/P	69 - 126
1884988	Simazine	114	120	P/P	51 - 152
1884988	Terbufos	103	100	P/P	54 - 125
1884988	Terbutylazine	94.9	95.5	P/P	76 - 113

Reference Method: EPA 8276
Batch ID: P322674

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885181	Chlordane	91.6	91.3	P/P	42 - 113
1885181	Toxaphene	97.9	99.0	P/P	40 - 120

Reference Method: EPA 8321B
Batch ID: P322530

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884563	Sucralose	76.5	76.6	P/P	40 - 140

Reference Method: EPA 8321B
Batch ID: P322576

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885156	Glyphosate	53.2	53.8	P/P	40 - 140

Reference Method: EPA 8321B
Batch ID: P322599

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884219	Pyraclostrobin	84.4	90.4	P/P	40 - 140

Reference Method: SM 4500 F-C-97
Batch ID: P323007

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1885134	Fluoride	98.9	98.9	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P322613

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884977	Organic Carbon	102		P	85 - 115

Reference Method: USGS O-2060-01
Batch ID: P322600

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884497	2,4-D	109	103	P/P	40 - 140

Quality Assurance Report Matrix Spike Accuracy

Reference Method: USGS O-2060-01
Batch ID: P322600

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1884497	Acetaminophen	70.8	55.6	P/P	40 - 140
1884497	Bentazon	160	160	F/F	40 - 140
1884497	Carbamazepine	83.2	102	P/P	40 - 140
1884497	Diuron	81.6	73.2	P/P	40 - 140
1884497	Fenuron	89.2	82.0	P/P	40 - 140
1884497	Fluridone	85.2	71.6	P/P	40 - 140
1884497	Imidacloprid	102	85.6	P/P	40 - 160
1884497	Linuron	74.0	87.6	P/P	40 - 140
1884497	MCP	122	104	P/P	40 - 140
1884497	Primidone	92.4	65.6	P/P	40 - 140
1884497	Triclopyr	96.0	87.6	P/P	40 - 140

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P322687

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884976	Turbidity	1.92	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P322663

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884721	Calcium	0.503	Spike	P	0 - 20
1884721	Iron	1.04	Spike	P	0 - 20
1884721	Magnesium	0.507	Spike	P	0 - 20
1884721	Potassium	0.382	Spike	P	0 - 20
1884721	Sodium	1.06	Spike	P	0 - 20
1884721	Zinc	1.24	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P322663

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884721	Arsenic	1.61	Spike	P	0 - 20
1884721	Cadmium	1.16	Spike	P	0 - 20
1884721	Chromium	1.34	Spike	P	0 - 20
1884721	Copper	0.431	Spike	P	0 - 20
1884721	Lead	0.0289	Spike	P	0 - 20
1884721	Nickel	1.40	Spike	P	0 - 20
1884721	Silver	0.429	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P323118

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885069	Chloride	0.223	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P323121

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885069	Sulfate	0.119	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P322607

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884977	Ammonia-N	5.51	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P322633

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884960	Kjeldahl Nitrogen	0.241	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322721

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884980	NO2NO3-N	1.01	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P322706

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885054	Total-P	4.65	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322609

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884981	O-Phosphate-P	1.43	Spike	P	0 - 20

Reference Method: EPA 8270D
Batch ID: P322674

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885130	Aldrin	1.13	Spike	P	0 - 30
1885130	Alpha-BHC	3.78	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D

Batch ID: P322674

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885130	Alpha-Chlordane	1.77	Spike	P	0 - 30
1885130	Beta-BHC	13.4	Spike	P	0 - 30
1885130	Carbophenothion	5.22	Spike	P	0 - 30
1885130	Chlorothalonil	28.1	Spike	P	0 - 30
1885130	Cypermethrin	11.7	Spike	P	0 - 30
1885130	DDD-p,p'	7.87	Spike	P	0 - 30
1885130	DDE-p,p'	3.48	Spike	P	0 - 30
1885130	DDT-p,p'	10.0	Spike	P	0 - 30
1885130	Delta-BHC	20.0	Spike	P	0 - 30
1885130	Dicofol	8.60	Spike	P	0 - 30
1885130	Dieldrin	1.39	Spike	P	0 - 30
1885130	Endosulfan I	9.30	Spike	P	0 - 30
1885130	Endosulfan II	2.83	Spike	P	0 - 30
1885130	Endosulfan Sulfate	8.63	Spike	P	0 - 30
1885130	Endrin	1.10	Spike	P	0 - 30
1885130	Endrin Aldehyde	4.64	Spike	P	0 - 30
1885130	Endrin Ketone	5.01	Spike	P	0 - 30
1885130	Gamma-BHC	4.17	Spike	P	0 - 30
1885130	Gamma-Chlordane	3.32	Spike	P	0 - 30
1885130	Heptachlor	1.54	Spike	P	0 - 30
1885130	Heptachlor Epoxide	2.40	Spike	P	0 - 30
1885130	Methoxychlor	12.3	Spike	P	0 - 30
1885130	Mirex	8.58	Spike	P	0 - 30
1885130	Permethrin	7.73	Spike	P	0 - 30
1885130	Trifluralin	2.79	Spike	P	0 - 30
LFB	Aldrin	27.8	LCS	P	0 - 30
LFB	Alpha-BHC	0.511	LCS	P	0 - 30
LFB	Alpha-Chlordane	2.22	LCS	P	0 - 30
LFB	Beta-BHC	6.69	LCS	P	0 - 30
LFB	Carbophenothion	6.90	LCS	P	0 - 30
LFB	Chlorothalonil	14.0	LCS	P	0 - 30
LFB	Cypermethrin	0.665	LCS	P	0 - 30
LFB	DDD-p,p'	1.03	LCS	P	0 - 30
LFB	DDE-p,p'	0.412	LCS	P	0 - 30
LFB	DDT-p,p'	3.17	LCS	P	0 - 30
LFB	Delta-BHC	7.67	LCS	P	0 - 30
LFB	Dicofol	8.00	LCS	P	0 - 30
LFB	Dieldrin	0.0500	LCS	P	0 - 30
LFB	Endosulfan I	0.593	LCS	P	0 - 30
LFB	Endosulfan II	2.48	LCS	P	0 - 30
LFB	Endosulfan Sulfate	5.38	LCS	P	0 - 30
LFB	Endrin	0.759	LCS	P	0 - 30
LFB	Endrin Aldehyde	1.40	LCS	P	0 - 30
LFB	Endrin Ketone	2.42	LCS	P	0 - 30
LFB	Gamma-BHC	4.50	LCS	P	0 - 30
LFB	Gamma-Chlordane	0.242	LCS	P	0 - 30
LFB	Heptachlor	0.136	LCS	P	0 - 30
LFB	Heptachlor Epoxide	0.329	LCS	P	0 - 30
LFB	Methoxychlor	11.4	LCS	P	0 - 30
LFB	Mirex	0.631	LCS	P	0 - 30
LFB	Permethrin	0.146	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D

Batch ID: P322674

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Trifluralin	7.54	LCS	P	0 - 30

Reference Method: EPA 8270D

Batch ID: P322762

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884988	Acetochlor	0.273	Spike	P	0 - 30
1884988	Alachlor	0.179	Spike	P	0 - 30
1884988	Ametryn	1.80	Spike	P	0 - 30
1884988	Atrazine	2.08	Spike	P	0 - 30
1884988	Atrazine Desethyl	15.6	Spike	P	0 - 30
1884988	Azinphos Methyl	7.53	Spike	P	0 - 30
1884988	Bromacil	0.386	Spike	P	0 - 30
1884988	Butylate	18.7	Spike	P	0 - 30
1884988	Chlorpyrifos Ethyl	7.98	Spike	P	0 - 30
1884988	Chlorpyrifos Methyl	7.99	Spike	P	0 - 30
1884988	Cyanazine	1.45	Spike	P	0 - 30
1884988	Demeton	7.33	Spike	P	0 - 30
1884988	Diazinon	0.0384	Spike	P	0 - 30
1884988	Disulfoton	5.21	Spike	P	0 - 30
1884988	EPTC	15.2	Spike	P	0 - 30
1884988	Ethion	35.4	Spike	F	0 - 30
1884988	Ethoprop	5.33	Spike	P	0 - 30
1884988	Fenamiphos	0.748	Spike	P	0 - 30
1884988	Fipronil	1.41	Spike	P	0 - 30
1884988	Fipronil Sulfide	9.18	Spike	P	0 - 30
1884988	Fipronil Sulfone	6.79	Spike	P	0 - 30
1884988	Fonofos	8.96	Spike	P	0 - 30
1884988	Hexazinone	1.94	Spike	P	0 - 30
1884988	Malathion	6.35	Spike	P	0 - 30
1884988	Metalaxyl	2.12	Spike	P	0 - 30
1884988	Metolachlor	0.273	Spike	P	0 - 30
1884988	Metribuzin	0.332	Spike	P	0 - 30
1884988	Mevinphos	11.0	Spike	P	0 - 30
1884988	Molinate	8.57	Spike	P	0 - 30
1884988	Norflurazon	0.820	Spike	P	0 - 30
1884988	Parathion Ethyl	1.44	Spike	P	0 - 30
1884988	Parathion Methyl	0.353	Spike	P	0 - 30
1884988	Pendimethalin	0.510	Spike	P	0 - 30
1884988	Phorate	4.17	Spike	P	0 - 30
1884988	Prometon	0.0282	Spike	P	0 - 30
1884988	Prometryn	2.48	Spike	P	0 - 30
1884988	Simazine	5.39	Spike	P	0 - 30
1884988	Terbufos	2.97	Spike	P	0 - 30
1884988	Terbutylazine	0.676	Spike	P	0 - 30
LFB	Acetochlor	2.40	LCS	P	0 - 30
LFB	Alachlor	2.53	LCS	P	0 - 30
LFB	Ametryn	6.41	LCS	P	0 - 30
LFB	Atrazine	7.65	LCS	P	0 - 30
LFB	Atrazine Desethyl	11.1	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D

Batch ID: P322762

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Azinphos Methyl	3.28	LCS	P	0 - 30
LFB	Bromacil	26.3	LCS	P	0 - 30
LFB	Butylate	27.8	LCS	P	0 - 30
LFB	Chlorpyrifos Ethyl	8.68	LCS	P	0 - 30
LFB	Chlorpyrifos Methyl	4.32	LCS	P	0 - 30
LFB	Cyanazine	4.50	LCS	P	0 - 30
LFB	Demeton	5.57	LCS	P	0 - 30
LFB	Diazinon	0.738	LCS	P	0 - 30
LFB	Disulfoton	1.86	LCS	P	0 - 30
LFB	EPTC	22.7	LCS	P	0 - 30
LFB	Ethion	21.0	LCS	P	0 - 30
LFB	Ethoprop	1.57	LCS	P	0 - 30
LFB	Fenamiphos	7.54	LCS	P	0 - 30
LFB	Fipronil	5.31	LCS	P	0 - 30
LFB	Fipronil Sulfide	6.85	LCS	P	0 - 30
LFB	Fipronil Sulfone	8.91	LCS	P	0 - 30
LFB	Fonofos	1.97	LCS	P	0 - 30
LFB	Hexazinone	8.51	LCS	P	0 - 30
LFB	Malathion	7.85	LCS	P	0 - 30
LFB	Metalaxyl	4.87	LCS	P	0 - 30
LFB	Metolachlor	5.48	LCS	P	0 - 30
LFB	Metribuzin	2.36	LCS	P	0 - 30
LFB	Mevinphos	3.30	LCS	P	0 - 30
LFB	Molinate	10.6	LCS	P	0 - 30
LFB	Norflurazon	4.44	LCS	P	0 - 30
LFB	Parathion Ethyl	9.17	LCS	P	0 - 30
LFB	Parathion Methyl	2.81	LCS	P	0 - 30
LFB	Pendimethalin	8.66	LCS	P	0 - 30
LFB	Phorate	5.46	LCS	P	0 - 30
LFB	Prometon	8.04	LCS	P	0 - 30
LFB	Prometryn	4.13	LCS	P	0 - 30
LFB	Simazine	4.92	LCS	P	0 - 30
LFB	Terbufos	2.19	LCS	P	0 - 30
LFB	Terbuthylazine	3.66	LCS	P	0 - 30

Reference Method: EPA 8276

Batch ID: P322674

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885181	Chlordane	0.369	Spike	P	0 - 30
1885181	Toxaphene	1.13	Spike	P	0 - 30
LFB	Chlordane	0.947	LCS	P	0 - 30
LFB	Toxaphene	6.83	LCS	P	0 - 30

Reference Method: EPA 8321B

Batch ID: P322530

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884563	Sucralose	0.131	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P322576

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885156	Glyphosate	1.12	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P322599

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884219	Pyraclostrobin	6.86	Spike	P	0 - 30

Reference Method: SM 2120 B
Batch ID: P322589

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884979	Color (true)	1.08	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P323004

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885134	Total Alkalinity	0.0234	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P322986

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884976	TDS	2.72	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P323007

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1885134	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P322613

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884977	Organic Carbon	0.290	Spike	P	0 - 20

Reference Method: USGS O-2060-01
Batch ID: P322600

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884497	2,4-D	5.28	Spike	P	0 - 30
1884497	Acetaminophen	24.1	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: USGS O-2060-01
Batch ID: P322600

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1884497	Bentazon	0.250	Spike	P	0 - 30
1884497	Carbamazepine	20.3	Spike	P	0 - 30
1884497	Diuron	10.9	Spike	P	0 - 30
1884497	Fenuron	8.41	Spike	P	0 - 30
1884497	Fluridone	17.3	Spike	P	0 - 30
1884497	Imidacloprid	17.9	Spike	P	0 - 30
1884497	Linuron	16.8	Spike	P	0 - 30
1884497	MCP	15.2	Spike	P	0 - 30
1884497	Primidone	33.9	Spike	F	0 - 30
1884497	Triclopyr	9.15	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 1884985
Field Sample ID: G2CE0108

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Glyphosate 13C 15N	99.4	P	50 - 150
EPA 8321B	Sucralose-d6	84.5	P	40 - 150
USGS O-2060-01	2,4-D-d3	113	P	40 - 150
USGS O-2060-01	Acetaminophen-d4	58.4	P	30 - 150
USGS O-2060-01	Carbamazepine-d10	58.0	P	40 - 160
USGS O-2060-01	Diuron-d6	54.8	P	40 - 150
USGS O-2060-01	Primidone-d5	68.0	P	40 - 150

Lab Sample ID: 1884986
Field Sample ID: G2CE0108

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Diuron-d6	144	P	40 - 150

Lab Sample ID: 1884987
Field Sample ID: G2CE0108

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Decachlorobiphenyl	65.2	P	51 - 123
EPA 8270D	Tetrachloro-m-xylene	38.0	P	31 - 111
EPA 8276	Decachlorobiphenyl	47.0	P	41 - 106

Lab Sample ID: 1884988
Field Sample ID: G2CE0108

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Malathion-d7	92.3	P	62 - 132
EPA 8270D	Simazine-d10	93.3	P	72 - 127

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A75503

Included Lab Sample IDs: 1884976, 1884979

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	102	P/P	85 - 115
Color (true)	102	102	P/P	85 - 115

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A75505

Included Lab Sample IDs: 1884977, 1884980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A75506

Included Lab Sample IDs: 1884978, 1884981

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	99.8	P/P	90 - 110
O-Phosphate-P	99.7	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A75508

Included Lab Sample IDs: 1884977, 1884980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.3	96.3	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A75544

Included Lab Sample IDs: 1884976, 1884979

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A75553

Included Lab Sample IDs: 1884977, 1884980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	95.0	94.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A75568

Included Lab Sample IDs: 1884977, 1884980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	105	P/P	90 - 110
Kjeldahl Nitrogen	105	110	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A75569

Included Lab Sample IDs: 1884989

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	95.5	95.4	P/P	90 - 110
Cadmium	94.5	93.4	P/P	90 - 110
Chromium	98.6	98.3	P/P	90 - 110
Copper	96.0	96.0	P/P	90 - 110
Lead	97.8	97.0	P/P	90 - 110
Nickel	96.4	95.4	P/P	90 - 110
Silver	95.0	94.9	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A75575

Included Lab Sample IDs: 1884986

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Pyraclostrobin	75.2	123	P/P	60 - 150

Reference Method: EPA 8321B

Run ID: A75581

Included Lab Sample IDs: 1884985

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Glyphosate	124	129	P/P	70 - 130

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A75585

Included Lab Sample IDs: 1884977, 1884980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	101	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A75593

Included Lab Sample IDs: 1884989

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	98.8	99.3	P/P	90 - 110
Iron	100	97.7	P/P	90 - 110
Magnesium	99.8	97.5	P/P	90 - 110
Potassium	100	98.1	P/P	90 - 110
Sodium	98.9	98.1	P/P	90 - 110
Zinc	99.7	98.3	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A75665

Included Lab Sample IDs: 1884976, 1884979

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	107	104	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97

Run ID: A75665

Included Lab Sample IDs: 1884976, 1884979

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A75673

Included Lab Sample IDs: 1884985

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	109	124	P/P	60 - 150

Reference Method: USGS O-2060-01

Run ID: A75682

Included Lab Sample IDs: 1884985

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetaminophen	130	123	P/P	60 - 150
Carbamazepine	117	118	P/P	60 - 150
Diuron	114	110	P/P	60 - 150
Fenuron	110	112	P/P	60 - 150
Fluridone	106	107	P/P	60 - 150
Imidacloprid	94.0	96.8	P/P	60 - 150
Linuron	145	150	P/P	60 - 150
Primidone	122	80.0	P/P	60 - 150

Reference Method: USGS O-2060-01

Run ID: A75705

Included Lab Sample IDs: 1884985

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	137	120	P/P	50 - 150
Bentazon	94.2	91.4	P/P	50 - 150
MCP	123	133	P/P	50 - 150
Triclopyr	113	103	P/P	50 - 150

Reference Method: EPA 8270D

Run ID: A75707

Included Lab Sample IDs: 1884987

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	100		P	80 - 120
Alpha-BHC	101		P	80 - 120
Alpha-Chlordane	98.8		P	80 - 120
Beta-BHC	93.9		P	80 - 120
Carbophenothion	82.4		P	80 - 120
Chlorothalonil	119		P	80 - 120
Cypermethrin	91.0		P	80 - 120
DDD-p,p'	96.9		P	80 - 120
DDE-p,p'	102		P	80 - 120
DDT-p,p'	104		P	80 - 120
Delta-BHC	101		P	80 - 120
Dicofol	108		P	80 - 120
Dieldrin	98.2		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270D

Run ID: A75707

Included Lab Sample IDs: 1884987

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Endosulfan I	101		P	80 - 120
Endosulfan II	91.7		P	80 - 120
Endosulfan Sulfate	84.2		P	80 - 120
Endrin	103		P	80 - 120
Endrin Aldehyde	91.0		P	80 - 120
Endrin Ketone	89.8		P	80 - 120
Gamma-BHC	96.0		P	80 - 120
Gamma-Chlordane	99.8		P	80 - 120
Heptachlor	102		P	80 - 120
Heptachlor Epoxide	98.1		P	80 - 120
Methoxychlor	107		P	80 - 120
Mirex	100		P	80 - 120
Permethrin	90.8		P	80 - 120
Trifluralin	103		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A75723

Included Lab Sample IDs: 1884976, 1884979

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	103	P/P	90 - 110
Sulfate	97.1	101	P/P	90 - 110

Reference Method: EPA 8270D

Run ID: A75825

Included Lab Sample IDs: 1884988

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	102		P	80 - 120
Alachlor	103		P	80 - 120
Ametryn	114		P	80 - 120
Atrazine	99.0		P	80 - 120
Atrazine Desethyl	108		P	80 - 120
Azinphos Methyl	113		P	80 - 120
Bromacil	110		P	80 - 120
Butylate	107		P	80 - 120
Chlorpyrifos Ethyl	93.7		P	80 - 120
Chlorpyrifos Methyl	94.0		P	80 - 120
Cyanazine	113		P	80 - 120
Demeton	102		P	80 - 120
Diazinon	103		P	80 - 120
Disulfoton	106		P	80 - 120
EPTC	109		P	80 - 120
Ethion	92.4		P	80 - 120
Ethoprop	101		P	80 - 120
Fenamiphos	95.6		P	80 - 120
Fipronil	103		P	80 - 120
Fipronil Sulfide	95.4		P	80 - 120
Fipronil Sulfone	98.0		P	80 - 120
Fonofos	98.5		P	80 - 120
Hexazinone	99.3		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270D
 Run ID: A75825
 Included Lab Sample IDs: 1884988

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Malathion	93.6		P	80 - 120
Metalaxyl	103		P	80 - 120
Metolachlor	107		P	80 - 120
Metribuzin	111		P	80 - 120
Mevinphos	80.5		P	80 - 120
Molinate	112		P	80 - 120
Norflurazon	100		P	80 - 120
Parathion Ethyl	90.2		P	80 - 120
Parathion Methyl	88.6		P	80 - 120
Pendimethalin	104		P	80 - 120
Phorate	102		P	80 - 120
Prometon	112		P	80 - 120
Prometryn	117		P	80 - 120
Simazine	109		P	80 - 120
Terbufos	94.0		P	80 - 120
Terbutylazine	101		P	80 - 120

Reference Method: EPA 8276
 Run ID: A75867
 Included Lab Sample IDs: 1884987

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	98.1		P	80 - 120
Toxaphene	102		P	80 - 120

* Pass/Fail determinations are made for each bracketing calibration verification check.
 Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
 Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				1.92	
EPA 200.7 Rev. 4.4	Calcium	102	100 99.5			0.503
	Iron	104	104 103 104			1.04
	Magnesium	103	101 98.4			0.507
	Potassium	102	101 103 103			0.382
	Sodium	102	102 101			1.06
	Zinc	99.7	98.8 96.8 98.0			1.24
EPA 200.8 Rev. 5.4	Arsenic	96.9	96.6 98.8 97.2			1.61
	Cadmium	93.7	92.5 96.3 95.2			1.16
	Chromium	96.6	95.1 98.1 96.7			1.34
	Copper	95.4	92.2 92.9 92.5			0.431
	Lead	99.2	97.9 98.3 98.3			0.0289
	Nickel	97.6	93.6 95.4 94.0			1.40
	Silver	98.1	98.1 99.2 98.8			0.429
EPA 300.0 Rev. 2.1	Chloride	101	99.0 98.7		0.223	
	Sulfate	97.3	95.6		0.119	

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		LCS	Precision	MS
							SMP	
EPA 350.1 Rev. 2.0	Ammonia-N	101		106	100			5.51
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104		103	103			0.241
EPA 353.2 Rev. 2.0	NO2NO3-N	98.5		99.5	98.5			1.01
EPA 365.1 Rev. 2.0	Total-P	107		108	102			4.65
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.4		97.0	98.4			1.43
EPA 8270D	Acetochlor	92.7	90.5	94.5	94.2	2.40		0.273
	Alachlor	93.2	90.9	98.7	98.5	2.53		0.179
	Aldrin	46.9	62.0	71.7	72.5	27.8		1.13
	Alpha-BHC	86.3	85.9	88.2	84.9	0.511		3.78
	Alpha-Chlordane	92.8	94.9	98.7	97.0	2.22		1.77
	Ametryn	113	106	128	130	6.41		1.80
	Atrazine	104	96.6			7.65		2.08
	Atrazine Desethyl	84.2	75.4	67.8	56.4	11.1		15.6
	Azinphos Methyl	144	149	185	171	3.28		7.53
	Beta-BHC	88.8	83.0	91.9	105	6.69		13.4
	Bromacil	99.7	76.5	83.0	83.3	26.3		0.386
	Butylate	72.2	54.6	68.8	57.0	27.8		18.7
	Carbophenothion	96.6	90.2	122	116	6.90		5.22
	Chlorothalonil	118	102	60.9	80.9	14.0		28.1
	Chlorpyrifos Ethyl	103	94.1	107	98.9	8.68		7.98
	Chlorpyrifos Methyl	97.2	93.1	95.5	88.2	4.32		7.99
	Cyanazine	121	116	120	118	4.50		1.45
	Cypermethrin	89.8	89.2	101	89.9	0.665		11.7
	DDD-p,p'	92.7	91.7	93.9	86.8	1.03		7.87
	DDE-p,p'	98.8	99.2	103	99.8	0.412		3.48
	DDT-p,p'	85.2	87.9	92.9	84.0	3.17		10.0
	Delta-BHC	105	97.3	133	109	7.67		20.0
	Demeton	104	98.2	115	107	5.57		7.33
	Diazinon	90.6	89.9	97.8	97.8	0.738		0.0384
	Dicofol	82.4	89.3	89.1	97.1	8.00		8.60
	Dieldrin	93.9	93.9	98.8	97.5	0.0500		1.39
	Disulfoton	117	115	137	130	1.86		5.21
	Endosulfan I	97.1	97.7	111	122	0.593		9.30
	Endosulfan II	93.2	91.0	108	105	2.48		2.83
	Endosulfan Sulfate	85.8	81.3	95.6	87.7	5.38		8.63
	Endrin	103	104	103	101	0.759		1.10
	Endrin Aldehyde	93.5	92.2	56.3	58.9	1.40		4.64
	Endrin Ketone	78.8	76.9	89.1	84.8	2.42		5.01
	EPTC	70.6	56.2	70.3	60.4	22.7		15.2
	Ethion	119	96.5	108	75.3	21.0		35.4
	Ethoprop	93.0	91.6	105	99.6	1.57		5.33
	Fenamiphos	122	113	89.9	89.3	7.54		0.748
	Fipronil	103	97.4	103	101	5.31		1.41
	Fipronil Sulfide	102	95.4	96.1	86.2	6.85		9.18
	Fipronil Sulfone	96.2	88.0	86.3	79.1	8.91		6.79
	Fonofos	91.0	89.2	103	94.2	1.97		8.96
	Gamma-BHC	91.1	87.1	112	116	4.50		4.17
	Gamma-Chlordane	92.3	92.5	98.7	95.5	0.242		3.32
	Heptachlor	84.5	84.4	81.8	80.5	0.136		1.54
	Heptachlor Epoxide	92.7	93.0	98.6	96.2	0.329		2.40
	Hexazinone	96.9	89.0	93.8	95.6	8.51		1.94
	Malathion	117	109	113	106	7.85		6.35
	Metalaxyl	93.9	89.5	94.7	96.7	4.87		2.12
	Methoxychlor	90.9	81.1	96.4	85.2	11.4		12.3

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270D	Metolachlor	102	96.6	97.0	96.7	5.48	0.273
	Metribuzin	125	128	130	129	2.36	0.332
	Mevinphos	114	110	123	110	3.30	11.0
	Mirex	81.4	80.9	91.2	83.7	0.631	8.58
	Molinate	75.6	68.0	79.6	73.1	10.6	8.57
	Norflurazon	115	110	108	107	4.44	0.820
	Parathion Ethyl	99.9	91.1	95.3	93.9	9.17	1.44
	Parathion Methyl	85.1	87.5	95.2	95.5	2.81	0.353
	Pendimethalin	103	94.4	104	104	8.66	0.510
	Permethrin	83.1	83.2	99.1	91.8	0.146	7.73
	Phorate	102	97.0	123	118	5.46	4.17
	Prometon	121	111	114	114	8.04	0.0282
	Prometryn	113	108	114	117	4.13	2.48
	Simazine	99.8	95.0	114	120	4.92	5.39
	Terbufos	85.7	83.8	103	100	2.19	2.97
	Terbuthylazine	93.8	90.5	94.9	95.5	3.66	0.676
	Trifluralin	116	108	140	136	7.54	2.79
EPA 8276	Chlordane	91.0	91.8	91.6	91.3	0.947	0.369
	Toxaphene	92.3	98.8	97.9	99.0	6.83	1.13
EPA 8321B	Glyphosate	131		53.2	53.8		1.12
	Pyraclostrobin	112		84.4	90.4		6.86
	Sucralose	99.6		76.5	76.6		0.131
SM 2120 B	Color (true)					1.08	
SM 2320 B-97	Total Alkalinity	105				0.0234	
SM 2540 C-97	TDS					2.72	
SM 4500 F-C-97	Fluoride	97.4		98.9	98.9		0.0
SM 5310 B-00	Organic Carbon	97.8		102			0.290
USGS O-2060-01	2,4-D	104		109	103		5.28
	Acetaminophen	53.2		70.8	55.6		24.1
	Bentazon	83.6		160	160		0.250
	Carbamazepine	121		83.2	102		20.3
	Diuron	87.6		81.6	73.2		10.9
	Fenuron	98.0		89.2	82.0		8.41
	Fluridone	75.6		85.2	71.6		17.3
	Imidacloprid	85.6		102	85.6		17.9
	Linuron	92.4		74.0	87.6		16.8
	MCPP	105		122	104		15.2
	Primidone	86.8		92.4	65.6		33.9
	Triclopyr	94.8		96.0	87.6		9.15

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1884976, 1884979
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	1884989
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	1884989
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1884976, 1884979
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1884976, 1884979

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1884977, 1884980
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1884977, 1884980
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1884977, 1884980
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1884977, 1884980
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1884978, 1884981
EPA 8270D / E31780	Organochlorine pesticides in water matrices by GC/MS-SIM	1884987
EPA 8270D / E31780	Organonitrogen and phosphorus pesticides (ultra trace level) in water matrices by GC/MS/MS	1884988
EPA 8276 / E31780	Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization	1884987
EPA 8321B / E31780	Glyphosate in water matrices by HPLC/MS/MS	1884985
EPA 8321B / E31780	Pyraclostrobin in water matrices by HPLC/MS/MS	1884986
EPA 8321B / E31780	Water soluble organic compounds by HPLC/MS/MS	1884985
SM 2120 B / E31780	True Color measured at 450 nm	1884976, 1884979
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1884976, 1884979
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1884976, 1884979
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1884976, 1884979
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1884976, 1884979
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1884977, 1884980
USGS O-2060-01 / E31780	Chlorinated (phenoxy acid) herbicides in water matrices by HPLC/MS/MS	1884985
USGS O-2060-01 / E31780	Urea herbicides and other chemicals in water matrices by HPLC/MS/MS	1884985

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/31/2017			03/31/2017 15:55	Blanca Fach	1884976, 1884979
EPA 200.7 Rev. 4.4	03/31/2017	04/03/2017	Elliott D. Healy	04/05/2017	Martin Acosta	1884989
EPA 200.8 Rev. 5.4	03/31/2017	04/03/2017	Elliott D. Healy	04/04/2017	Betina Topolski	1884989
EPA 300.0 Rev. 2.1	03/31/2017			04/11/2017	Ping Hua	1884976, 1884979
EPA 350.1 Rev. 2.0	03/31/2017			03/31/2017	Julie Michelle Frank	1884977, 1884980
EPA 351.2 Rev. 2.0	03/31/2017	04/03/2017	Adrian Shelby	04/04/2017	Joseph Suarez	1884977, 1884980
EPA 353.2 Rev. 2.0	03/31/2017			04/04/2017	Beverly Y. Sanders	1884977, 1884980
EPA 365.1 Rev. 2.0	03/31/2017	04/04/2017	Adrian Shelby	04/05/2017	Lipi Saha	1884977, 1884980
EPA 365.1 Rev. 2.0 dissolved	03/31/2017			03/31/2017 15:05	Anthony C. Onicha	1884981
	03/31/2017			03/31/2017 15:30	Anthony C. Onicha	1884978
EPA 8270D	03/31/2017	04/04/2017	Fe-Val Siddell	04/07/2017	Vandana T. Nagar	1884987
	03/31/2017	04/04/2017	Rasheda Ghaffari	04/05/2017	Irina Carbone	1884988
EPA 8276	03/31/2017	04/04/2017	Fe-Val Siddell	04/15/2017	Vandana T. Nagar	1884987
EPA 8321B	03/31/2017	04/03/2017	Fe-Val Siddell	04/05/2017	Juyoung Kim	1884986
	03/31/2017	04/05/2017	Mohammad Ghaffari	04/05/2017	Mohammad Ghaffari	1884985
	03/31/2017	04/07/2017	Mohammad Ghaffari	04/07/2017	Mohammad Ghaffari	1884985
SM 2120 B	03/31/2017			03/31/2017 14:24	Julia Storbeck	1884979
	03/31/2017			03/31/2017 14:40	Julia Storbeck	1884976
SM 2320 B-97	03/31/2017			04/07/2017	Dale L. Simmons	1884976, 1884979
SM 2540 C-97	03/31/2017			04/05/2017	Yijie Li	1884976, 1884979
SM 2540 D-97	03/31/2017			04/05/2017	Yijie Li	1884976, 1884979
SM 4500 F-C-97	03/31/2017			04/07/2017	Dale L. Simmons	1884976, 1884979
SM 5310 B-00	03/31/2017			03/31/2017	Christopher Armour	1884977, 1884980
USGS O-2060-01	03/31/2017	04/03/2017	Hoor Shaik	04/08/2017	Juyoung Kim	1884985
	03/31/2017	04/03/2017	Hoor Shaik	04/10/2017	Juyoung Kim	1884985