

Chemical Analysis Report

CEN-DIST-2017-05-11-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Groover Branch**
Request ID: **RQ-2017-05-08-74**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 06-JUN-2017 11:01



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Groover Branch @ Airport Rd Bridge

Collection Date/Time: 05/10/2017 10:40

Field ID: G5CE0040

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1898476	EPA 180.1 Rev. 2.0	Turbidity	2.2		NTU	P325129	
	EPA 300.0 Rev. 2.1	Chloride	64		mg Cl/L	P325828	
		Sulfate	12		mg SO4/L	P325832	
	SM 2120 B	Color (true)	43		PCU	P325136	
	SM 2320 B-97	Total Alkalinity	212		mg CaCO3/L	P325591	
	SM 2540 C-97	TDS	381	A	mg/L	P325653	
	SM 2540 D-97	TSS	5	IA	mg/L	P325532	
	SM 4500 F-C-97	Fluoride	0.15		mg F/L	P325594	
1898477	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P325121	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.95		mg N/L	P325450	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.034		mg N/L	P325425	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P325251	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P325516	
1898478	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.044		mg P/L	P325171	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P325129

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P325828

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P325832

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P325121

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P325450

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P325425

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P325251

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P325171

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P325136

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P325591

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P325653

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P325532

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P325594

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P325516

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P325828

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P325832

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P325121

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P325450

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P325425

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P325251

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P325171

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P325591

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P325594

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	94.2		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P325516

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	89.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P325828

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898480	Chloride	99.4		P	90 - 110
1898482	Chloride	98.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P325832

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898480	Sulfate	97.1		P	90 - 110
1898482	Sulfate	96.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P325121

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898477	Ammonia-N	105	106	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P325450

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898619	Kjeldahl Nitrogen	95.0	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P325425

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898477	NO2NO3-N	100	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P325251

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898485	Total-P	98.9	97.3	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P325171

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898389	O-Phosphate-P	96.0	96.6	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P325594

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898378	Fluoride	98.7	95.7	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P325516

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1898243	Organic Carbon	86.1	88.0	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P325129

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898482	Turbidity	0.690	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P325828

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898482	Chloride	0.0329	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P325832

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898482	Sulfate	0.709	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P325121

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898477	Ammonia-N	1.28	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P325450

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898619	Kjeldahl Nitrogen	3.42	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P325425

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P325425

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898477	NO2NO3-N	1.27	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P325251

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898485	Total-P	0.886	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P325171

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898389	O-Phosphate-P	0.623	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P325136

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898480	Color (true)	0.809	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P325591

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898378	Total Alkalinity	3.21	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P325653

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898476	TDS	1.05	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P325594

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898378	Fluoride	2.57	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P325516

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1898243	Organic Carbon	1.33	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A76358

Included Lab Sample IDs: 1898477

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	103	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A76360

Included Lab Sample IDs: 1898476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A76361

Included Lab Sample IDs: 1898476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	100	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A76368

Included Lab Sample IDs: 1898478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.1	96.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A76452

Included Lab Sample IDs: 1898477

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A76480

Included Lab Sample IDs: 1898477

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	99.0	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A76527

Included Lab Sample IDs: 1898477

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.4	94.0	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 2320 B-97
 Run ID: A76561
 Included Lab Sample IDs: 1898476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Run ID: A76561
 Included Lab Sample IDs: 1898476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	95.5	98.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Run ID: A76562
 Included Lab Sample IDs: 1898477

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	107	107	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Run ID: A76672
 Included Lab Sample IDs: 1898476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	98.6	94.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
 Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
 Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.690	
EPA 300.0 Rev. 2.1	Chloride	101	98.8 99.4		0.0329	
	Sulfate	97.3	96.6 97.1		0.709	
EPA 350.1 Rev. 2.0	Ammonia-N	102	105 106			1.28
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	95.0 102			3.42
EPA 353.2 Rev. 2.0	NO2NO3-N	102	100 102			1.27
EPA 365.1 Rev. 2.0	Total-P	101	98.9 97.3			0.886
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	96.0 96.6			0.623
SM 2120 B	Color (true)				0.809	
SM 2320 B-97	Total Alkalinity	104			3.21	
SM 2540 C-97	TDS				1.05	
SM 4500 F-C-97	Fluoride	94.2	98.7 95.7			2.57
SM 5310 B-00	Organic Carbon	89.8	86.1 88.0			1.33

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1898476
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1898476
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1898476
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1898477
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1898477
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1898477
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1898477
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1898478
SM 2120 B / E31780	True Color measured at 450 nm	1898476
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1898476
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1898476
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1898476
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1898476
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1898477

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/11/2017			05/11/2017 15:50	Sima Feizi	1898476
EPA 300.0 Rev. 2.1	05/11/2017			05/25/2017	Ping Hua	1898476
EPA 350.1 Rev. 2.0	05/11/2017			05/11/2017	Julie Michelle Frank	1898477
EPA 351.2 Rev. 2.0	05/11/2017	05/18/2017	Adrian Shelby	05/19/2017	Joseph Suarez	1898477
EPA 353.2 Rev. 2.0	05/11/2017			05/17/2017	Beverly Y. Sanders	1898477
EPA 365.1 Rev. 2.0	05/11/2017	05/15/2017	Vijayalakshmi Reddy	05/16/2017	Lipi Saha	1898477
EPA 365.1 Rev. 2.0 dissolved	05/11/2017			05/11/2017 15:32	Anthony C. Onicha	1898478
SM 2120 B	05/11/2017			05/11/2017 17:04	Julia Storbeck	1898476
SM 2320 B-97	05/11/2017			05/20/2017	Dale L. Simmons	1898476
SM 2540 C-97	05/11/2017			05/15/2017	Sima Feizi	1898476
SM 2540 D-97	05/11/2017			05/15/2017	Sima Feizi	1898476
SM 4500 F-C-97	05/11/2017			05/20/2017	Dale L. Simmons	1898476
SM 5310 B-00	05/11/2017			05/16/2017	Christopher Armour	1898477